



OFFICE NAME: Larry R. Goldstein, DDS, PC
 ADDRESS: 2606 Abbey Court PHONE NUMBER: (770)475-3210
Alpharetta, GA 30004 FAX: (888)844-4037
 CONTACT PERSON: Larry Goldstein COMPLIANCE OFFICER EMAIL(s) blackbirdsix@gr
 OTHER EMAIL(s): _____

DUWL Treatment (Select TABLET, STRAW or OTHER):

A. TABLET: BluTab ICX Citrisil Citrisil Blue
 How long since last shock treatment?: ___ Days or Never Shocked
 Shock Product Used:
 Sterilix Mint-A-Kleen Citrisil Shock Bleach Other: _____

B. STRAW: Dentapure Sterisil Straw Install Date: _____ Shock Before Install? Yes No

C. OTHER WATER TREATMENT: _____

D. WE DON'T HAVE WATER BOTTLES

SOURCE WATER (we fill our water bottles from):

City Water Faucet (sink name or #) _____ Bottled Distilled
 Filter, Distiller, or RO Unit (brand): TuTraver Other (e.g. delivered): _____

Received on
JUL 03 2018
 By Richard

For Lab Use Only	Sample	Location (Circle One)	Room/ Chair/ Operator #	Device (Circle One)	Other Information
151	#1	Room / Chair / <u>Operator</u>	<u># 1</u>	<u>A/W Syringe</u> / Handpiece / Scaler	
152	#2	Room / Chair / <u>Operator</u>	<u># 2</u>	<u>A/W Syringe</u> / Handpiece / Scaler	
153	#3	Room / Chair / <u>Operator</u>	<u># 3</u>	<u>A/W Syringe</u> / Handpiece / Scaler	
154	#4	Room / Chair / <u>Operator</u>	<u># 3</u>	A/W Syringe / Handpiece / <u>Scaler</u>	

Waterline Testing Results

ProEdge Sample #	Sample #	Microbial Count	Result
		(CFU/mL)	
#151	#1	0	Pass
#152	#2	0	Pass
#153	#3	5	Pass
#154	#4	10	Pass

Test Date: 07-03-2018
Lab Technician: CK

The CDC's guideline for bacteria in water used as a coolant/irrigant for non-surgical dental procedures, should be as low as reasonably achievable, and, at a minimum, less than 500 CFU/mL.