

DENTAL WATER SAFETY

Waterline Test Safety Report

Practice Information:

Larry Goldstein DDS
2606 Abbey Ct, Alpharetta, GA 30004
770-475-3210

Practice-Reported Treatment Protocol:

Shock:	Shock Date:	Treatment:	Source Water:
Sterilex ULTRA	7/27/2019	BluTab	Filter/Distiller

Test Date:	Report Date:
7/2/2019	7/9/2019

ProEdge Dental Water Labs approves the following results as accurate and complete utilizing our advanced and proven neutralization methods.


Richard Vigil, MS
Laboratory Director


Kellie Thimmes
Water Safety Specialist

					Water Lab Results		
ProEdge Sample #	Vial #	Location	Room/Chair/Operatory #	Device or Source Water	Microbial Count (CFU/mL)	Waterline Result	*Safety Level
19-33319	1	Chair	1	A/W Syringe	0	PASS	✓
19-33320	2	Operatory	1	Scaler	15	PASS	✓
19-33321	3	Chair	2	A/W Syringe	5	PASS	✓
19-33322	4	Chair	3	Handpiece	10	PASS	✓



Questions?

Our team of dental water experts are available to help explain your results and take the next steps to achieve or maintain safe water.

Call or Email
888.843.3343
info@proedgedental.com

TNTC = Too Numerous to Count. QNS = Quantity Not Sufficient (of sample water). Pass / Fail results are based on the ≤ 500 CFU/mL standard set by the CDC and ADA.
*Safety Level: Red = Immediate Shock & Retest | Yellow = Immediate Shock | Green = Continue Treatment Protocol. See attached document for more information about your report.
ProEdge Dental Water Labs neutralization methods have been validated by an independent research lab following ASTM E1054-08. For full reports, visit ProEdgeDental.com.

413-416



THEIR PROTECTION
YOUR ASSURANCE

DENTAL WATER SAFETY
Waterline Test Submission Form

ProEdgeDental.com | 888.843.3343 | 7042 S Revere Parkway #400, Centennial, CO 80112

FILL OUT THIS FORM COMPLETELY FOR EVERY TEST TO ENSURE YOUR SAMPLES CAN BE PROCESSED. KEEP A COPY FOR YOUR RECORDS.

Practice Information: New Customer Returning Customer

Practice Name: Larry Goldstein, DDS, PC Contact Person: Larry Goldstein

Practice Address: 2606 Addey Court Contact Title/Position: owner

City, State & Zip: Alpharetta, GA 30004 Contact Email: blackbirdsix@gmail.com

Phone: (770) 475-3210 Fax: (678) 544-4037 Compliance Officer Email: _____

Waterline Treatment Protocol Information (Select All That Apply):

A. TREATMENT PRODUCT(S) USED: BluTab ICX Citrisil/Citrisil Blue Sterisil Straw Sterisil Straw Other: _____

B. SHOCK PRODUCT(S) USED: Bleach Citrisil Shock Steriliz Ultra Other: _____

C. OTHER TREATMENT SYSTEMS: Sterisil System O-So Pure Vista System Other: _____

Install Date: _____ Date of Last Shock: 27 JUN 19

Shock Prior to Install: Yes No Never Shocked:

D. USING WATER BOTTLES: Yes No

E. SOURCE WATER (We fill our water bottles from...): City Water/Tap Other (e.g. Delivered): _____ In-Office Filter, Distiller or R/O Unit (Brand: Water)

Test Sampling Information:

Sampling Date: 1 JUL 19

† Do not sample on Fridays or before a holiday. Ship same day as sampled.

FOLLOW CLOSELY TO ENSURE ACCURATE TEST RESULTS

- Fill out this form completely every time you test.
- Do not sample or ship on Fridays or the day before a holiday.
- Ship samples on the same day they are taken.
- Do not include more than 16 vials in each shipment.

FOR LAB USE ONLY ProEdge Sample #	Vial Number	Location (Circle One)	Room/Chair/Operator #	Device or Source Water (Circle One)	Notes or Other Information
413	1	Room <u>Chair</u> Operator	1	<u>AW Syringe</u> Handpiece Scaler Source Water	
414	2	Room Chair <u>Operator</u>	1	AW Syringe Handpiece <u>Scaler</u> Source Water	
415	3	Room <u>Chair</u> Operator	2	<u>AW Syringe</u> Handpiece Scaler Source Water	
416	4	Room <u>Chair</u> Operator	3	AW Syringe <u>Handpiece</u> Scaler Source Water	
	5	Room Chair Operator		AW Syringe Handpiece Scaler Source Water	
	6	Room Chair Operator		AW Syringe Handpiece Scaler Source Water	
	7	Room Chair Operator		AW Syringe Handpiece Scaler Source Water	
	8	Room Chair Operator		AW Syringe Handpiece Scaler Source Water	
	9	Room Chair Operator		AW Syringe Handpiece Scaler Source Water	
	10	Room Chair Operator		AW Syringe Handpiece Scaler Source Water	
	11	Room Chair Operator		AW Syringe Handpiece Scaler Source Water	
	12	Room Chair Operator		AW Syringe Handpiece Scaler Source Water	
	13	Room Chair Operator		AW Syringe Handpiece Scaler Source Water	
	14	Room Chair Operator		AW Syringe Handpiece Scaler Source Water	
	15	Room Chair Operator		AW Syringe Handpiece Scaler Source Water	
	16	Room Chair Operator		AW Syringe Handpiece Scaler Source Water	

The guidelines outlined by the Center for Disease Control (CDC) for bacteria in water used as a coolant/irrigant for non-surgical dental procedures should be as low as reasonably achievable, and, at a minimum ≤ 500 CFU/mL. Your practice's water samples will be tested against this CDC standard for dental water safety.

Questions? Call us today at 888.843.3343

6690