



813-A Woodcrest Drive · Houston, Texas · 77018-2127
 (713)695-1939 · Fax: (713)695-3001
 http://www.thermotemp.com

Certificate Of Calibration

Calibrated for :P & B Testing
 6645 West Tidwell
 Houston Texas,77092

Report No:	120457	Calibration Interval:	1 Year
PO. No.:	86563	Calibration Date:	02/09/2024
Company No:	7088	Calibration Due Date:	02/09/2025
Description:	ROCKWELL	Serial No:	7088
Model No:	AR-12	Manufacturer:	Clark
Temperature:	64°F	Humidity :	50.00 %RH
Cal. Procedure:	CP-02		

THIS DOCUMENT IS TO CERTIFY THAT THE AFOREMENTIONED HARDNESS TESTER PASSED VERIFICATION BY INDIRECT METHOD IN THE RANGES LISTED BELOW USING A TEST BLOCK CALIBRATED IN THE STANDARDIZATION LABORATORY OF THE BLOCK MANUFACTURER (SUNTEC). ALL TESTS WERE PERFORMED IN ACCORDANCE WITH OUR QUALITY MANUAL AND THE LATEST REVISION OF ASTM E18 AND / OR E103. THERMO-TEMP INC. QUALITY ASSURANCE PROGRAM FOLLOWS THE REQUIREMENTS STATED IN ANSI / NCSL Z540-1.

This report shall not be reproduced, except in full, without the written approval of Thermo-Temp Incorporated

Asset Requirements: 0

VERIFICATION TEST DATA

TEST BLOCK	SERIAL NO	HARDNESS TESTER READINGS					(AS FOUND/AS LEFT)		
		RD1	RD2	RD3	RD4	RD5	Average	Error	Repeatability
HR C 25.80+/-1.00	222400 15	25.40	25.30	25.30	25.40	25.20	25.32	-0.48	0.20
HR C 41.90+/-1.00	224002 56	41.50	41.40	41.50	41.40	41.40	41.44	-0.46	0.10
HR C 60.40+/-0.50	226101 89	60.20	60.00	60.10	60.20	60.10	60.12	-0.28	0.20
HR Bw 43.50+/-2.50	224000 85	44.00	43.70	44.00	44.30	43.80	43.96	0.46	0.60
HR Bw 74.60+/-2.50	227002 31	74.60	75.00	74.80	74.90	74.80	74.82	0.22	0.40
HR Bw 96.20+/-1.00	229501 23	96.00	96.20	96.20	96.00	96.00	96.08	-0.12	0.20

Buck Snider

Calibration Performed By:			Certificate Authorized By:		
Jeromy Holman	Technician		Ashley Holman	2/14/2024	
Name	Title	Signature	Quality Control	Date	Signature