

# State Environmental Laboratory Services Division

EPA DRINKING WATER CERTIFICATION #OK00013 NH ELAP ACCREDITATION #2338 FOR METHOD EPA 1623.1

## **Report of Analysis**

General Inquiries: (Toll-Free) 1-866-412-3057 (E-mail) selsd@deq.ok.gov

EMAIL TO SETH LARMAN

SETH@INEXHOMEINSPECTION.COM

**PROJECT SUMMARY** 

Project: INTERI-001\_0017

Customer: INTERIOR EXTERIOR HOME INSPECT - SETH LARMAN Program: Private

INTERI-001 Subprogram: Private

**AUTHORIZING SIGNATURE** 

Bryan Gisch, Environmental Chemical Lab Scientist

**PROJECT SAMPLE SUMMARY** 

**Project Status: Complete** 

Sample Number <u>Description</u>

PRIV-1546720-01 BATHROOM SINK

**ID Code** 

All rights reserved. Report may not be reproduced, except in full, without the written approval of the SELSD. Results relate only to the specific sample aliquots tested.



# State Environmental Laboratory Services Division

EPA DRINKING WATER CERTIFICATION #OK00013 NH ELAP ACCREDITATION #2338 FOR METHOD EPA 1623.1

## **Report of Analysis**

General Inquiries: (Toll-Free) 1-866-412-3057 (E-mail) selsd@deq.ok.gov

## **SAMPLE INFORMATION**

Sample Number: PRIV-1546720-01 Collected By: SL

 Sample Address:
 20771 JUANITA LN. TECUMSEH, OK. 74873
 Collected:
 12/28/20 10:45 am

Description: BATHROOM SINK Received: 12/28/20 11:44 am

Receipt Temp: 16.9 °C

QA Code:

### **TEST RESULTS**

Analysis Method: SM 9223B Analysis: SM9223B Total Coliforms & E. coli, P/A

Component Name	Result	Unit	Qualifiers	Reviewer	Analyzed On
E. coli	Absent			BLN	12/29/2020
Total Coliform	Absent			BLN	12/29/2020