









Intacs® & Intacs® SK Corneal Implantation Instruments

<i>Product Description</i>	Intacs® Surgical Instruments consist of a series of ophthalmic instruments that are designed for the placement of Intacs® Corneal Implants.			
<i>Indications</i>	Intacs® Instruments are indicated for use in patients undergoing ophthalmic surgery requiring a lamellar dissection of the cornea.			
<i>Presentation</i>	Supplied non-sterile.			
				
<i>Reference</i>	KV10445	KV10478-2	KV10606	KV1160
<i>Instrument</i>	Intacs® Forceps	Vacuum Centering Guide, Sloped	Pocketing Hook	Symmetric Glide
<i>Material</i>	Titanium	Stainless steel	Titanium	Titanium
<i>Characteristics</i>	<ul style="list-style-type: none"> Notched tip. Facilitates easy removal of the segments from the carrier. 	<ul style="list-style-type: none"> Port for attachment to a vacuum source. Equipped with fins on the bottom. Centers and stabilizes the Corneal Separators onto the cornea during the dissection of the intrastromal tunnels. 	<ul style="list-style-type: none"> Alloy probe. Initiates the delamination of the stroma. 	<ul style="list-style-type: none"> Bifurcated glide. Guides in both the CW and CCW Corneal Separators.
				
<i>Reference</i>	KV10730-1	KV10730-2	KV10732	KV10733
<i>Instrument</i>	Corneal Separator, CW, 0.9 mm	Corneal Separator, CCW, 0.9 mm	Inspection Gauge, 0.9 mm	Procedure Marker, 0.9 mm
<i>Material</i>	Stainless steel		Ultem	Stainless steel
<i>Characteristics</i>	<ul style="list-style-type: none"> Also known as "Corneal Dissectors". 210° blade attached to the body by a spoke. Creates a CW and CCW annular subsurface tunnel. 		<ul style="list-style-type: none"> Also known as "Separator or Dissector Inspection Gauge". Inspects the concentricity and planarity of the 0.9mm Corneal Separator blades. 	<ul style="list-style-type: none"> Also known as "Incision & Placement Marker". Marks the incision. Aligns the Vacuum Centering Guide.

Additional (not Intacs specific) instruments may be needed for Intacs surgical procedure. Please contact us for additional information.