



Certificate of Conformity



according to
Low Voltage Directive 2006/95/EC

Certificate No.: M1 69240506 0001

Report No.: 28209697 002

Certificate Holder: Steriwave Plc.
Airport House Purley Way
CR0 0XZ Croydon Grt London
United Kingdom

Manufacturer: Steriwave Plc.
Airport House Purley Way
CR0 0XZ Croydon Grt London
United Kingdom

Product: Microwave boiler


Identification: Type: HC 001
Rated voltage: 230C 50Hz
Rated power input: 3kW

Factory: PLC Szoftverfejlesztő Kft.
Seltenhoffer Frigyes út 2/D
9400 Sopron Hungary

Tested according to: EN 60335-1:2002+A1+A11+A2+A12+A13
EN 60335-2-25:2002+A1+A2, Sub-Clause 19.11.2, 19.13 and
Clause 32
EN 62233:2008

This certificate refers to the above mentioned product. This is to certify that the test sample is in conformity with the requirements of Annex I of the Directive. The holder of the certificate is entitled to use this certificate for the declaration of conformity according to Annex III of the Directive. This certificate does not imply assessment of the production of the product and does not permit to use of a TÜV Rheinland or MEEI mark. The certificate is valid up to 3 years provided that the test specifications and construction of the product remain unchanged during this period.

Budapest, 2010.09.27


Ferencne Schmidt

TÜV Rheinland — MEEI Kft. Certification — H-1132 Budapest, Váci str. 48/A-B — www.meei.hu
The CE marking may be used if all relevant and effective EC Directives are complied with

