

SUPERVISOR TEST




CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 01/05/2024

OPERATOR NAME:
JAMES J SNELL
Permit #: 38312
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:53
INTERNAL STANDARD	VERIFIED	08:53
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:54
BLANK TEST	0.00	08:54
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:55
BLANK TEST	0.00	08:56
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:56
BLANK TEST	0.00	08:57



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 01/10/2024

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:06
INTERNAL STANDARD	VERIFIED	18:06
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	18:07
BLANK TEST	0.00	18:08
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	18:08
BLANK TEST	0.00	18:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	18:09
BLANK TEST	0.00	18:10



SUPERVISOR TEST




CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 01/16/2024

OPERATOR NAME:
CASEY PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:32
INTERNAL STANDARD	VERIFIED	21:32
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:32
BLANK TEST	0.00	21:33
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:34
BLANK TEST	0.00	21:34
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:35
BLANK TEST	0.00	21:36



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 01/22/2024

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:19
INTERNAL STANDARD	VERIFIED	22:19
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:20
BLANK TEST	0.00	22:21
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:21
BLANK TEST	0.00	22:22
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:22
BLANK TEST	0.00	22:23



SOLUTION CHANGE




CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 01/30/2024

OPERATOR NAME:
CASEY PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:54
INTERNAL STANDARD	VERIFIED	20:54
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	20:54
BLANK TEST	0.00	20:55
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	20:55
BLANK TEST	0.00	20:56
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	20:56
BLANK TEST	0.00	20:57



SUPERVISOR TEST



 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 02/04/2024

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:07
INTERNAL STANDARD	VERIFIED	22:07
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:08
BLANK TEST	0.00	22:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:09
BLANK TEST	0.00	22:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:10
BLANK TEST	0.00	22:11



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 02/09/2024

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:30
INTERNAL STANDARD	VERIFIED	05:30
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:31
BLANK TEST	0.00	05:31
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:32
BLANK TEST	0.00	05:32
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:33
BLANK TEST	0.00	05:34



SUPERVISOR TEST




CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 02/15/2024

OPERATOR NAME:
CASEY PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:18
INTERNAL STANDARD	VERIFIED	22:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:18
BLANK TEST	0.00	22:19
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:19
BLANK TEST	0.00	22:20
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:20
BLANK TEST	0.00	22:21



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 02/21/2024

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:18
INTERNAL STANDARD	VERIFIED	22:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:18
BLANK TEST	0.00	22:19
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:19
BLANK TEST	0.00	22:20
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:20
BLANK TEST	0.00	22:21



SUPERVISOR TEST



 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 02/27/2024

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:24
INTERNAL STANDARD	VERIFIED	22:24
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:24
BLANK TEST	0.00	22:25
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:25
BLANK TEST	0.00	22:26
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:26
BLANK TEST	0.00	22:27



SOLUTION CHANGE


CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 02/29/2024

OPERATOR NAME:
CASEY PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:04
INTERNAL STANDARD	VERIFIED	22:04
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:04
BLANK TEST	0.00	22:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:05
BLANK TEST	0.00	22:06
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:06
BLANK TEST	0.00	22:07



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 03/05/2024

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:07
INTERNAL STANDARD	VERIFIED	22:07
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:08
BLANK TEST	0.00	22:08
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:09
BLANK TEST	0.00	22:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:10
BLANK TEST	0.00	22:11



SUPERVISOR TEST



 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 03/10/2024

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:33
INTERNAL STANDARD	VERIFIED	21:33
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:34
BLANK TEST	0.00	21:34
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:35
BLANK TEST	0.00	21:35
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:36
BLANK TEST	0.00	21:37



SUPERVISOR TEST




CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 03/16/2024

OPERATOR NAME:
CASEY PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:02
INTERNAL STANDARD	VERIFIED	22:02
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:02
BLANK TEST	0.00	22:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:03
BLANK TEST	0.00	22:04
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:05
BLANK TEST	0.00	22:05



SUPERVISOR TEST


 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 03/22/2024

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:42
INTERNAL STANDARD	VERIFIED	21:42
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:42
BLANK TEST	0.00	21:43
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:43
BLANK TEST	0.00	21:44
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:45
BLANK TEST	0.00	21:45



SUPERVISOR TEST



 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 03/28/2024

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:59
INTERNAL STANDARD	VERIFIED	21:59
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:59
BLANK TEST	0.00	22:00
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:00
BLANK TEST	0.00	22:01
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:02
BLANK TEST	0.00	22:02



SOLUTION CHANGE


CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 03/31/2024

OPERATOR NAME:
CASEY PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:10
INTERNAL STANDARD	VERIFIED	22:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:11
BLANK TEST	0.00	22:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:12
BLANK TEST	0.00	22:13
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:13
BLANK TEST	0.00	22:14



SUPERVISOR TEST




CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 04/06/2024

OPERATOR NAME:
CASEY PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:28
INTERNAL STANDARD	VERIFIED	05:29
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:29
BLANK TEST	0.00	05:30
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:30
BLANK TEST	0.00	05:31
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:31
BLANK TEST	0.00	05:32



SUPERVISOR TEST


CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 04/09/2024

OPERATOR NAME:
CASEY PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:39
INTERNAL STANDARD	VERIFIED	21:39
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:40
BLANK TEST	0.00	21:41
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:41
BLANK TEST	0.00	21:42
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:42
BLANK TEST	0.00	21:43



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 04/19/2024

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:04
INTERNAL STANDARD	VERIFIED	06:04
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:05
BLANK TEST	0.00	06:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:06
BLANK TEST	0.00	06:06
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:07
BLANK TEST	0.00	06:08



SUPERVISOR TEST




CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 04/27/2024

OPERATOR NAME:
CASEY PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:02
INTERNAL STANDARD	VERIFIED	22:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:03
BLANK TEST	0.00	22:04
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:04
BLANK TEST	0.00	22:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:05
BLANK TEST	0.00	22:06



SOLUTION CHANGE




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 05/03/2024

OPERATOR NAME:
 JAMES J SNELL
 Permit #: 38312
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:59
INTERNAL STANDARD	VERIFIED	15:59
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	16:00
BLANK TEST	0.00	16:00
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	16:01
BLANK TEST	0.00	16:01
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	16:02
BLANK TEST	0.00	16:03



SUPERVISOR TEST




CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 05/09/2024

OPERATOR NAME:
CASEY PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:45
INTERNAL STANDARD	VERIFIED	23:45
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	23:45
BLANK TEST	0.00	23:46
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	23:46
BLANK TEST	0.00	23:47
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	23:47
BLANK TEST	0.00	23:48



SUPERVISOR TEST


CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 05/15/2024

OPERATOR NAME:
CASEY PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:24
INTERNAL STANDARD	VERIFIED	22:24
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:25
BLANK TEST	0.00	22:25
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:26
BLANK TEST	0.00	22:26
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:27
BLANK TEST	0.00	22:28



SUPERVISOR TEST




CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 05/22/2024

OPERATOR NAME:
CASEY PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:13
INTERNAL STANDARD	VERIFIED	05:13
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:14
BLANK TEST	0.00	05:15
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:15
BLANK TEST	0.00	05:16
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:16
BLANK TEST	0.00	05:17



SUPERVISOR TEST




CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 05/27/2024

OPERATOR NAME:
CASEY PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:06
INTERNAL STANDARD	VERIFIED	22:06
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:06
BLANK TEST	0.00	22:07
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:08
BLANK TEST	0.00	22:08
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:09
BLANK TEST	0.00	22:09



SOLUTION CHANGE


 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 06/02/2024

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:02
INTERNAL STANDARD	VERIFIED	22:02
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:03
BLANK TEST	0.00	22:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:04
BLANK TEST	0.00	22:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:05
BLANK TEST	0.00	22:06



SUPERVISOR TEST



 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 06/08/2024

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:27
INTERNAL STANDARD	VERIFIED	13:27
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:27
BLANK TEST	0.00	13:28
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	13:28
BLANK TEST	0.00	13:29
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	13:29
BLANK TEST	0.00	13:30



SUPERVISOR TEST




CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 06/14/2024

OPERATOR NAME:
CASEY PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:11
INTERNAL STANDARD	VERIFIED	22:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:12
BLANK TEST	0.00	22:12
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:13
BLANK TEST	0.00	22:14
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:14
BLANK TEST	0.00	22:15



SUPERVISOR TEST


 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 06/25/2024

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:39
INTERNAL STANDARD	VERIFIED	05:39
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:39
BLANK TEST	0.00	05:40
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:41
BLANK TEST	0.00	05:41
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	05:42
BLANK TEST	0.00	05:42



SOLUTION CHANGE



CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 06/30/2024

OPERATOR NAME:
CASEY PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:24
INTERNAL STANDARD	VERIFIED	06:25
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:25
BLANK TEST	0.00	06:26
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:26
BLANK TEST	0.00	06:27
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:27
BLANK TEST	0.00	06:28



SUPERVISOR TEST




CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 07/05/2024

OPERATOR NAME:
CASEY PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:17
INTERNAL STANDARD	VERIFIED	06:17
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:17
BLANK TEST	0.00	06:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:18
BLANK TEST	0.00	06:19
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	06:19
BLANK TEST	0.00	06:20



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 07/08/2024

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:08
INTERNAL STANDARD	VERIFIED	22:08
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:09
BLANK TEST	0.00	22:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:10
BLANK TEST	0.00	22:10
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:11
BLANK TEST	0.00	22:12



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 07/09/2024

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:12
INTERNAL STANDARD	VERIFIED	01:12
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:12
BLANK TEST	0.00	01:13
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:13
BLANK TEST	0.00	01:14
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:14
BLANK TEST	0.00	01:15



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 07/14/2024

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:32
INTERNAL STANDARD	VERIFIED	21:32
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:32
BLANK TEST	0.00	21:33
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:33
BLANK TEST	0.00	21:34
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:34
BLANK TEST	0.00	21:35



SUPERVISOR TEST


CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 07/20/2024

OPERATOR NAME:
CASEY PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:50
INTERNAL STANDARD	VERIFIED	23:50
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:50
BLANK TEST	0.00	23:51
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:51
BLANK TEST	0.00	23:52
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:52
BLANK TEST	0.00	23:53



SUPERVISOR TEST


CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 07/26/2024

OPERATOR NAME:
CASEY PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:35
INTERNAL STANDARD	VERIFIED	21:35
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	21:35
BLANK TEST	0.00	21:36
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	21:36
BLANK TEST	0.00	21:37
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	21:38
BLANK TEST	0.00	21:38



SOLUTION CHANGE




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 08/02/2024

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:08
INTERNAL STANDARD	VERIFIED	02:08
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	02:08
BLANK TEST	0.00	02:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	02:10
BLANK TEST	0.00	02:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	02:11
BLANK TEST	0.00	02:11



SUPERVISOR TEST


CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 08/08/2024

OPERATOR NAME:
CASEY PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:34
INTERNAL STANDARD	VERIFIED	21:34
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:35
BLANK TEST	0.00	21:35
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:36
BLANK TEST	0.00	21:37
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:37
BLANK TEST	0.00	21:38



SUPERVISOR TEST


 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 08/13/2024

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:13
INTERNAL STANDARD	VERIFIED	22:13
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:14
BLANK TEST	0.00	22:15
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:15
BLANK TEST	0.00	22:16
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:16
BLANK TEST	0.00	22:17



SUPERVISOR TEST




CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 08/20/2024

OPERATOR NAME:
CASEY PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:36
INTERNAL STANDARD	VERIFIED	05:36
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	05:36
BLANK TEST	0.00	05:37
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	05:37
BLANK TEST	0.00	05:38
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	05:38
BLANK TEST	0.00	05:39



SUPERVISOR TEST




CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 08/26/2024

OPERATOR NAME:
CASEY PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:10
INTERNAL STANDARD	VERIFIED	23:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:10
BLANK TEST	0.00	23:11
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	23:11
BLANK TEST	0.00	23:12
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	23:13
BLANK TEST	0.00	23:13



SOLUTION CHANGE




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 08/31/2024

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:07
INTERNAL STANDARD	VERIFIED	22:07
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:08
BLANK TEST	0.00	22:08
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:09
BLANK TEST	0.00	22:10
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:10
BLANK TEST	0.00	22:11



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 09/06/2024

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:03
INTERNAL STANDARD	VERIFIED	22:03
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:04
BLANK TEST	0.00	22:04
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:05
BLANK TEST	0.00	22:06
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:06
BLANK TEST	0.00	22:07



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 09/11/2024

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:51
INTERNAL STANDARD	VERIFIED	02:51
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	02:52
BLANK TEST	0.00	02:52
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	02:53
BLANK TEST	0.00	02:54
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	02:54
BLANK TEST	0.00	02:55



SUPERVISOR TEST




CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 09/18/2024

OPERATOR NAME:
CASEY PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:05
INTERNAL STANDARD	VERIFIED	22:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:05
BLANK TEST	0.00	22:06
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:06
BLANK TEST	0.00	22:07
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:07
BLANK TEST	0.00	22:08



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 09/24/2024

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:37
INTERNAL STANDARD	VERIFIED	22:37
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	22:37
BLANK TEST	0.00	22:38
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:38
BLANK TEST	0.00	22:39
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:39
BLANK TEST	0.00	22:40



SOLUTION CHANGE




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 09/30/2024

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:30
INTERNAL STANDARD	VERIFIED	22:30
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:30
BLANK TEST	0.00	22:31
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:32
BLANK TEST	0.00	22:32
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:33
BLANK TEST	0.00	22:33



SUPERVISOR TEST



 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 10/06/2024

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:07
INTERNAL STANDARD	VERIFIED	22:08
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:08
BLANK TEST	0.00	22:09
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:09
BLANK TEST	0.00	22:10
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:10
BLANK TEST	0.00	22:11



SUPERVISOR TEST


CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 10/12/2024

OPERATOR NAME:
CASEY PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:03
INTERNAL STANDARD	VERIFIED	22:03
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:03
BLANK TEST	0.00	22:04
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:05
BLANK TEST	0.00	22:05
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:06
BLANK TEST	0.00	22:07



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 10/18/2024

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:15
INTERNAL STANDARD	VERIFIED	22:15
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:15
BLANK TEST	0.00	22:16
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:16
BLANK TEST	0.00	22:17
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:17
BLANK TEST	0.00	22:18



SUPERVISOR TEST




CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 10/24/2024

OPERATOR NAME:
CASEY PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:10
INTERNAL STANDARD	VERIFIED	22:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:10
BLANK TEST	0.00	22:11
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:11
BLANK TEST	0.00	22:12
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:12
BLANK TEST	0.00	22:13



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 10/28/2024

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:28
INTERNAL STANDARD	VERIFIED	21:28
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	21:29
BLANK TEST	0.00	21:30
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	21:30
BLANK TEST	0.00	21:31
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:31
BLANK TEST	0.00	21:32



SUPERVISOR TEST




CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 11/06/2024

OPERATOR NAME:
CASEY PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:01
INTERNAL STANDARD	VERIFIED	22:02
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:02
BLANK TEST	0.00	22:03
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:03
BLANK TEST	0.00	22:04
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:04
BLANK TEST	0.00	22:05



SOLUTION CHANGE




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 11/01/2024

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:00
INTERNAL STANDARD	VERIFIED	01:00
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:00
BLANK TEST	0.00	01:01
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	01:01
BLANK TEST	0.00	01:02
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	01:02
BLANK TEST	0.00	01:03



SUPERVISOR TEST


CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 11/11/2024

OPERATOR NAME:
CASEY PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:41
INTERNAL STANDARD	VERIFIED	21:41
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	21:42
BLANK TEST	0.00	21:43
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	21:43
BLANK TEST	0.00	21:44
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	21:44
BLANK TEST	0.00	21:45



SUPERVISOR TEST




CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 11/17/2024

OPERATOR NAME:
CASEY PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:58
INTERNAL STANDARD	VERIFIED	22:58
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:58
BLANK TEST	0.00	22:59
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:59
BLANK TEST	0.00	23:00
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	23:00
BLANK TEST	0.00	23:01



SUPERVISOR TEST



CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 11/23/2024

OPERATOR NAME:
CASEY PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:07
INTERNAL STANDARD	VERIFIED	22:07
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:07
BLANK TEST	0.00	22:08
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:09
BLANK TEST	0.00	22:09
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:10
BLANK TEST	0.00	22:10



SUPERVISOR TEST


CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 11/29/2024

OPERATOR NAME:
CASEY PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:30
INTERNAL STANDARD	VERIFIED	22:30
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:31
BLANK TEST	0.00	22:32
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:32
BLANK TEST	0.00	22:33
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:33
BLANK TEST	0.00	22:34



SOLUTION CHANGE


CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 11/30/2024

OPERATOR NAME:
CASEY PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:13
INTERNAL STANDARD	VERIFIED	23:13
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	23:13
BLANK TEST	0.00	23:14
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	23:14
BLANK TEST	0.00	23:15
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	23:15
BLANK TEST	0.00	23:16



SUPERVISOR TEST


CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 12/05/2024

OPERATOR NAME:
CASEY PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:01
INTERNAL STANDARD	VERIFIED	22:01
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:02
BLANK TEST	0.00	22:02
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:03
BLANK TEST	0.00	22:03
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:04
BLANK TEST	0.00	22:05



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 12/10/2024

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:01
INTERNAL STANDARD	VERIFIED	22:01
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:01
BLANK TEST	0.00	22:02
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:02
BLANK TEST	0.00	22:03
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:04
BLANK TEST	0.00	22:04



SUPERVISOR TEST




CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 12/17/2024

OPERATOR NAME:
CASEY PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:43
INTERNAL STANDARD	VERIFIED	21:43
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	21:43
BLANK TEST	0.00	21:44
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	21:44
BLANK TEST	0.00	21:45
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	21:45
BLANK TEST	0.00	21:46



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 12/23/2024

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:07
INTERNAL STANDARD	VERIFIED	22:08
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	22:08
BLANK TEST	0.00	22:09
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	22:09
BLANK TEST	0.00	22:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:10
BLANK TEST	0.00	22:11



SOLUTION CHANGE


 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 12/31/2024

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:26
INTERNAL STANDARD	VERIFIED	22:26
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:27
BLANK TEST	0.00	22:27
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:28
BLANK TEST	0.00	22:29
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:29
BLANK TEST	0.00	22:30



SUPERVISOR TEST



 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 01/08/2025

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:20
INTERNAL STANDARD	VERIFIED	01:20
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:20
BLANK TEST	0.00	01:21
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:21
BLANK TEST	0.00	01:22
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:23
BLANK TEST	0.00	01:23



SUPERVISOR TEST



 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 01/12/2025

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:36
INTERNAL STANDARD	VERIFIED	02:36
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	02:36
BLANK TEST	0.00	02:37
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	02:38
BLANK TEST	0.00	02:38
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	02:39
BLANK TEST	0.00	02:39



SUPERVISOR TEST


CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 01/16/2025

OPERATOR NAME:
CASEY PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:04
INTERNAL STANDARD	VERIFIED	22:04
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:04
BLANK TEST	0.00	22:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:05
BLANK TEST	0.00	22:06
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:06
BLANK TEST	0.00	22:07



SUPERVISOR TEST


 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 01/23/2025

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:28
INTERNAL STANDARD	VERIFIED	05:28
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:28
BLANK TEST	0.00	05:29
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:29
BLANK TEST	0.00	05:30
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:30
BLANK TEST	0.00	05:31



SUPERVISOR TEST


 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 01/29/2025

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:40
INTERNAL STANDARD	VERIFIED	21:40
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	21:40
BLANK TEST	0.00	21:41
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	21:41
BLANK TEST	0.00	21:42
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:42
BLANK TEST	0.00	21:43



SOLUTION CHANGE



CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 01/31/2025

OPERATOR NAME:
CASEY PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:07
INTERNAL STANDARD	VERIFIED	22:08
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:08
BLANK TEST	0.00	22:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:09
BLANK TEST	0.00	22:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:10
BLANK TEST	0.00	22:11



SUPERVISOR TEST


CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 02/05/2025

OPERATOR NAME:
CASEYU PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:40
INTERNAL STANDARD	VERIFIED	21:40
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:40
BLANK TEST	0.00	21:41
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:41
BLANK TEST	0.00	21:42
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:42
BLANK TEST	0.00	21:43



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 02/09/2025

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:13
INTERNAL STANDARD	VERIFIED	22:13
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:13
BLANK TEST	0.00	22:14
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:14
BLANK TEST	0.00	22:15
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:15
BLANK TEST	0.00	22:16



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 02/17/2025

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:44
INTERNAL STANDARD	VERIFIED	21:44
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:44
BLANK TEST	0.00	21:45
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:45
BLANK TEST	0.00	21:46
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:46
BLANK TEST	0.00	21:47



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 02/21/2025

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:45
INTERNAL STANDARD	VERIFIED	21:45
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:45
BLANK TEST	0.00	21:46
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:46
BLANK TEST	0.00	21:47
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:47
BLANK TEST	0.00	21:48



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 02/27/2025

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:09
INTERNAL STANDARD	VERIFIED	22:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:09
BLANK TEST	0.00	22:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:10
BLANK TEST	0.00	22:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:11
BLANK TEST	0.00	22:12



SOLUTION CHANGE




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 03/01/2025

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:21
INTERNAL STANDARD	VERIFIED	23:22
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	23:22
BLANK TEST	0.00	23:23
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	23:23
BLANK TEST	0.00	23:24
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	23:24
BLANK TEST	0.00	23:25



SUPERVISOR TEST



 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 03/05/2025

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:08
INTERNAL STANDARD	VERIFIED	22:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:08
BLANK TEST	0.00	22:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:09
BLANK TEST	0.00	22:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:11
BLANK TEST	0.00	22:11



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 03/11/2025

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:37
INTERNAL STANDARD	VERIFIED	21:37
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:37
BLANK TEST	0.00	21:38
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:38
BLANK TEST	0.00	21:39
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:39
BLANK TEST	0.00	21:40



SUPERVISOR TEST


 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 03/17/2025

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:01
INTERNAL STANDARD	VERIFIED	22:01
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:01
BLANK TEST	0.00	22:02
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:02
BLANK TEST	0.00	22:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:03
BLANK TEST	0.00	22:04



SUPERVISOR TEST


 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 03/23/2025

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:43
INTERNAL STANDARD	VERIFIED	21:43
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:44
BLANK TEST	0.00	21:44
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:45
BLANK TEST	0.00	21:45
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:46
BLANK TEST	0.00	21:47



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 03/30/2025

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:57
INTERNAL STANDARD	VERIFIED	21:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	21:57
BLANK TEST	0.00	21:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	21:58
BLANK TEST	0.00	21:59
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:59
BLANK TEST	0.00	22:00



SOLUTION CHANGE




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 04/01/2025

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:31
INTERNAL STANDARD	VERIFIED	22:31
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:31
BLANK TEST	0.00	22:32
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:32
BLANK TEST	0.00	22:33
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:33
BLANK TEST	0.00	22:34



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 04/05/2025

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:03
INTERNAL STANDARD	VERIFIED	22:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:03
BLANK TEST	0.00	22:04
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:04
BLANK TEST	0.00	22:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:06
BLANK TEST	0.00	22:06



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 04/10/2025

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:37
INTERNAL STANDARD	VERIFIED	21:37
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:37
BLANK TEST	0.00	21:38
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:38
BLANK TEST	0.00	21:39
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:40
BLANK TEST	0.00	21:40



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 04/16/2025

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:36
INTERNAL STANDARD	VERIFIED	21:36
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:36
BLANK TEST	0.00	21:37
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:37
BLANK TEST	0.00	21:38
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:38
BLANK TEST	0.00	21:39



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 04/23/2025

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:29
INTERNAL STANDARD	VERIFIED	00:30
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:30
BLANK TEST	0.00	00:31
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:31
BLANK TEST	0.00	00:32
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:32
BLANK TEST	0.00	00:33



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 04/29/2025

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:04
INTERNAL STANDARD	VERIFIED	04:04
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	04:05
BLANK TEST	0.00	04:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	04:06
BLANK TEST	0.00	04:07
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	04:07
BLANK TEST	0.00	04:08



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 05/04/2025

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:32
INTERNAL STANDARD	VERIFIED	21:32
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:32
BLANK TEST	0.00	21:33
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:33
BLANK TEST	0.00	21:34
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:34
BLANK TEST	0.00	21:35



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 05/10/2025

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:04
INTERNAL STANDARD	VERIFIED	22:04
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:04
BLANK TEST	0.00	22:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:05
BLANK TEST	0.00	22:06
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:06
BLANK TEST	0.00	22:07



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 05/16/2025

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:29
INTERNAL STANDARD	VERIFIED	22:30
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:30
BLANK TEST	0.00	22:31
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:31
BLANK TEST	0.00	22:32
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:32
BLANK TEST	0.00	22:33



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 05/22/2025

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:14
INTERNAL STANDARD	VERIFIED	22:14
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:14
BLANK TEST	0.00	22:15
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:15
BLANK TEST	0.00	22:16
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:16
BLANK TEST	0.00	22:17



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 05/28/2025

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:41
INTERNAL STANDARD	VERIFIED	21:41
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	21:41
BLANK TEST	0.00	21:42
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	21:42
BLANK TEST	0.00	21:43
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:44
BLANK TEST	0.00	21:44



SOLUTION CHANGE




CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 05/31/2025

OPERATOR NAME:
CASEY PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:04
INTERNAL STANDARD	VERIFIED	22:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:05
BLANK TEST	0.00	22:06
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:06
BLANK TEST	0.00	22:07
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:07
BLANK TEST	0.00	22:08



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 06/09/2025

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:38
INTERNAL STANDARD	VERIFIED	21:39
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:39
BLANK TEST	0.00	21:40
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:40
BLANK TEST	0.00	21:41
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:41
BLANK TEST	0.00	21:42



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 06/15/2025

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:44
INTERNAL STANDARD	VERIFIED	21:44
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:45
BLANK TEST	0.00	21:45
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:46
BLANK TEST	0.00	21:47
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:47
BLANK TEST	0.00	21:48



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 06/20/2025

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:04
INTERNAL STANDARD	VERIFIED	22:04
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:04
BLANK TEST	0.00	22:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:05
BLANK TEST	0.00	22:06
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:06
BLANK TEST	0.00	22:07



SUPERVISOR TEST




CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 06/25/2025

OPERATOR NAME:
CASEY PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:41
INTERNAL STANDARD	VERIFIED	02:41
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:41
BLANK TEST	0.00	02:42
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	02:42
BLANK TEST	0.00	02:43
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	02:43
BLANK TEST	0.00	02:44



SOLUTION CHANGE




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 06/30/2025

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:43
INTERNAL STANDARD	VERIFIED	01:43
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:43
BLANK TEST	0.00	01:44
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:44
BLANK TEST	0.00	01:45
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:45
BLANK TEST	0.00	01:46



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 07/05/2025

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:55
INTERNAL STANDARD	VERIFIED	06:55
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:56
BLANK TEST	0.00	06:56
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:57
BLANK TEST	0.00	06:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:58
BLANK TEST	0.00	06:59



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 07/10/2025

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:45
INTERNAL STANDARD	VERIFIED	21:45
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:45
BLANK TEST	0.00	21:46
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:46
BLANK TEST	0.00	21:47
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:48
BLANK TEST	0.00	21:48



SUPERVISOR TEST




CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 07/15/2025

OPERATOR NAME:
CASEY PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:39
INTERNAL STANDARD	VERIFIED	21:39
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:39
BLANK TEST	0.00	21:40
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:40
BLANK TEST	0.00	21:41
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:42
BLANK TEST	0.00	21:42



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 07/27/2025

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:39
INTERNAL STANDARD	VERIFIED	21:39
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	21:39
BLANK TEST	0.00	21:40
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	21:40
BLANK TEST	0.00	21:41
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	21:41
BLANK TEST	0.00	21:42



SOLUTION CHANGE




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 07/30/2025

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:52
INTERNAL STANDARD	VERIFIED	22:53
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:53
BLANK TEST	0.00	22:54
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:54
BLANK TEST	0.00	22:55
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:55
BLANK TEST	0.00	22:56



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 08/06/2025

OPERATOR NAME:
 JAMES J SNELL
 Permit #: 38312
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:08
INTERNAL STANDARD	VERIFIED	06:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:09
BLANK TEST	0.00	06:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:10
BLANK TEST	0.00	06:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:11
BLANK TEST	0.00	06:12



SUPERVISOR TEST




CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 08/11/2025

OPERATOR NAME:
CASEY PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:18
INTERNAL STANDARD	VERIFIED	21:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:18
BLANK TEST	0.00	21:19
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	21:20
BLANK TEST	0.00	21:20
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:21
BLANK TEST	0.00	21:21



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 08/18/2025

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:46
INTERNAL STANDARD	VERIFIED	05:47
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:47
BLANK TEST	0.00	05:48
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:48
BLANK TEST	0.00	05:49
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	05:49
BLANK TEST	0.00	05:50



SUPERVISOR TEST


CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 08/23/2025

OPERATOR NAME:
CASEY PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:49
INTERNAL STANDARD	VERIFIED	23:49
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	23:49
BLANK TEST	0.00	23:50
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	23:50
BLANK TEST	0.00	23:51
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	23:51
BLANK TEST	0.00	23:52



SOLUTION CHANGE




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 09/01/2025

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:57
INTERNAL STANDARD	VERIFIED	07:57
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:57
BLANK TEST	0.00	07:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:58
BLANK TEST	0.00	07:59
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:00
BLANK TEST	0.00	08:00



SUPERVISOR TEST




CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 09/08/2025

OPERATOR NAME:
CASEY PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:30
INTERNAL STANDARD	VERIFIED	05:30
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:30
BLANK TEST	0.00	05:31
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:31
BLANK TEST	0.00	05:32
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:32
BLANK TEST	0.00	05:33



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 09/16/2025

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:41
INTERNAL STANDARD	VERIFIED	17:41
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	17:41
BLANK TEST	0.00	17:42
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	17:42
BLANK TEST	0.00	17:43
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	17:43
BLANK TEST	0.00	17:44



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 09/25/2025

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:00
INTERNAL STANDARD	VERIFIED	22:00
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:01
BLANK TEST	0.00	22:02
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:02
BLANK TEST	0.00	22:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:03
BLANK TEST	0.00	22:04



SOLUTION CHANGE




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 10/01/2025

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:41
INTERNAL STANDARD	VERIFIED	21:41
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:42
BLANK TEST	0.00	21:42
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:43
BLANK TEST	0.00	21:44
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:44
BLANK TEST	0.00	21:45



SUPERVISOR TEST




CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 10/11/2025

OPERATOR NAME:
CASEY PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:05
INTERNAL STANDARD	VERIFIED	06:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:05
BLANK TEST	0.00	06:06
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:06
BLANK TEST	0.00	06:07
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:07
BLANK TEST	0.00	06:08



SUPERVISOR TEST



 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 10/17/2025

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:04
INTERNAL STANDARD	VERIFIED	06:04
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:04
BLANK TEST	0.00	06:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:05
BLANK TEST	0.00	06:06
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:06
BLANK TEST	0.00	06:07



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 10/23/2025

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:03
INTERNAL STANDARD	VERIFIED	06:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:04
BLANK TEST	0.00	06:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:05
BLANK TEST	0.00	06:06
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:06
BLANK TEST	0.00	06:07



SUPERVISOR TEST




CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 10/29/2025

OPERATOR NAME:
CASEY PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:49
INTERNAL STANDARD	VERIFIED	08:49
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:49
BLANK TEST	0.00	08:50
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:50
BLANK TEST	0.00	08:51
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:51
BLANK TEST	0.00	08:52



SOLUTION CHANGE




CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 11/01/2025

OPERATOR NAME:
CASEY PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:21
INTERNAL STANDARD	VERIFIED	06:21
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	06:21
BLANK TEST	0.00	06:22
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:22
BLANK TEST	0.00	06:23
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	06:23
BLANK TEST	0.00	06:24



SUPERVISOR TEST


CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 11/10/2025

OPERATOR NAME:
CASEY PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:18
INTERNAL STANDARD	VERIFIED	09:19
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	09:19
BLANK TEST	0.00	09:20
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	09:20
BLANK TEST	0.00	09:21
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	09:21
BLANK TEST	0.00	09:22



SUPERVISOR TEST



CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 11/16/2025

OPERATOR NAME:
CASEY PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:05
INTERNAL STANDARD	VERIFIED	06:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:06
BLANK TEST	0.00	06:06
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	06:07
BLANK TEST	0.00	06:08
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	06:08
BLANK TEST	0.00	06:09



SUPERVISOR TEST


CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 11/22/2025

OPERATOR NAME:
CASEY PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:41
INTERNAL STANDARD	VERIFIED	05:41
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	05:41
BLANK TEST	0.00	05:42
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	05:42
BLANK TEST	0.00	05:43
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	05:43
BLANK TEST	0.00	05:44



SOLUTION CHANGE




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 12/01/2025

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:03
INTERNAL STANDARD	VERIFIED	06:03
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	06:03
BLANK TEST	0.00	06:04
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:05
BLANK TEST	0.00	06:05
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	06:06
BLANK TEST	0.00	06:06



SUPERVISOR TEST




CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 12/06/2025

OPERATOR NAME:
CASEY PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:23
INTERNAL STANDARD	VERIFIED	06:24
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	06:24
BLANK TEST	0.00	06:25
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	06:25
BLANK TEST	0.00	06:26
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:26
BLANK TEST	0.00	06:27



SUPERVISOR TEST




CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 12/17/2025

OPERATOR NAME:
CASEY PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:53
INTERNAL STANDARD	VERIFIED	05:53
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:53
BLANK TEST	0.00	05:54
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:54
BLANK TEST	0.00	05:55
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:56
BLANK TEST	0.00	05:56



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 12/24/2025

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:24
INTERNAL STANDARD	VERIFIED	01:24
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:24
BLANK TEST	0.00	01:25
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:25
BLANK TEST	0.00	01:26
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:26
BLANK TEST	0.00	01:27



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 12/29/2025

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:38
INTERNAL STANDARD	VERIFIED	05:38
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:39
BLANK TEST	0.00	05:39
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:40
BLANK TEST	0.00	05:41
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:41
BLANK TEST	0.00	05:42



SOLUTION CHANGE




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 01/03/2026

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:23
INTERNAL STANDARD	VERIFIED	14:23
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	14:23
BLANK TEST	0.00	14:24
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	14:25
BLANK TEST	0.00	14:25
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	14:26
BLANK TEST	0.00	14:27



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 01/10/2026

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:17
INTERNAL STANDARD	VERIFIED	16:17
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	16:18
BLANK TEST	0.00	16:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	16:19
BLANK TEST	0.00	16:20
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	16:20
BLANK TEST	0.00	16:21



SUPERVISOR TEST




CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 01/23/2026

OPERATOR NAME:
A H SMITH
Permit #: 44366
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:18
INTERNAL STANDARD	VERIFIED	22:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:18
BLANK TEST	0.00	22:19
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:19
BLANK TEST	0.00	22:20
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:20
BLANK TEST	0.00	22:21



[Handwritten signature]

SOLUTION CHANGE




CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 02/02/2026

OPERATOR NAME:
CASEY PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:29
INTERNAL STANDARD	VERIFIED	14:29
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	14:30
BLANK TEST	0.00	14:31
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	14:31
BLANK TEST	0.00	14:32
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	14:32
BLANK TEST	0.00	14:33



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 02/08/2026

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:18
INTERNAL STANDARD	VERIFIED	16:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	16:19
BLANK TEST	0.00	16:19
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	16:20
BLANK TEST	0.00	16:21
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	16:21
BLANK TEST	0.00	16:22



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 02/16/2026

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:09
INTERNAL STANDARD	VERIFIED	14:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	14:10
BLANK TEST	0.00	14:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	14:11
BLANK TEST	0.00	14:12
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	14:12
BLANK TEST	0.00	14:13



SUPERVISOR TEST




CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 02/23/2026

OPERATOR NAME:
CASEY PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:43
INTERNAL STANDARD	VERIFIED	14:43
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	14:43
BLANK TEST	0.00	14:44
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	14:44
BLANK TEST	0.00	14:45
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	14:45
BLANK TEST	0.00	14:46



SOLUTION CHANGE




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 03/01/2026

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:32
INTERNAL STANDARD	VERIFIED	19:32
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	19:33
BLANK TEST	0.00	19:34
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	19:34
BLANK TEST	0.00	19:35
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	19:35
BLANK TEST	0.00	19:36



SUPERVISOR TEST




CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 03/16/2026

OPERATOR NAME:
CASEY PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:44
INTERNAL STANDARD	VERIFIED	19:45
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	19:45
BLANK TEST	0.00	19:46
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	19:46
BLANK TEST	0.00	19:47
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	19:47
BLANK TEST	0.00	19:48



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 03/22/2026

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:05
INTERNAL STANDARD	VERIFIED	11:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	11:06
BLANK TEST	0.00	11:07
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	11:07
BLANK TEST	0.00	11:08
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	11:08
BLANK TEST	0.00	11:09



SOLUTION CHANGE




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 03/31/2026

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 26050

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:22
INTERNAL STANDARD	VERIFIED	20:22
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	20:23
BLANK TEST	0.00	20:23
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	20:24
BLANK TEST	0.00	20:25
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	20:25
BLANK TEST	0.00	20:26



SUPERVISOR TEST


 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 04/05/2026

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 26050

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:29
INTERNAL STANDARD	VERIFIED	17:29
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	17:29
BLANK TEST	0.00	17:30
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	17:30
BLANK TEST	0.00	17:31
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	17:31
BLANK TEST	0.00	17:32



SUPERVISOR TEST


 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 04/11/2026

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 26050

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:20
INTERNAL STANDARD	VERIFIED	18:20
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	18:21
BLANK TEST	0.00	18:21
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	18:22
BLANK TEST	0.00	18:22
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	18:23
BLANK TEST	0.00	18:24



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 04/16/2026

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 26050

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:54
INTERNAL STANDARD	VERIFIED	22:54
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:55
BLANK TEST	0.00	22:55
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:56
BLANK TEST	0.00	22:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:57
BLANK TEST	0.00	22:58



SUPERVISOR TEST



 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 04/27/2026

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 26050

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:16
INTERNAL STANDARD	VERIFIED	22:16
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:16
BLANK TEST	0.00	22:17
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:17
BLANK TEST	0.00	22:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:18
BLANK TEST	0.00	22:19



SOLUTION CHANGE



CICERO TOWN
POLICE DEPT
DATAMASTER dmt: 107808

Date: 04/30/2026

OPERATOR NAME:
CASEY PLACITO
Permit #: 47415
Operator Agency/Dept: CICERO TOWN PD
Reference Standard #: 26050

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:58
INTERNAL STANDARD	VERIFIED	21:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	21:58
BLANK TEST	0.00	21:59
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	21:59
BLANK TEST	0.00	22:00
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:00
BLANK TEST	0.00	22:01



SOLUTION CHANGE




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 05/13/2026

OPERATOR NAME:
 JAMES J SNELL
 Permit #: 38312
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 26050

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:34
INTERNAL STANDARD	VERIFIED	15:34
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	15:35
BLANK TEST	0.00	15:35
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	15:36
BLANK TEST	0.00	15:36
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	15:37
BLANK TEST	0.00	15:38



SUPERVISOR TEST



 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 05/21/2026

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 26050

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:00
INTERNAL STANDARD	VERIFIED	14:00
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	14:01
BLANK TEST	0.00	14:02
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	14:02
BLANK TEST	0.00	14:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	14:03
BLANK TEST	0.00	14:04



SOLUTION CHANGE


 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 05/31/2026

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 26120

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:49
INTERNAL STANDARD	VERIFIED	18:49
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	18:49
BLANK TEST	0.00	18:50
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	18:50
BLANK TEST	0.00	18:51
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	18:51
BLANK TEST	0.00	18:52



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 06/15/2026

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 26120

Nominal = 0.10 g/210L

BLANK TEST		0.00	14:01
INTERNAL STANDARD	VERIFIED		14:01
REFERENCE STANDARD	TEMP	34.02c	
REFERENCE STANDARD		0.09	14:01
BLANK TEST		0.00	14:02
REFERENCE STANDARD	TEMP	34.02c	
REFERENCE STANDARD		0.10	14:03
BLANK TEST		0.00	14:03
REFERENCE STANDARD	TEMP	34.02c	
REFERENCE STANDARD		0.10	14:04
BLANK TEST		0.00	14:05



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 06/21/2026

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 26120

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:24
INTERNAL STANDARD	VERIFIED	19:24
REFERENCE STANDARD	TEMP 34.02c	
REFERENCE STANDARD	0.09	19:24
BLANK TEST	0.00	19:25
REFERENCE STANDARD	TEMP 34.02c	
REFERENCE STANDARD	0.09	19:25
BLANK TEST	0.00	19:26
REFERENCE STANDARD	TEMP 34.02c	
REFERENCE STANDARD	0.09	19:26
BLANK TEST	0.00	19:27



SUPERVISOR TEST



 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 06/25/2026

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 26120

Nominal = 0.10 g/210L

BLANK TEST		0.00	15:42
INTERNAL STANDARD	VERIFIED		15:42
REFERENCE STANDARD	TEMP	34.02c	
REFERENCE STANDARD		0.09	15:43
BLANK TEST		0.00	15:44
REFERENCE STANDARD	TEMP	34.02c	
REFERENCE STANDARD		0.09	15:44
BLANK TEST		0.00	15:45
REFERENCE STANDARD	TEMP	34.02c	
REFERENCE STANDARD		0.09	15:45
BLANK TEST		0.00	15:46



SOLUTION CHANGE




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 07/03/2026

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 26120

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:09
INTERNAL STANDARD	VERIFIED	17:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	17:10
BLANK TEST	0.00	17:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	17:11
BLANK TEST	0.00	17:12
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	17:12
BLANK TEST	0.00	17:13



SUPERVISOR TEST




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 07/14/2026

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 26120

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:37
INTERNAL STANDARD	VERIFIED	13:37
REFERENCE STANDARD	TEMP 34.02c	
REFERENCE STANDARD	0.09	13:37
BLANK TEST	0.00	13:38
REFERENCE STANDARD	TEMP 34.02c	
REFERENCE STANDARD	0.09	13:38
BLANK TEST	0.00	13:39
REFERENCE STANDARD	TEMP 34.02c	
REFERENCE STANDARD	0.10	13:40
BLANK TEST	0.00	13:40



SOLUTION CHANGE




 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 08/02/2026

OPERATOR NAME:
 CASEY PLACITO
 Permit #: 47415
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 26190

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:52
INTERNAL STANDARD	VERIFIED	16:52
REFERENCE STANDARD	TEMP	34.02c
REFERENCE STANDARD	0.10	16:53
BLANK TEST	0.00	16:54
REFERENCE STANDARD	TEMP	34.02c
REFERENCE STANDARD	0.10	16:54
BLANK TEST	0.00	16:55
REFERENCE STANDARD	TEMP	34.02c
REFERENCE STANDARD	0.10	16:55
BLANK TEST	0.00	16:56



SUPERVISOR TEST

 CICERO TOWN
 POLICE DEPT
 DATAMASTER dmt: 107808

Date: 08/11/2026

OPERATOR NAME:
 KYLE A HARRINGTON
 Permit #: 48283
 Operator Agency/Dept: CICERO TOWN PD
 Reference Standard #: 26190

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:26
INTERNAL STANDARD	VERIFIED	08:26
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:26
BLANK TEST	0.00	08:27
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	08:28
BLANK TEST	0.00	08:28
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:29
BLANK TEST	0.00	08:30

