

SUPERVISOR TEST

-----  
CICERO TOWN  
POLICE DEPT  
DATAMASTER dmt: 107808  
-----

Date: 11/17/2024

OPERATOR NAME:  
CASEY PLACITO  
Permit #: 47415  
Operator Agency/Dept: CICERO TOWN PD  
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:58
INTERNAL STANDARD	VERIFIED	22:58
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:58
BLANK TEST	0.00	22:59
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:59
BLANK TEST	0.00	23:00
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	23:00
BLANK TEST	0.00	23:01

