



HiGrade Technik
Jonas Boehnke
4444 - 212 street
Langley, BC
Canada, V3A 5S8
www.higrade.tech

Calibration Report

Date: July 1, 2019	Project: Moosehorn
HiGrade Project #: 2019-DPI-001	Project Location: Alaska
Prepared by: Jonas Boehnke	Client:
Email: jonas@highgrade.tech	Email:
Phone: 1 604 3678221	Phone:

Scan Tool DPI 1901	Fire Assay Method FAS-415, 30g fire assay, grav. finish
Image Scan Settings 1200 dpi	Assay Operator MSA Labs
Image Adjust Settings Moosehorn	Certificate # YVR1910189
Image Processing Settings Moosehorn	
Data Output Settings Moosehorn	

Sample #	DPI [g/t]	Deviation [%]	Fire Assay [g/t]
1	0	0.0%	0
2	28	3.7%	27
3	237	3.9%	228
4	777	2.2%	760
5	541	-0.4%	543

Thickness Calibration

Particle Size [mesh]	Thickness Value [micron]
> 20 mesh	123
20-40 mesh	83
40-100 mesh	53
< 100 mesh	23

Summary

The optical DPI analysis correlates well with the industry standard fire assay. The deviation between the two methods is -0.4% to 3.9%. Five subsamples of table concentrate (approx 1g each) were analyzed with the non-destructive DPI tool, then blended with 24 g of clean silica sand and submitted for fire assays to MSALabs. Blending with silica sand homogenizes the sample and ensures that the maximum permissible fire assay grade (1000 g/t) is not exceeded. The sample contained no interfering minerals or metals, such as pyrite.

Reconciliation of Assay methods

