

CERTIFICATE OF GENERAL CALIBRATION

Quincy Lab Inc. hereby certifies that all of our Ovens, Incubators, and Warming instruments have been tested and calibrated using equipment with accuracies traceable to the National Institute of Standards and Technology (NIST). This is within the given limitations of the institute's calibration service or derived from accepted values of natural physical constants. Quincy Lab Inc. further certifies that all testing and test results were executed in accordance with, and conform to, the referenced tests and standards of general testing practices where applicable.

All systems receive a visual inspection prior to shipment to ensure proper assembly. They are also operationally tested for functionality, accuracy, and safety.

Performance Test:

Control temperature accuracy is calculated against a NIST (National Institute of Standards & Technology), traceable device, measured at the center of the chamber, to ensure it is within stated specified tolerances.



Quincy Lab, Inc.
109 Shore Drive
Burr Ridge, IL 60527