



grady.surgicalsolutions@gmail.com

Scanning Protocol for Surgical Guides

Intraoral Scan

- Both maxillary and mandibular full arch scans are needed for planning

CBCT Scan

- Complete maxillary and sinuses for maxillary arch guide planning
- Complete mandible for mandibular guides
- .2 to .4 mm slice thickness
- Patient in open position during scan – approximately 5mm in anterior