



Service
Maintenance &
Spares

Central Sterile Supply Department

ASNZS 4187:2014 Amd2:2019



Water Treatment

Water quality is of utmost importance for Centralised Sterile Supply Departments. Water treatment equipment must be carefully selected to meet the requirements of ASNZS 4187:2014 Amd2. These systems can vary in complexity requiring specialised understanding and support. Healthcare facilities rely on the quality of the water used in Reprocessing of Reusable Medical Devices (RMDs) so the quality of the services they receive must be of the highest standards.

Our Services



Parts and Consumables

- Water Softener Salt
- Ion Exchange Resins
- Reverse Osmosis (RO) Membranes
- Electro-De-Ionisation (EDI) cells
- Endotoxin Filters
- Ultraviolet (UV) Lamps and sleeves
- Cartridge Filters
- Activated Carbon



Routine Maintenance

- Calibration of instrumentation
- Replacement of filters, UV lamps and Endotoxin Filters
- Inhouse hardness, conductivity and pH testing



Water Testing

NATA Accredited Testing for.

- pH
- Conductivity
- Total Hardness
- Chloride
- Iron
- Phosphates
- Silicates
- Total Viable Counts
- Endotoxin



Breakdown Support

- Priority mobilisation to site to rectify issues
- Access to spares for returning the system back to service
- Afterhours calls and site visits

Call us for appointment and consultation (02) 8582 7081

