



Biological Sterilizer Monitoring Program
Test Results
Week 1 (01/01/2018) - Week 13 (04/01/2018)

Account Id: C-132560
Customer: Larry Goldstein, D.D.S.

2606 Abbey Ct
Alpharetta, GA 30004-6023

Sterilizer	Test #	Result Date	Result	Result Description
#01 STEAM AUTOCLAVE	16838641	3/30/2018	Pass	No growth of test spores, growth of control spores
	16838640	3/23/2018	Pass	No growth of test spores, growth of control spores
	16838639	3/16/2018	Pass	No growth of test spores, growth of control spores
	16838638	3/9/2018	Pass	No growth of test spores, growth of control spores
	16838637	2/22/2018	Pass	No growth of test spores, growth of control spores
	16838636	2/14/2018	Unable to Process 3	No test strip present
	16838635	2/8/2018	Pass	No growth of test spores, growth of control spores
	16838634	2/1/2018	Pass	No growth of test spores, growth of control spores
	16838633	1/26/2018	Pass	No growth of test spores, growth of control spores
	16838632	1/19/2018	Pass	No growth of test spores, growth of control spores
	16838631	1/15/2018	Pass	No growth of test spores, growth of control spores
	16838630	1/4/2018	Pass	No growth of test spores, growth of control spores
	16838629	1/2/2018	Pass	No growth of test spores, growth of control spores

Data displayed in this report is based on information provided to Mesa Labs by the client(s) at the time the tests are processed.