





THEIR PROTECTION  
YOUR ASSURANCE

ProEdgeDental.com | 888.843.3343 | 7042 S Revere Parkway #400, Centennial, CO 80112

## DENTAL WATER SAFETY

Waterline Test Submission Form

FILL OUT ENTIRE FORM COMPLETELY FOR EVERY TEST - FOLLOW TESTING INSTRUCTIONS CLOSELY

**Practice Information:**  New Customer  Returning Customer

Practice Name: Larry R Goldstein, DDS, PC Contact Person: Larry Goldstein

Practice Address: 2606 Abbey Court Contact Title/Position: owner

City, State & Zip: Alpharetta, GA 30004 Contact Email: blackbirdsix@gmail.com  
\* Test results will be sent to this email

Phone: (770) 475-9210 Fax: \_\_\_\_\_ Compliance Officer Email: blackbird6@pm.me  
\* Test results will be CCed to this email

### Waterline Treatment Protocol Information (Select All That Apply):

**A. TREATMENT PRODUCT(S) USED:**  
 BluTab  ICX  
 Citrisil/Citrisil Blue  DentaPure Straw/Cartridge  
 Steril Straw  Other: \_\_\_\_\_

**B. SHOCK PRODUCT(S) USED:**  
 Bleach  Citrisil Shock  
 Sterilex Ultra  Other: \_\_\_\_\_

**C. OTHER TREATMENT SYSTEMS:**  
 Steril System  O-So Pure  
 Vista System  
 Other: Received on

Install Date: \_\_\_\_\_ Date of Last Shock: 24 Dec 19  
 Shock Prior to Install:  Yes  No Never Shocked:

**D. USING WATER BOTTLES:**  Yes  No

**E. SOURCE WATER (We fill our water bottles from...):**  
 City Water/Tap  Other (e.g. Delivered): \_\_\_\_\_  
 Bottle Distilled  In-Office Filter (Distiller) or R/O Unit (Brand: Juflwater)

**JAN 03 2020**  
**by Madison**

### Test Sampling Information:

SAMPLING DATE: 30 Dec 19 2 JAN 20 DO NOT RETURN SHIP SAMPLES TO LAB ON FRIDAYS

What date did you take water samples from your office?

FOR LAB USE ONLY ProEdge Sample #	Vial Number	Location (Circle One)	Room/Chair/Operator #	Device or Source Water (Circle One)	Notes or Other Information
<u>59</u>	1	Room   <u>Chair</u>   Operator	<u>1</u>	AW Syringe   <u>Handpiece</u>   Scaler   Source Water	
<u>60</u>	2	Room   <u>Chair</u>   Operator	<u>2</u>	<u>AW Syringe</u>   Handpiece   Scaler   Source Water	
<u>61</u>	3	Room   <u>Chair</u>   Operator	<u>3</u>	<u>AW Syringe</u>   Handpiece   Scaler   Source Water	
<u>62</u>	4	Room   Chair   <u>Operator</u>	<u>3</u>	AW Syringe   Handpiece   <u>Scaler</u>   Source Water	<u>Caution</u>
	5	Room   Chair   Operator		AW Syringe   Handpiece   Scaler   Source Water	
	6	Room   Chair   Operator		AW Syringe   Handpiece   Scaler   Source Water	
	7	Room   Chair   Operator		AW Syringe   Handpiece   Scaler   Source Water	
	8	Room   Chair   Operator		AW Syringe   Handpiece   Scaler   Source Water	
	9	Room   Chair   Operator		AW Syringe   Handpiece   Scaler   Source Water	
	10	Room   Chair   Operator		AW Syringe   Handpiece   Scaler   Source Water	
	11	Room   Chair   Operator		AW Syringe   Handpiece   Scaler   Source Water	
	12	Room   Chair   Operator		AW Syringe   Handpiece   Scaler   Source Water	
	13	Room   Chair   Operator		AW Syringe   Handpiece   Scaler   Source Water	
	14	Room   Chair   Operator		AW Syringe   Handpiece   Scaler   Source Water	
	15	Room   Chair   Operator		AW Syringe   Handpiece   Scaler   Source Water	
	16	Room   Chair   Operator		AW Syringe   Handpiece   Scaler   Source Water	

The guidelines outlined by the Center for Disease Control (CDC) for bacteria in water used as a coolant/irrigant for non-surgical dental procedures should be as low as reasonably achievable, and, at a minimum  $\leq 500$  CFU/mL. Your practice's water samples will be tested against this CDC standard for dental water safety.  
DUWL75F04

**Questions?**  
Call us today at 888.843.3343

3924