

This report is generated for:  
Name : GOLDSTEIN, LARRY  
Address: 2606 ABBEY COURT  
ALPHARETTA, GA-30004

## Spore Test Results

Test Period :01/01/2020 - 12/31/2020  
Report Generated: 01/04/2021



Autoclave Name: Autoclave AD8379      Serial #: AD8379      Type: Steam

Test Done by	Test Date	Test Type	Result	Addressed On	Note
Unknown	12/17/2020	BIOlogical Mail-In	Pass		
Unknown	12/02/2020	BIOlogical Mail-In	Pass		
Unknown	11/19/2020	BIOlogical Mail-In	Pass		
Unknown	11/13/2020	BIOlogical Mail-In	Pass		
Unknown	11/05/2020	BIOlogical Mail-In	Pass		
Unknown	10/30/2020	BIOlogical Mail-In	Pass		
Unknown	10/29/2020	BIOlogical Mail-In	Pass		
Unknown	10/15/2020	BIOlogical Mail-In	Pass		
Unknown	10/09/2020	BIOlogical Mail-In	Pass		
Unknown	10/02/2020	BIOlogical Mail-In	Pass		
Unknown	09/29/2020	BIOlogical Mail-In	Pass		
Unknown	09/17/2020	BIOlogical Mail-In	Pass		



This report is generated for:  
**Name : GOLDSTEIN, LARRY**  
**Address: 2606 ABBEY COURT**  
**ALPHARETTA, GA-30004**

**Spore Test Results**  
 Test Period :01/01/2020 - 12/31/2020  
 Report Generated: 01/04/2021



Continued from previous page

**Autoclave Name: Autoclave AD8379      Serial #: AD8379      Type: Steam**

Test Done by	Test Date	Test Type	Result	Addressed On	Note
Unknown	09/15/2020	BIOlogical Mail-In	Pass		
Unknown	09/04/2020	BIOlogical Mail-In	Pass		
Unknown	08/26/2020	BIOlogical Mail-In	Pass		
Unknown	08/21/2020	BIOlogical Mail-In	Pass		
Unknown	07/31/2020	BIOlogical Mail-In	Pass		
Unknown	07/23/2020	BIOlogical Mail-In	Pass		
Unknown	07/16/2020	BIOlogical Mail-In	Pass		
Unknown	07/09/2020	BIOlogical Mail-In	Pass		
Unknown	07/03/2020	BIOlogical Mail-In	Pass		
Unknown	06/26/2020	BIOlogical Mail-In	Pass		
Unknown	06/19/2020	BIOlogical Mail-In	Pass		
Unknown	06/12/2020	BIOlogical Mail-In	Pass		



This report is generated for:  
**Name : GOLDSTEIN, LARRY**  
**Address: 2606 ABBEY COURT**  
**ALPHARETTA, GA-30004**

## Spore Test Results

Test Period :01/01/2020 - 12/31/2020  
 Report Generated: 01/04/2021



Continued from previous page

**Autoclave Name: Autoclave AD8379      Serial #: AD8379      Type: Steam**

Test Done by	Test Date	Test Type	Result	Addressed On	Note
Unknown	06/05/2020	BIOLogical Mail-In	Pass		
Unknown	05/29/2020	BIOLogical Mail-In	Pass		
Larry R. Goldstein, D.D.S.	05/21/2020	In-Office	Pass		
Larry R. Goldstein, D.D.S.	05/14/2020	In-Office	Pass		
Larry R. Goldstein, D.D.S.	05/07/2020	In-Office	Pass		
Larry R. Goldstein, D.D.S>	05/01/2020	In-Office	Pass		
Larry R. Goldstein, DDS	04/24/2020	In-Office	Pass		
Larry R. Goldstein, DDS	04/17/2020	In-Office	Pass		
Larry R. Goldstein, DDS	04/10/2020	In-Office	Pass		
Unknown	04/08/2020	BIOLogical Mail-In	Pass		
Unknown	03/31/2020	BIOLogical Mail-In	Pass		
Unknown	03/24/2020	BIOLogical Mail-In	Pass		



This report is generated for:  
Name : GOLDSTEIN, LARRY  
Address: 2606 ABBEY COURT  
ALPHARETTA, GA-30004

## Spore Test Results

Test Period :01/01/2020 - 12/31/2020  
Report Generated: 01/04/2021



Continued from previous page

Autoclave Name: Autoclave AD8379      Serial #: AD8379      Type: Steam

Test Done by	Test Date	Test Type	Result	Addressed On	Note
Unknown	03/19/2020	BIOlogical Mail-In	Pass		
Unknown	03/12/2020	BIOlogical Mail-In	Pass		
Unknown	03/04/2020	BIOlogical Mail-In	Pass		
Unknown	02/28/2020	BIOlogical Mail-In	Pass		
Unknown	02/15/2020	BIOlogical Mail-In	Pass		
Unknown	02/06/2020	BIOlogical Mail-In	Pass		
Unknown	01/23/2020	BIOlogical Mail-In	Pass		
Unknown	01/19/2020	BIOlogical Mail-In	Pass		
Unknown	01/09/2020	BIOlogical Mail-In	Pass		
Unknown	01/02/2020	BIOlogical Mail-In	Pass		
Unknown	01/02/2020	BIOlogical Mail-In	Pass		

