



ProEdgeDental.com | 888.843.3343 | 7042 S Revere Pkwy #400, Centennial, CO 80112

# R2A DENTAL WATER SAFETY REPORT

## Larry R. Goldstein DDS

Alpharetta, GA	Phone number 770-475-3210	Test date 12/30/2025	Report date 01/05/2026
----------------	------------------------------	-------------------------	---------------------------

ProEdge Dental Water Lab Technicians approve the following results accurate and complete utilizing gold-standard testing and laboratory methods.

  
 Colton Knepp  
 Quality Assurance





  
 Hillary Sinclair  
 Customer Delight

### Waterline Treatment Info

Source water Filter, Distiller, or R/O Unit	Shocks used Liquid Ultra (Pink Stuff)	Shock date 12/05/2025	Treatment used BluTab
--	--	--------------------------	--------------------------

### SAMPLES SUBMITTED

### LABORATORY TEST RESULTS (CFU/mL)

ProEdge sample #	Operatory #	Sample source	CDC compliance	Microbial count	0	5-200	201-500	501-2,500	2,501-10,000	10,001-40,000	40,001-100,000	TNTC
41	1	AW Syringe	Pass	0								
42	2	AW Syringe	Pass	0								
43	2	Highspeed handpiece	Pass	0								
44	3	AW Syringe	Pass	0								

---

QNS = Quantity not sufficient (not enough water present in vile to complete testing) PASS/FAIL results are based on the  $\leq 500$  CFU/mL standard set by the CDC. RED = Immediate shock and retest. YELLOW = Immediate shock. GREEN = Continue treatment protocol. Learn more on the second page of this lab report. ProEdge Dental Water Labs flow cytometry testing method has been validated by an independent study conducted by the University of Colorado. For more info, visit [ProEdgeDental.com/Flo](http://ProEdgeDental.com/Flo)