

From: Laboratory laboratory@biadiagnostics.com
 Subject: Your Test Results: From Bia Diagnostics
 Date: Jul 6, 2023 at 12:08:50 PM
 To: Kaitlyn Corbett kaitlyn@greenmontcraftcannabisco.com

Please note, terpene and pathogen results will be sent separately at a later date.



Please see attached Certificate of Analysis for your test results. Feel free to contact us with any questions.



Office: 802-540-0148 | Fax: 802-540-0147
 480 HERCULES DR. COLCHESTER, VT 05446

Certificate of Analysis


Company: Greenmont Craft Cannabis Co. 305 Rt 15 Jericho, VT 05465	Sample ID: Harvest Lot #1 Lot: N/A Matrix: Flower	Report Date: 7/6/2023 Date Analyzed: 6/28/2023 Analyst: 045 Report ID: C230627AL
Customer ID: 220215-1	Date Sampled: N/A	
Grower License #: SCLT0233	Date Received: 6/27/2023	

Pesticides/Mycotoxins Summary

Category II Residual Pesticide	LOQ (ppm)	Concentration (ppm)
Abamectin	0.0100	<LOQ
Acephate	0.0010	<LOQ
Acequinocyl	0.0010	<LOQ
Azoxystrobin	0.0010	<LOQ
Bifenazate	0.0010	<LOQ
Bifenthrin	0.0010	<LOQ
Carbaryl	0.0010	<LOQ
Cypermethrin	0.0100	<LOQ
Etoxazole	0.0010	<LOQ
Imidacloprid	0.0010	<LOQ
Myclobutanil	0.0010	<LOQ
Pyrethrin I	0.0010	<LOQ
Pyrethrin II	0.0010	<LOQ
Spinosyn A	0.0010	<LOQ
Spinosyn D	0.0010	<LOQ

Category II Mycotoxin	LOQ (ppm)	Concentration (ppm)
Ochratoxin A	0.0020	NOT TESTED
Aflatoxin B1	0.0002	NOT TESTED
Alfatoxin B2	0.0010	NOT TESTED
Alfatoxin G1	0.0002	NOT TESTED
Alfatoxin G2	0.0010	NOT TESTED

Category I Residual Pesticide	LOQ (ppm)	Concentration (ppm)
Chlorpyrifos	0.0010	<LOQ
Imazalil	0.0010	<LOQ



15.62%
Percent Moisture



LOQ = The lowest quantity this method can reliably detect. Any pesticide or mycotoxins that was not detected is assumed to be less than the stated LOQ (<LOQ).

All results reflect dry weight of material, based on % moisture of the sample.

ppb = parts per billion

Pesticides/Mycotoxin Methodology: Liquid Chromatography with Tandem Mass Spectrometry using PerkinElme QSight® LX50 UHPLC and QSight 220 Mass Spectrometer

All moisture analysis is determined by loss-on-drying measurement using OHAUS Model MB90 Moisture Content Readers.

Certified by: *Luke E. M.*
Luke Emerson Mason (Laboratory Director, Bia Diagnostics)

This report shall not be reproduced except in full without approval of the laboratory. This is to provide assurance that parts of a report are not taken out of context.
Results apply to the samples as received.

(802) 540-0148 laboratory@biadiagnostics.com

