

NeoArch Checklist

Surgeon: _____ Today's Date: _____

Patient: _____ Surgery Date: _____

Circle **Upper** or **Lower** Arch

Helix	Implants	8mm	10mm	11.5mm	13mm	16mm
3.5 mm						
3.75 mm						
4.0 mm						
4.3 mm						
5.0 mm						
6.0 mm						

Tooth #'s & implants



GM Healing Abutment

Profile	0.8 mm	1.5 mm	2.5 mm	3.5 mm	4.5 mm	5.5 mm
Ø 3.3	106.207	106.208	106.209	106.210	106.211	106.212
Ø 4.5	106.213	106.214	106.215	106.216	106.217	106.218

Use the manual Neo Screwdriver (104.060). Do not exceed the insertion torque of 10 N.cm.

GM Customizable Healing Abutments

Profile	1.5 mm	2.5 mm	3.5 mm	4.5 mm	5.5 mm	6.5 mm
Ø 5.5	106.223	106.224	106.225	106.226	106.227	
Ø 7.0	106.228	106.229	106.230	106.231	106.232	

GM Cover Screw

	0 mm	2 mm
	117.021	117.022

Use the manual Neo Screwdriver (104.060). Do not exceed the insertion torque of 10 N.cm.

Credit Card: _____

EXP: _____ Code: _____ name: _____



	GM Mini Conical Abutment : Insertion torque: 32 N.cm.	1.5mm _____ 2.5mm _____ 3.5mm _____
	17°/30° GM Exact Mini Conical Abutment : Insertion torque: 20 N.cm.	17 Degree: 1.5mm _____ 2.5mm _____ 3.5mm _____ 30 Degree: 1.5mm _____ 2.5mm _____ 3.5mm _____ 45 Degree: 1.5 _____ 2.5mm _____
	Mini Conical Abutment Protection Cylinder	Qty: _____
	Temp. coping	Qty: _____
	Impression coping	Qty: _____
	Lab Analog	Qty: _____

"All-on-X" tools

A04 Metal bendable guide

qty: _____



qty: _____ qty: _____ qty: _____

17° 30° 45°

128.032 128.033 128.034

Bone Profiler & guide pin

qty: _____

