

# GM CoCr Abutment

The set includes one GM CoCr Abutment, one Titanium Screw and one GM Implant Analog.  
Interocclusal height of 12.0 mm. Customizable up to 5.0 mm.



Single-unit  
screw-retained  
prosthesis



OR  
Single-unit  
cement-retained  
prosthesis

036



Consider in addition 1.5 - 2.0 mm for the restorative material  
Minimum interocclusal space of 4.9 mm from the mucosa level



GM CoCr Abutment Set

Ø 3.5 / 3.75

118.309

Ø 4.0 / 4.3

118.310

Ø 5.0 / 6.0

118.311

## Installation Sequence

