




GM Temporary Abutment


Customizable area made of titanium

A minimum height of 4 mm of the customizable area must be kept

With retentive grooves for acrylic material and allows customization



Temporary Protheses



Consider in addition 1.5 - 2.0 mm for the restorative material

Accessories

Replacement Sterile
Screws



Titanium 116.286



Installation Sequence

