

November 30, 2022

John Jackson
River Park Utilities Management Association,
Inc.
106 Glenn Street
Crescent City, FL 32112

RE: Project: RPUMA DBP
Pace Project No.: 35760550

Dear John Jackson:

Enclosed are the analytical results for sample(s) received by the laboratory on November 17, 2022. The results relate only to the samples included in this report. Results reported herein conform to the applicable TNI/NELAC Standards and the laboratory's Quality Manual, where applicable, unless otherwise noted in the body of the report.

The test results provided in this final report were generated by each of the following laboratories within the Pace Network:

- Pace Analytical Services - Ormond Beach

If you have any questions concerning this report, please feel free to contact me.

Sincerely,



Jeff Baylor
jeff.baylor@pacelabs.com
(386)672-5668
Project Manager

Enclosures



REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

CERTIFICATIONS

Project: RPUMA DBP

Pace Project No.: 35760550

Pace Analytical Services Ormond Beach

8 East Tower Circle, Ormond Beach, FL 32174

Alaska DEC- CS/UST/LUST

Alabama Certification #: 41320

Colorado Certification: FL NELAC Reciprocity

Connecticut Certification #: PH-0216

Delaware Certification: FL NELAC Reciprocity

Florida Certification #: E83079

Georgia Certification #: 955

Guam Certification: FL NELAC Reciprocity

Hawaii Certification: FL NELAC Reciprocity

Illinois Certification #: 200068

Indiana Certification: FL NELAC Reciprocity

Kansas Certification #: E-10383

Kentucky Certification #: 90050

Louisiana Certification #: FL NELAC Reciprocity

Louisiana Environmental Certificate #: 05007

Maine Certification #: FL01264

Maryland Certification: #346

Massachusetts Certification #: M-FL1264

Michigan Certification #: 9911

Mississippi Certification: FL NELAC Reciprocity

Missouri Certification #: 236

Montana Certification #: Cert 0074

Nebraska Certification: NE-OS-28-14

New Hampshire Certification #: 2958

New Jersey Certification #: FL022

New York Certification #: 11608

North Carolina Environmental Certificate #: 667

North Carolina Certification #: 12710

North Dakota Certification #: R-216

Ohio DEP 87780

Oklahoma Certification #: D9947

Pennsylvania Certification #: 68-00547

Puerto Rico Certification #: FL01264

South Carolina Certification: #96042001

Tennessee Certification #: TN02974

Texas Certification: FL NELAC Reciprocity

US Virgin Islands Certification: FL NELAC Reciprocity

Virginia Environmental Certification #: 460165

West Virginia Certification #: 9962C

Wisconsin Certification #: 399079670

Wyoming (EPA Region 8): FL NELAC Reciprocity

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

SAMPLE SUMMARY

Project: RPUMA DBP

Pace Project No.: 35760550

Lab ID	Sample ID	Matrix	Date Collected	Date Received
35760550001	131 Iowa St POE	Drinking Water	11/17/22 09:08	11/17/22 11:28
35760550002	148 Hicks	Drinking Water	11/17/22 09:30	11/17/22 11:28
35760550003	210 S Hayes	Drinking Water	11/17/22 09:22	11/17/22 11:28
35760550004	Live Oak Loop (Flush)	Drinking Water	11/17/22 09:41	11/17/22 11:28

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

SAMPLE ANALYTE COUNT

Project: RPUMA DBP

Pace Project No.: 35760550

Lab ID	Sample ID	Method	Analysts	Analytes Reported	Laboratory
35760550001	131 Iowa St POE	EPA 552.3	TSW	7	PASI-O
		EPA 524.2	JIF	8	PASI-O
35760550002	148 Hicks	EPA 552.3	TSW	7	PASI-O
		EPA 524.2	JIF	8	PASI-O
35760550003	210 S Hayes	EPA 552.3	TSW	7	PASI-O
		EPA 524.2	JIF	8	PASI-O
35760550004	Live Oak Loop (Flush)	EPA 552.3	TSW	7	PASI-O
		EPA 524.2	JIF	8	PASI-O

PASI-O = Pace Analytical Services - Ormond Beach

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: RPUMA DBP
Pace Project No.: 35760550

Sample: 131 Iowa St POE **Lab ID:** 35760550001 Collected: 11/17/22 09:08 Received: 11/17/22 11:28 Matrix: Drinking Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
552.3 Haloacetic Acids									
Analytical Method: EPA 552.3 Preparation Method: EPA 552.3									
Pace Analytical Services - Ormond Beach									
Dibromoacetic Acid	0.43 U	ug/L	1.0	0.43	1	11/23/22 23:12	11/26/22 20:59	631-64-1	
Dichloroacetic Acid	18.1	ug/L	1.0	0.24	1	11/23/22 23:12	11/26/22 20:59	79-43-6	
Haloacetic Acids (Total)	25.5	ug/L	1.0	0.90	1	11/23/22 23:12	11/26/22 20:59		
Monobromoacetic Acid	0.29 U	ug/L	1.0	0.29	1	11/23/22 23:12	11/26/22 20:59	79-08-3	
Monochloroacetic Acid	3.9	ug/L	1.0	0.90	1	11/23/22 23:12	11/26/22 20:59	79-11-8	
Trichloroacetic Acid	3.5	ug/L	1.0	0.26	1	11/23/22 23:12	11/26/22 20:59	76-03-9	
Surrogates									
2,3-Dibromopropanoic Acid (S)	101	%	70-130		1	11/23/22 23:12	11/26/22 20:59	600-05-5	
524.2 THM									
Analytical Method: EPA 524.2									
Pace Analytical Services - Ormond Beach									
Bromodichloromethane	0.46 I	ug/L	1.0	0.37	1		11/18/22 18:09	75-27-4	
Bromoform	0.35 U	ug/L	1.0	0.35	1		11/18/22 18:09	75-25-2	
Chloroform	9.9	ug/L	1.0	0.44	1		11/18/22 18:09	67-66-3	
Dibromochloromethane	0.47 U	ug/L	1.0	0.47	1		11/18/22 18:09	124-48-1	
Total Trihalomethanes (Calc.)	10.3	ug/L	1.0	0.47	1		11/18/22 18:09		
Surrogates									
4-Bromofluorobenzene (S)	88	%	70-130		1		11/18/22 18:09	460-00-4	
Toluene-d8 (S)	95	%	70-130		1		11/18/22 18:09	2037-26-5	
1,2-Dichlorobenzene-d4 (S)	100	%	70-130		1		11/18/22 18:09	2199-69-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: RPUMA DBP

Pace Project No.: 35760550

Sample: 148 Hicks **Lab ID: 35760550002** Collected: 11/17/22 09:30 Received: 11/17/22 11:28 Matrix: Drinking Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
552.3 Haloacetic Acids									
Analytical Method: EPA 552.3 Preparation Method: EPA 552.3									
Pace Analytical Services - Ormond Beach									
Dibromoacetic Acid	0.43 U	ug/L	1.0	0.43	1	11/23/22 23:12	11/26/22 21:21	631-64-1	
Dichloroacetic Acid	2.6	ug/L	1.0	0.24	1	11/23/22 23:12	11/26/22 21:21	79-43-6	
Haloacetic Acids (Total)	5.7	ug/L	1.0	0.90	1	11/23/22 23:12	11/26/22 21:21		
Monobromoacetic Acid	0.29 U	ug/L	1.0	0.29	1	11/23/22 23:12	11/26/22 21:21	79-08-3	
Monochloroacetic Acid	0.90 U	ug/L	1.0	0.90	1	11/23/22 23:12	11/26/22 21:21	79-11-8	
Trichloroacetic Acid	3.1	ug/L	1.0	0.26	1	11/23/22 23:12	11/26/22 21:21	76-03-9	
Surrogates									
2,3-Dibromopropanoic Acid (S)	115	%	70-130		1	11/23/22 23:12	11/26/22 21:21	600-05-5	
524.2 THM									
Analytical Method: EPA 524.2									
Pace Analytical Services - Ormond Beach									
Bromodichloromethane	0.62 I	ug/L	1.0	0.37	1		11/18/22 18:34	75-27-4	
Bromoform	0.35 U	ug/L	1.0	0.35	1		11/18/22 18:34	75-25-2	
Chloroform	12.8	ug/L	1.0	0.44	1		11/18/22 18:34	67-66-3	
Dibromochloromethane	0.47 U	ug/L	1.0	0.47	1		11/18/22 18:34	124-48-1	
Total Trihalomethanes (Calc.)	13.4	ug/L	1.0	0.47	1		11/18/22 18:34		
Surrogates									
4-Bromofluorobenzene (S)	100	%	70-130		1		11/18/22 18:34	460-00-4	
Toluene-d8 (S)	92	%	70-130		1		11/18/22 18:34	2037-26-5	
1,2-Dichlorobenzene-d4 (S)	103	%	70-130		1		11/18/22 18:34	2199-69-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: RPUMA DBP
Pace Project No.: 35760550

Sample: 210 S Hayes **Lab ID: 35760550003** Collected: 11/17/22 09:22 Received: 11/17/22 11:28 Matrix: Drinking Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
552.3 Haloacetic Acids									
Analytical Method: EPA 552.3 Preparation Method: EPA 552.3									
Pace Analytical Services - Ormond Beach									
Dibromoacetic Acid	0.43 U	ug/L	1.0	0.43	1	11/23/22 23:12	11/26/22 21:44	631-64-1	
Dichloroacetic Acid	14.6	ug/L	1.0	0.24	1	11/23/22 23:12	11/26/22 21:44	79-43-6	
Haloacetic Acids (Total)	22.4	ug/L	1.0	0.90	1	11/23/22 23:12	11/26/22 21:44		
Monobromoacetic Acid	0.29 U	ug/L	1.0	0.29	1	11/23/22 23:12	11/26/22 21:44	79-08-3	
Monochloroacetic Acid	4.2	ug/L	1.0	0.90	1	11/23/22 23:12	11/26/22 21:44	79-11-8	
Trichloroacetic Acid	3.6	ug/L	1.0	0.26	1	11/23/22 23:12	11/26/22 21:44	76-03-9	
Surrogates									
2,3-Dibromopropanoic Acid (S)	98	%	70-130		1	11/23/22 23:12	11/26/22 21:44	600-05-5	
524.2 THM									
Analytical Method: EPA 524.2									
Pace Analytical Services - Ormond Beach									
Bromodichloromethane	0.73 I	ug/L	1.0	0.37	1		11/18/22 18:59	75-27-4	
Bromoform	0.35 U	ug/L	1.0	0.35	1		11/18/22 18:59	75-25-2	
Chloroform	13.3	ug/L	1.0	0.44	1		11/18/22 18:59	67-66-3	
Dibromochloromethane	0.47 U	ug/L	1.0	0.47	1		11/18/22 18:59	124-48-1	
Total Trihalomethanes (Calc.)	14.0	ug/L	1.0	0.47	1		11/18/22 18:59		
Surrogates									
4-Bromofluorobenzene (S)	93	%	70-130		1		11/18/22 18:59	460-00-4	
Toluene-d8 (S)	95	%	70-130		1		11/18/22 18:59	2037-26-5	
1,2-Dichlorobenzene-d4 (S)	100	%	70-130		1		11/18/22 18:59	2199-69-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: RPUMA DBP

Pace Project No.: 35760550

Sample: Live Oak Loop (Flush) **Lab ID: 35760550004** Collected: 11/17/22 09:41 Received: 11/17/22 11:28 Matrix: Drinking Water

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
552.3 Haloacetic Acids									
Analytical Method: EPA 552.3 Preparation Method: EPA 552.3									
Pace Analytical Services - Ormond Beach									
Dibromoacetic Acid	0.43 U	ug/L	1.0	0.43	1	11/23/22 23:12	11/26/22 22:06	631-64-1	
Dichloroacetic Acid	6.0	ug/L	1.0	0.24	1	11/23/22 23:12	11/26/22 22:06	79-43-6	
Haloacetic Acids (Total)	9.7	ug/L	1.0	0.90	1	11/23/22 23:12	11/26/22 22:06		
Monobromoacetic Acid	0.29 U	ug/L	1.0	0.29	1	11/23/22 23:12	11/26/22 22:06	79-08-3	
Monochloroacetic Acid	2.0	ug/L	1.0	0.90	1	11/23/22 23:12	11/26/22 22:06	79-11-8	
Trichloroacetic Acid	1.7	ug/L	1.0	0.26	1	11/23/22 23:12	11/26/22 22:06	76-03-9	
Surrogates									
2,3-Dibromopropanoic Acid (S)	107	%	70-130		1	11/23/22 23:12	11/26/22 22:06	600-05-5	
524.2 THM									
Analytical Method: EPA 524.2									
Pace Analytical Services - Ormond Beach									
Bromodichloromethane	0.71 I	ug/L	1.0	0.37	1		11/18/22 19:24	75-27-4	
Bromoform	0.35 U	ug/L	1.0	0.35	1		11/18/22 19:24	75-25-2	
Chloroform	13.8	ug/L	1.0	0.44	1		11/18/22 19:24	67-66-3	
Dibromochloromethane	0.47 U	ug/L	1.0	0.47	1		11/18/22 19:24	124-48-1	
Total Trihalomethanes (Calc.)	14.6	ug/L	1.0	0.47	1		11/18/22 19:24		
Surrogates									
4-Bromofluorobenzene (S)	88	%	70-130		1		11/18/22 19:24	460-00-4	
Toluene-d8 (S)	86	%	70-130		1		11/18/22 19:24	2037-26-5	
1,2-Dichlorobenzene-d4 (S)	99	%	70-130		1		11/18/22 19:24	2199-69-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

QUALITY CONTROL DATA

Project: RPUMA DBP
Pace Project No.: 35760550

QC Batch: 872724	Analysis Method: EPA 524.2
QC Batch Method: EPA 524.2	Analysis Description: 524.2 THM MSV
	Laboratory: Pace Analytical Services - Ormond Beach

Associated Lab Samples: 35760550001, 35760550002, 35760550003, 35760550004

METHOD BLANK: 4804227 Matrix: Water
Associated Lab Samples: 35760550001, 35760550002, 35760550003, 35760550004

Parameter	Units	Blank Result	Reporting Limit	MDL	Analyzed	Qualifiers
Bromodichloromethane	ug/L	0.37 U	1.0	0.37	11/18/22 11:04	
Bromoform	ug/L	0.35 U	1.0	0.35	11/18/22 11:04	
Chloroform	ug/L	0.44 U	1.0	0.44	11/18/22 11:04	
Dibromochloromethane	ug/L	0.47 U	1.0	0.47	11/18/22 11:04	
Total Trihalomethanes (Calc.)	ug/L	0.47 U	1.0	0.47	11/18/22 11:04	
1,2-Dichlorobenzene-d4 (S)	%	99	70-130		11/18/22 11:04	
4-Bromofluorobenzene (S)	%	97	70-130		11/18/22 11:04	
Toluene-d8 (S)	%	92	70-130		11/18/22 11:04	

Parameter	Units	4804228		4804229			% Rec Limits	RPD	Max RPD	Qualifiers
		Spike Conc.	LCS Result	LCSD Result	LCS % Rec	LCSD % Rec				
Bromodichloromethane	ug/L	40	41.3	40.0	103	100	70-130	3	20	
Bromoform	ug/L	40	38.2	42.9	95	107	70-130	12	20	
Chloroform	ug/L	40	40.3	41.1	101	103	70-130	2	20	
Dibromochloromethane	ug/L	40	42.0	46.9	105	117	70-130	11	20	
Total Trihalomethanes (Calc.)	ug/L	160	162	171	101	107	70-130	5	20	
1,2-Dichlorobenzene-d4 (S)	%				98	97	70-130			
4-Bromofluorobenzene (S)	%				90	109	70-130			
Toluene-d8 (S)	%				96	95	70-130			

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

QUALITY CONTROL DATA

Project: RPUMA DBP
Pace Project No.: 35760550

QC Batch: 874274 Analysis Method: EPA 552.3
QC Batch Method: EPA 552.3 Analysis Description: 5523 Haloacetic Acids
Laboratory: Pace Analytical Services - Ormond Beach
Associated Lab Samples: 35760550001, 35760550002, 35760550003, 35760550004

METHOD BLANK: 4812169 Matrix: Water
Associated Lab Samples: 35760550001, 35760550002, 35760550003, 35760550004

Parameter	Units	Blank Result	Reporting Limit	MDL	Analyzed	Qualifiers
Dibromoacetic Acid	ug/L	0.43 U	1.0	0.43	11/26/22 18:47	
Dichloroacetic Acid	ug/L	0.24 U	1.0	0.24	11/26/22 18:47	
Haloacetic Acids (Total)	ug/L	0.90 U	1.0	0.90	11/26/22 18:47	
Monobromoacetic Acid	ug/L	0.29 U	1.0	0.29	11/26/22 18:47	
Monochloroacetic Acid	ug/L	0.90 U	1.0	0.90	11/26/22 18:47	
Trichloroacetic Acid	ug/L	0.26 U	1.0	0.26	11/26/22 18:47	
2,3-Dibromopropanoic Acid (S)	%	101	70-130		11/26/22 18:47	

LABORATORY CONTROL SAMPLE: 4812170

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Dibromoacetic Acid	ug/L	10	10.0	100	70-130	
Dichloroacetic Acid	ug/L	10	10.1	101	70-130	
Haloacetic Acids (Total)	ug/L	50	49.4	99	70-130	
Monobromoacetic Acid	ug/L	10	9.9	99	70-130	
Monochloroacetic Acid	ug/L	10	9.4	94	70-130	
Trichloroacetic Acid	ug/L	10	10	100	70-130	
2,3-Dibromopropanoic Acid (S)	%			106	70-130	

MATRIX SPIKE & MATRIX SPIKE DUPLICATE: 4812171 4812172

Parameter	Units	35761770001		MSD		MS		MSD		% Rec Limits	RPD	Max RPD	Qual
		Result	MS Spike Conc.	MSD Spike Conc.	Result	MSD Result	% Rec	% Rec					
Dibromoacetic Acid	ug/L	0.0038 mg/L	10	10	14.1	13.6	102	98	70-130	3	30		
Dichloroacetic Acid	ug/L	0.0010 mg/L	10	10	11.6	11.0	106	100	70-130	6	30		
Haloacetic Acids (Total)	ug/L	0.0052 mg/L	50	50	58.2	55.5	106	101	70-130	5	30		
Monobromoacetic Acid	ug/L	0.00029 U mg/L	10	10	11.4	11.1	114	111	70-130	3	30		
Monochloroacetic Acid	ug/L	0.00090 U mg/L	10	10	10.1	9.4	101	94	70-130	6	30		
Trichloroacetic Acid	ug/L	0.00030 I mg/L	10	10	11.0	10.4	107	101	70-130	5	30		
2,3-Dibromopropanoic Acid (S)	%						103	97	70-130		30		

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

QUALIFIERS

Project: RPUMA DBP

Pace Project No.: 35760550

DEFINITIONS

DF - Dilution Factor, if reported, represents the factor applied to the reported data due to dilution of the sample aliquot.

ND - Not Detected at or above adjusted reporting limit.

TNTC - Too Numerous To Count

MDL - Adjusted Method Detection Limit.

PQL - Practical Quantitation Limit.

RL - Reporting Limit - The lowest concentration value that meets project requirements for quantitative data with known precision and bias for a specific analyte in a specific matrix.

S - Surrogate

1,2-Diphenylhydrazine decomposes to and cannot be separated from Azobenzene using Method 8270. The result for each analyte is a combined concentration.

Consistent with EPA guidelines, unrounded data are displayed and have been used to calculate % recovery and RPD values.

LCS(D) - Laboratory Control Sample (Duplicate)

MS(D) - Matrix Spike (Duplicate)

DUP - Sample Duplicate

RPD - Relative Percent Difference

NC - Not Calculable.

SG - Silica Gel - Clean-Up

U - Indicates the compound was analyzed for, but not detected.

N-Nitrosodiphenylamine decomposes and cannot be separated from Diphenylamine using Method 8270. The result reported for each analyte is a combined concentration.

Reported results are not rounded until the final step prior to reporting. Therefore, calculated parameters that are typically reported as "Total" may vary slightly from the sum of the reported component parameters.

Pace Analytical is TNI accredited. Contact your Pace PM for the current list of accredited analytes.

TNI - The NELAC Institute.

ANALYTE QUALIFIERS

I The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.

U Compound was analyzed for but not detected.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

QUALITY CONTROL DATA CROSS REFERENCE TABLE

Project: RPUMA DBP

Pace Project No.: 35760550

Lab ID	Sample ID	QC Batch Method	QC Batch	Analytical Method	Analytical Batch
35760550001	131 Iowa St POE	EPA 552.3	874274	EPA 552.3	874458
35760550002	148 Hicks	EPA 552.3	874274	EPA 552.3	874458
35760550003	210 S Hayes	EPA 552.3	874274	EPA 552.3	874458
35760550004	Live Oak Loop (Flush)	EPA 552.3	874274	EPA 552.3	874458
35760550001	131 Iowa St POE	EPA 524.2	872724		
35760550002	148 Hicks	EPA 524.2	872724		
35760550003	210 S Hayes	EPA 524.2	872724		
35760550004	Live Oak Loop (Flush)	EPA 524.2	872724		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

Pace

WO#: 35760550
 PM: JSB Due Date: 11/30/22
 CLIENT: 37-RPUMA

Project #
 Project Manager:
 Client:

Date and Initials of person:
 Examining contents:
 Label: NPI
 Deliver:
 pH:

Thermometer Used: T-399 Date: 11-17-22 Time: 1136 Initials: NPI

State of Origin: _____ For WV projects, all containers verified to ≤6 °C

Cooler #1 Temp.°C <u>9.8</u> (Visual) <u>0.0</u> (Correction Factor) <u>9.8</u> (Actual)	<input checked="" type="checkbox"/> Samples on ice, cooling process has begun
Cooler #2 Temp.°C <u>12.2</u> (Visual) _____ (Correction Factor) <u>12.2</u> (Actual)	<input type="checkbox"/> Samples on ice, cooling process has begun
Cooler #3 Temp.°C _____ (Visual) _____ (Correction Factor) _____ (Actual)	<input type="checkbox"/> Samples on ice, cooling process has begun
Cooler #4 Temp.°C _____ (Visual) _____ (Correction Factor) _____ (Actual)	<input type="checkbox"/> Samples on ice, cooling process has begun
Cooler #5 Temp.°C _____ (Visual) _____ (Correction Factor) _____ (Actual)	<input type="checkbox"/> Samples on ice, cooling process has begun
Cooler #6 Temp.°C _____ (Visual) _____ (Correction Factor) _____ (Actual)	<input type="checkbox"/> Samples on ice, cooling process has begun

Recheck for OOT °C _____ (Visual) _____ (Correction Factor) _____ (Actual) Time: _____ Initials: _____

Courier: Fed Ex UPS USPS Client Commercial Pace Other _____

Shipping Method: First Overnight Priority Overnight Standard Overnight Ground International Priority
 Other _____

Billing: Recipient Sender Third Party Credit Card Unknown

Tracking # _____

Custody Seal on Cooler/Box Present: Yes No Seals intact: Yes No Ice: Wet Blue Melted None

Packing Material: Bubble Wrap Bubble Bags None Other _____

Samples shorted to lab (If Yes, complete) Shorted Date: 11-17-22 Shorted Time: 1139 Qty: 1 *SPST*

Comments:

Chain of Custody Present	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	
Chain of Custody Filled Out	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	
Relinquished Signature & Sampler Name COC	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	
Samples Arrived within Hold Time	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	
Rush TAT requested on COC	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	
Sufficient Volume	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	
Correct Containers Used	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	
Containers Intact	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	
Sample Labels match COC (sample IDs & date/time of collection)	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	
All containers needing acid/base preservation have been checked.	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	Preservation Information: Preservative: _____ Lot #/Trace #: _____ Date: _____ Time: _____ Initials: _____
All Containers needing preservation are found to be in compliance with EPA recommendation: Exceptions: Vials, Microbiology, O&G, PFAS	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	
Headspace in VOA Vials? (>6mm):	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	
Trip Blank Present:	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	

Comments/ Resolution (use back for additional comments):

