

- ◆ Universal sample interface for ALL FT-IR analyzers
- ◆ No Sample Preparation
- ◆ Full pH range
- ◆ Upgradeable
- ◆ Quick Clean-up



DuraSAMPLIR II™

DuraSAMPLIR II is the fully upgradable version of the popular **DuraSAMPLIR™** diamond ATR accessory for FT-IR spectrometers. By adding **DuraSAMPLIR II DiComp™** capability to your instrument today, you will enjoy the convenience of rapid sample change-overs, high energy throughput with outstanding sensitivity, chemical inertness for “worry-free” analysis of all solid and liquid samples, and abrasion resistance for long-term reliability.

Samples are pressed against the diamond IRE with the simple twist of a knob. The press supplies constant force for controlled compression without rotating the sample contact tip to prevent stripping or sheering of your sample. For samples requiring controlled, repeatable contact force, your **DuraSAMPLIR II** may be ordered or upgraded to include the force readout and overload alarm option.



DuraSAMPLIR II is available with a variety of DiComp DuraDisk™ models for enhanced sensitivity and to improve your quantitative results. Single and multi-reflection disks have been engineered using diamond, the world’s most durable substance, and a ZnSe substrate for maximum IR signal.

For applications requiring visual confirmation of contact in the analysis area, your **DuraSAMPLIR II** may be upgraded to add **DuraScope™** video imaging capability. Accurate positioning of fibers, contamination, ink spots, and other small features can now be accomplished easily and with complete confidence in your results.

DuraSamplIR II Options and Upgrades

DuraDisk (Choose one)

Single Reflection DuraDisk™

Diamond Thickness: 0.5mm

Applications: Samples requiring high pressure and/or high throughput.

Size: 1mm diamond diameter with 200um "sweet spot."

Mount: Flush for optimum sample contact

DuraScope Compatible: Yes

Temperature: 100 degrees C. Max.

Range: 4000-650

Diamond Absorption: Usually not visible in spectrum. 1900 - 2200cm⁻¹ region undisturbed.

Triple Reflection DuraDisk™

Diamond Thickness: 0.35mm

Applications: For samples requiring increased sensitivity

Size: 2mm diamond diameter with 2mm "sweet spot" with concentric ring.

Mount: Flush for optimum sample contact

DuraScope Compatible: Yes

Temperature: 100 degrees C. Max.

Range: 4000-650

Diamond Absorption: Diamond absorption region (1900 - 2200cm⁻¹) useful for strong peaks.

Nine Reflection DuraDisk™

Diamond Thickness: 0.5mm

Applications: Samples requiring high sensitivity. Ideal for liquids.

Size: 4mm diamond diameter with 4mm "sweet spot" with 3 concentric rings.

Mount: Recessed sampling area for liquid retention.

DuraScope Compatible: No

Temperature: 100 degrees C. Max.

Range: 4000-650

Diamond Absorption: Diamond absorption region (1900 - 2200cm⁻¹) useless.

Force Applicator Option

Adds the functionality of integrated force indicator and overload alarm. Ideal for analysis of samples requiring uniform and repeatable contact. Force applicator features interchangeable tips.



Force Applicator Tip Options

Syringe - Ideal for dispensing and retrieving small measurable amounts of liquid.

Acrylic Rod - Allows light onto sample area for DuraScope applications.

Stainless Steel Rod - for high pressure contact.




DuraScope Video Imaging Option

With the DuraScope upgrade, your DuraSamplIR II equipped spectrometer will enjoy the imaging capabilities of video microscopy. You will view the sample through the DiComp sensor allowing you to verify the area of the sample you are analyzing and ensuring adequate contact.

With the addition of a video capture card, you can import video images into your computer system allowing you to include and archive pictures of your subject matter. Reports and presentations can now include these images.



Fiber positioned for analysis.



Fiber compressed for analysis

Represented by:



15 Great Pasture Rd.
Danbury, CT 06810-8153

Tel: (203) 207-9700 Fax: (203) 207-9780

Email: info@sensir.com

Visit our Web site at www.sensir.com