

## Client Sample Results

Client: Alabama Dept. Environmental Management  
 Project/Site: ADEM PFC Sampling - Centre

TestAmerica Job ID: 280-93528-2

**Client Sample ID: CENTRE-FINISHED**

**Lab Sample ID: 280-93528-1**

Date Collected: 02/01/17 11:30

Matrix: Water

Date Received: 02/02/17 09:50

Method: DV-LC-0012 - Perfluorinated Hydrocarbons									
Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorobutane Sulfonate (PFBS)	0.16		0.020	0.0082	ug/L		02/07/17 10:12	02/08/17 16:22	1
Perfluorobutanoic acid (PFBA)	0.038		0.020	0.0097	ug/L		02/07/17 10:12	02/08/17 16:22	1
Perfluorodecane sulfonate (PFDS)	0.0091	U	0.020	0.0091	ug/L		02/07/17 10:12	02/08/17 16:22	1
Perfluorodecanoic acid (PFDA)	0.0078	U	0.020	0.0078	ug/L		02/07/17 10:12	02/08/17 16:22	1
Perfluorododecanoic acid (PFDoA)	0.015	U	0.030	0.015	ug/L		02/07/17 10:12	02/08/17 16:22	1
Perfluoroheptanoic acid (PFHpA)	0.039		0.030	0.013	ug/L		02/07/17 10:12	02/08/17 16:22	1
Perfluorohexane Sulfonate (PFHxS)	0.0071	J	0.030	0.0069	ug/L		02/07/17 10:12	02/08/17 16:22	1
Perfluorohexanoic acid (PFHxA)	0.20		0.020	0.0029	ug/L		02/07/17 10:12	02/08/17 16:22	1
Perfluorononanoic acid (PFNA)	0.017	U	0.040	0.017	ug/L		02/07/17 10:12	02/08/17 16:22	1
Perfluorooctanoic acid (PFOA)	0.049		0.020	0.0097	ug/L		02/07/17 10:12	02/08/17 16:22	1
Perfluorooctane Sulfonate (PFOS)	0.070		0.030	0.013	ug/L		02/07/17 10:12	02/08/17 16:22	1
Perfluoropentanoic acid (PFPA)	0.24		0.030	0.011	ug/L		02/07/17 10:12	02/08/17 16:22	1
Perfluorotetradecanoic acid (PFTeA)	0.015	U	0.030	0.015	ug/L		02/07/17 10:12	02/08/17 16:22	1
Perfluorotridecanoic Acid (PFTriA)	0.018	U	0.040	0.018	ug/L		02/07/17 10:12	02/08/17 16:22	1
Perfluoroundecanoic acid (PFUnA)	0.0069	U	0.020	0.0069	ug/L		02/07/17 10:12	02/08/17 16:22	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
13C8 PFOA	109		60 - 155				02/07/17 10:12	02/08/17 16:22	1
13C8 PFOS	111		45 - 130				02/07/17 10:12	02/08/17 16:22	1

Method: PFC -FOSA - FOSA in Water (LC/MS/MS)									
Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorooctane Sulfonamide (FOSA)	0.0055	U	0.048	0.0055	ug/L		02/07/17 10:10	02/08/17 20:16	1