

Default Detection Limits

Client: Alabama Dept. Environmental Management
 Project/Site: ADEM PFC Sampling - Centre

TestAmerica Job ID: 280-93143-2

Method: DV-LC-0012 - Perfluorinated Hydrocarbons

Prep: 3535

Analyte	RL	MDL	Units	Method
Perfluorobutane Sulfonate (PFBS)	0.020	0.0082	ug/L	DV-LC-0012
Perfluorobutanoic acid (PFBA)	0.020	0.0098	ug/L	DV-LC-0012
Perfluorodecane sulfonate (PFDS)	0.020	0.0092	ug/L	DV-LC-0012
Perfluorodecanoic acid (PFDA)	0.020	0.0078	ug/L	DV-LC-0012
Perfluorododecanoic acid (PFDoA)	0.030	0.015	ug/L	DV-LC-0012
Perfluoroheptanoic acid (PFHpA)	0.030	0.013	ug/L	DV-LC-0012
Perfluorohexane Sulfonate (PFHxS)	0.030	0.0070	ug/L	DV-LC-0012
Perfluorohexanoic acid (PFHxA)	0.020	0.0029	ug/L	DV-LC-0012
Perfluorononanoic acid (PFNA)	0.040	0.017	ug/L	DV-LC-0012
Perfluorooctane Sulfonate (PFOS)	0.030	0.013	ug/L	DV-LC-0012
Perfluorooctanoic acid (PFOA)	0.020	0.0098	ug/L	DV-LC-0012
Perfluoropentanoic acid (PFPA)	0.030	0.011	ug/L	DV-LC-0012
Perfluorotetradecanoic acid (PFTeA)	0.030	0.015	ug/L	DV-LC-0012
Perfluorotridecanoic Acid (PFTriA)	0.040	0.018	ug/L	DV-LC-0012
Perfluoroundecanoic acid (PFUnA)	0.020	0.0069	ug/L	DV-LC-0012

Method: PFC -FOSA - FOSA in Water (LC/MS/MS)

Prep: 3535

Analyte	RL	MDL	Units	Method
Perfluorooctane Sulfonamide (FOSA)	0.050	0.0057	ug/L	PFC -FOSA