



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

CROSS COUNTRY CONTRACTORS, INC.  
1045 N. Carrollton Ave.  
Baton Rouge, Louisiana 70806  
Joe Amorello Phone: 1 800 742 2040

CALIBRATION

Valid To: February 28, 2028

Certificate Number: 6303.01

In recognition of the successful completion of the A2LA evaluation process (including an assessment of the organization's compliance with R205 – A2LA's Calibration Program Requirements), accreditation is granted to this laboratory to perform the following calibrations<sup>1, 5</sup>:

I. Mechanical

| Parameter/Equipment            | Range                                                                                  | CMC <sup>2, 3, 4</sup> (±)                                                    | Comments                                           |
|--------------------------------|----------------------------------------------------------------------------------------|-------------------------------------------------------------------------------|----------------------------------------------------|
| Torque – Measuring Equipment   |                                                                                        |                                                                               |                                                    |
| Wrenches                       | (10 to 100) lbf·in<br>(50 to 500) lbf·in<br>(25 to 250) lbf·ft<br>(100 to 1000) lbf·ft | 4 lbf·in (4 %)<br>16 lbf·in (3.2 %)<br>9 lbf·ft (3.6 %)<br>7.2 lbf·ft (0.7 %) | Torque transducer                                  |
| Pneumatic                      | (750 to 6000) lbf·ft                                                                   | 98 lbf·ft (1.3 %)                                                             | AWS transducer<br>(pressure to torque correlation) |
| Hydraulic                      | (3000 to 30 000) lbf·ft                                                                | 350 lbf·ft (1.8 %)                                                            | AWS transducer<br>(pressure to torque correlation) |
| Pressure – Measuring Equipment | (up to 40 000) psig                                                                    | 0.12 % of reading                                                             | Master pressure gauge                              |

<sup>1</sup> This laboratory offers commercial calibration service.

<sup>2</sup> Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMC's represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of  $k = 2$ . The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

<sup>3</sup> In the statement of CMC, percentages are percentage of reading, unless otherwise indicated.

<sup>4</sup> The type of instrument or material being calibrated is defined by the parameter. This indicates the laboratory is capable of calibrating instruments that measure or generate the values in the ranges indicated for the listed measurement parameter.

<sup>5</sup> This scope meets A2LA's *P112 Flexible Scope Policy*.



## Accredited Laboratory

A2LA has accredited

**CROSS COUNTRY CONTRACTORS, INC**

*Batton Rouge, LA*

for technical competence in the field of

**Calibration**

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 11<sup>th</sup> day of May 2026.

A blue ink signature of Mr. Trace McInturff.

Mr. Trace McInturff, Vice President, Accreditation Services  
For the Accreditation Council  
Certificate Number 6303.01  
Valid to Febraury 29, 2028

*For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.*