


Clinical Laboratory of San Bernardino, Inc.



Mountain Water Company, LLC.
P. O. Box 4096
Big Bear Lake CA, 92315

Project: Routine
Sub Project: Title 21-Annual
Project Manager: Eric Heino

Work Order: 24G2638
Received: 07/24/24 16:00
Reported: 08/21/24

Bottled Water Product **24G2638-01 (Drinking Water)** **Sample Date:** 07/24/24 8:30 **Sampler:** Not Listed

Analyte	Method	Result	Units	Rep. Limit	Prepared	Analyzed	Batch	Qualifier
---------	--------	--------	-------	------------	----------	----------	-------	-----------

Microbiology Analyses

Total Coliform (10 Tube)	SM 9221	ND	MPN/100 mL	1.1	07/25/24	07/27/24	2431035	
Fecal Coliform (10 Tube)	SM 9221	ND	MPN/100 mL	1.1	07/25/24	07/27/24	2431035	

General Physical Analyses

Apparent Color	SM 2120BM	ND	Color Units	3.0	07/25/24	07/25/24	2430173	
Odor Threshold	EPA 140.1-M	ND	TON	1	07/25/24	07/25/24	2430173	
Turbidity	EPA 180.1	ND	NTU	0.10	07/25/24	07/25/24	2430173	


General Chemical Analyses

Alkalinity, Total (as CaCO ₃)	SM 2320 B	160	mg/L	5.0	08/05/24	08/05/24	2430132	
Bicarbonate (HCO ₃)	SM 2320 B	190	mg/L	5.0	08/05/24	08/05/24	2430132	
Carbonate (CO ₃)	SM 2320B	ND	mg/L	5.0	08/05/24	08/05/24	2430132	
Chloride (Cl)	EPA 300.0	10	mg/L	1.0	07/25/24	07/25/24	2430121	
Cl Res Total	SM 4500CIG	ND	mg/L	0.04	07/25/24	07/25/24	2430157	
Cyanide (CN)	SM4500CNF	ND	ug/L	100	07/29/24	07/29/24	2431010	
Specific Conductance (E.C.)	SM 2510B	370	umhos/cm	2.0	07/25/24	08/05/24	2430132	
Fluoride (F)	EPA 300.0	ND	mg/L	0.10	07/25/24	07/25/24	2430121	
Hydroxide (OH)	SM 2320B	ND	mg/L	5.0	08/05/24	08/05/24	2430132	
MBAS (LAS Mole. Wt 340.0)	SM 5540C	ND	mg/L	0.10	07/25/24	07/25/24	2430144	
Nitrate as N (NO ₃ -N)	EPA 300.0	ND	mg/L	0.40	07/25/24	07/25/24	2430121	
Nitrate as N (NO ₃ -N)	EPA 300.0	ND	mg/L	0.40	07/25/24	07/25/24	2430121	
Nitrate + Nitrite (as N)	EPA 300.0	ND	mg/L	0.40	07/25/24	07/25/24	2430121	
Nitrite as N (NO ₂ -N)	EPA 300.0	ND	mg/L	0.40	07/25/24	07/25/24	2430121	
Perchlorate (ClO ₄)	EPA 314.0	ND	ug/L	1.0	07/30/24	07/30/24	2431082	
pH (Lab)	SM 4500HB	7.1	pH Units		07/25/24	08/05/24	2430132	
Temp. (C)	SM 4500 HB	21.9	°C		07/25/24	08/05/24	2430132	
Sulfate (SO ₄)	EPA 300.0	8.9	mg/L	0.50	07/25/24	07/25/24	2430121	
Total Filterable Residue/TDS	SM 2540C	240	mg/L	5.0	07/31/24	08/02/24	2431088	

Metals

Aluminum (Al)	EPA 200.7	ND	ug/L	50	08/02/24	08/02/24	2431178	
Antimony (Sb)	EPA 200.8	ND	ug/L	6.0	08/01/24	08/01/24	2431139	
Arsenic (As)	EPA 200.8	ND	ug/L	2.0	08/01/24	08/01/24	2431139	
Barium (Ba)	EPA 200.7	ND	ug/L	100	08/02/24	08/02/24	2431178	
Beryllium (Be)	EPA 200.8	ND	ug/L	1.0	08/01/24	08/01/24	2431139	
Boron (B)	EPA 200.7	ND	ug/L	100	08/02/24	08/02/24	2431178	
Cadmium (Cd)	EPA 200.8	ND	ug/L	1.0	08/01/24	08/01/24	2431139	
Calcium (Ca)	EPA 200.7	48	mg/L	1.0	08/01/24	08/01/24	2431130	
Chromium (Total Cr)	EPA 200.8	ND	ug/L	10	08/01/24	08/01/24	2431139	
Copper (Cu)	EPA 200.7	ND	ug/L	50	08/02/24	08/02/24	2431178	
Iron (Fe)	EPA 200.7	ND	ug/L	100	08/02/24	08/02/24	2431178	
Lead (Pb)	EPA 200.8	ND	ug/L	5.0	08/01/24	08/01/24	2431139	
Magnesium (Mg)	EPA 200.7	11	mg/L	1.0	08/01/24	08/01/24	2431130	

Clinical Laboratory of San Bernardino, Inc.



Mountain Water Company, LLC.

P. O. Box 4096

Big Bear Lake CA, 92315

Project: Routine

Sub Project: Title 21-Annual

Project Manager: Eric Heino

Work Order: 24G2638

Received: 07/24/24 16:00

Reported: 08/21/24

Bottled Water Product

24G2638-01 (Drinking Water)

Sample Date: 07/24/24 8:30

Sampler: Not Listed

Analyte	Method	Result	Units	Rep. Limit	Prepared	Analyzed	Batch	Qualifier
---------	--------	--------	-------	------------	----------	----------	-------	-----------

Metals

Manganese (Mn)	EPA 200.7	ND	ug/L	20	08/02/24	08/02/24	2431178	
Mercury (Hg)	EPA 200.8	ND	ug/L	1.0	07/30/24	07/30/24	2431059	
Nickel (Ni)	EPA 200.8	ND	ug/L	10	08/01/24	08/01/24	2431139	
Potassium (K)	EPA 200.7	1.6	mg/L	1.0	08/01/24	08/01/24	2431130	
Selenium (Se)	EPA 200.8	ND	ug/L	5.0	08/01/24	08/01/24	2431139	
Silver (Ag)	EPA 200.8	ND	ug/L	10	08/01/24	08/01/24	2431139	
Sodium (Na)	EPA 200.7	14	mg/L	1.0	08/01/24	08/01/24	2431130	
Thallium (Tl)	EPA 200.8	ND	ug/L	1.0	08/01/24	08/01/24	2431139	
Vanadium (V)	EPA 200.8	3.2	ug/L	3.0	08/01/24	08/01/24	2431139	
Zinc (Zn)	EPA 200.7	ND	ug/L	50	08/02/24	08/02/24	2431178	

Calculated Analysis

Hardness, Total (as CaCO3)	Calculated	160	mg/L	6.6	08/01/24	08/01/24	[CALC]	
Total Anions	Calculated	3.58	meq/L		08/01/24	08/05/24	[CALC]	
Total Cations	Calculated	3.96	meq/L		08/01/24	08/01/24	[CALC]	
% difference	Calculated	9.9			08/01/24	08/05/24	[CALC]	


Radiochemistry Analyses

Gross Alpha	SM 7110C	ND	pCi/L	3.0	07/24/24	07/25/24	2430072	
Gross Alpha Counting Error	SM 7110C	0.52	pCi/L		07/24/24	07/25/24	2430072	
Gross Alpha Min Det Activity	SM 7110C	0.46	pCi/L		07/24/24	07/25/24	2430072	
Uranium	EPA 200.8	ND	pCi/L	1.0	07/29/24	07/29/24	2431022	

Volatile Organic Analyses

Vinyl Chloride (VC)	EPA 524.2	ND	ug/L	0.50	07/26/24	07/26/24	2430161	
Trichlorofluoromethane (FREON 11)	EPA 524.2	ND	ug/L	5.0	07/26/24	07/26/24	2430161	
1,1-Dichloroethylene (1,1-DCE)	EPA 524.2	ND	ug/L	0.50	07/26/24	07/26/24	2430161	
1,1,2-Trichloro-1,2,2-trifluoroethane	EPA 524.2	ND	ug/L	10	07/26/24	07/26/24	2430161	
Dichloromethane (Methylene Chloride)	EPA 524.2	ND	ug/L	0.50	07/26/24	07/26/24	2430161	
trans-1,2-Dichloroethylene (t-1,2-DCE)	EPA 524.2	ND	ug/L	0.50	07/26/24	07/26/24	2430161	
Methyl tert-Butyl Ether	EPA 524.2	ND	ug/L	3.0	07/26/24	07/26/24	2430161	
1,1-Dichloroethane (1,1-DCA)	EPA 524.2	ND	ug/L	0.50	07/26/24	07/26/24	2430161	
cis-1,2-Dichloroethylene (c-1,2-DCE)	EPA 524.2	ND	ug/L	0.50	07/26/24	07/26/24	2430161	
Chloroform (Trichloromethane)	EPA 524.2	1.9	ug/L	1.0	07/26/24	07/26/24	2430161	
Carbon Tetrachloride	EPA 524.2	ND	ug/L	0.50	07/26/24	07/26/24	2430161	
1,1,1-Trichloroethane (1,1,1-TCA)	EPA 524.2	ND	ug/L	0.50	07/26/24	07/26/24	2430161	
Benzene	EPA 524.2	ND	ug/L	0.50	07/26/24	07/26/24	2430161	
1,2-Dichloroethane (1,2-DCA)	EPA 524.2	ND	ug/L	0.50	07/26/24	07/26/24	2430161	
Trichloroethylene (TCE)	EPA 524.2	ND	ug/L	0.50	07/26/24	07/26/24	2430161	
1,2-Dichloropropane	EPA 524.2	ND	ug/L	0.50	07/26/24	07/26/24	2430161	
Bromodichloromethane	EPA 524.2	ND	ug/L	1.0	07/26/24	07/26/24	2430161	
Toluene	EPA 524.2	ND	ug/L	0.50	07/26/24	07/26/24	2430161	
Tetrachloroethylene (PCE)	EPA 524.2	ND	ug/L	0.50	07/26/24	07/26/24	2430161	
1,1,2-Trichloroethane (1,1,2-TCA)	EPA 524.2	ND	ug/L	0.50	07/26/24	07/26/24	2430161	

Clinical Laboratory of San Bernardino, Inc.



Mountain Water Company, LLC.

P. O. Box 4096

Big Bear Lake CA, 92315

Project: Routine

Sub Project: Title 21-Annual

Project Manager: Eric Heino

Work Order: 24G2638

Received: 07/24/24 16:00

Reported: 08/21/24

Bottled Water Product

24G2638-01 (Drinking Water)

Sample Date: 07/24/24 8:30

Sampler: Not Listed

Analyte	Method	Result	Units	Rep. Limit	Prepared	Analyzed	Batch	Qualifier
---------	--------	--------	-------	------------	----------	----------	-------	-----------

Volatile Organic Analyses

Dibromochloromethane	EPA 524.2	ND	ug/L	1.0	07/26/24	07/26/24	2430161	
Monochlorobenzene (Chlorobenzene)	EPA 524.2	ND	ug/L	0.50	07/26/24	07/26/24	2430161	
Ethyl Benzene	EPA 524.2	ND	ug/L	0.50	07/26/24	07/26/24	2430161	
m,p-Xylene	EPA 524.2	ND	ug/L	1.0	07/26/24	07/26/24	2430161	
cis-1,3-Dichloropropene	EPA 524.2	ND	ug/L	0.50	07/26/24	07/26/24	2430161	
trans-1,3-Dichloropropene	EPA 524.2	ND	ug/L	0.50	07/26/24	07/26/24	2430161	
o-Xylene	EPA 524.2	ND	ug/L	0.50	07/26/24	07/26/24	2430161	
Styrene	EPA 524.2	ND	ug/L	0.50	07/26/24	07/26/24	2430161	
Bromoform	EPA 524.2	ND	ug/L	1.0	07/26/24	07/26/24	2430161	
1,1,2,2-Tetrachloroethane	EPA 524.2	ND	ug/L	0.50	07/26/24	07/26/24	2430161	
1,4-Dichlorobenzene (p-DCB)	EPA 524.2	ND	ug/L	0.50	07/26/24	07/26/24	2430161	
1,2-Dichlorobenzene (o-DCB)	EPA 524.2	ND	ug/L	0.50	07/26/24	07/26/24	2430161	
1,2,4-Trichlorobenzene	EPA 524.2	ND	ug/L	0.50	07/26/24	07/26/24	2430161	
Total 1,3-Dichloropropene	EPA 524.2	ND	ug/L	0.50	07/26/24	07/26/24	2430161	
Total Trihalomethanes (TTHM)	EPA 524.2	1.9	ug/L	1.0	07/26/24	07/26/24	2430161	
Total Xylenes (m,p & o)	EPA 524.2	ND	ug/L	0.50	07/26/24	07/26/24	2430161	
Surrogate: 1,2-Dichlorobenzene-d4	EPA 524.2	83 %			07/26/24	07/26/24	2430161	
Surrogate: Bromofluorobenzene	EPA 524.2	84 %			07/26/24	07/26/24	2430161	


Semi-Volatile Organic Analyses / EPA 504

Ethylene Dibromide (EDB)	EPA 504.1	ND	ug/L	0.020	08/01/24	08/02/24	2431124	
Dibromochloropropane (DBCP)	EPA 504.1	ND	ug/L	0.010	08/01/24	08/02/24	2431124	

Synthetic Organic Analyses

Endrin	EPA 508.1	ND	ug/L	0.10	08/01/24	08/12/24	2431123	
Lindane (gamma-BHC)	EPA 508.1	ND	ug/L	0.20	08/01/24	08/12/24	2431123	
Methoxychlor	EPA 508.1	ND	ug/L	10	08/01/24	08/12/24	2431123	
Toxaphene	EPA 508.1	ND	ug/L	1.0	08/01/24	08/12/24	2431123	
Chlordane	EPA 508.1	ND	ug/L	0.10	08/01/24	08/12/24	2431123	
Heptachlor	EPA 508.1	ND	ug/L	0.010	08/01/24	08/12/24	2431123	
Heptachlor Epoxide	EPA 508.1	ND	ug/L	0.010	08/01/24	08/12/24	2431123	
Hexachlorobenzene	EPA 508.1	ND	ug/L	0.50	08/01/24	08/12/24	2431123	
Hexachlorocyclopentadiene	EPA 508.1	ND	ug/L	1.0	08/01/24	08/12/24	2431123	
Polychlorinated Biphenyls (PCBs)	EPA 508.1	ND	ug/L	0.50	08/01/24	08/12/24	2431123	
Surrogate: 4,4'-Dichlorobiphenyl	EPA 508.1	110 %			08/01/24	08/12/24	2431123	
Dalapon	EPA 515.4	ND	ug/L	10	08/06/24	08/18/24	2432072	
2,4,5-TP (SILVEX)	EPA 515.4	ND	ug/L	1.0	08/06/24	08/18/24	2432072	
Bentazon (BASAGRAN)	EPA 515.4	ND	ug/L	2.0	08/06/24	08/18/24	2432072	
Picloram	EPA 515.4	ND	ug/L	1.0	08/06/24	08/18/24	2432072	
2,4-D	EPA 515.4	ND	ug/L	10	08/06/24	08/18/24	2432072	
Pentachlorophenol (PCP)	EPA 515.4	ND	ug/L	0.20	08/06/24	08/18/24	2432072	
Dinoseb (DNBP)	EPA 515.4	ND	ug/L	2.0	08/06/24	08/18/24	2432072	
Surrogate: 2,4-Dichlorophenylacetic acid	EPA 515.4	98 %			08/06/24	08/18/24	2432072	
Alachlor (ALANEX)	EPA 525.2	ND	ug/L	1.0	07/25/24	08/02/24	2430119	

Clinical Laboratory of San Bernardino, Inc.



Mountain Water Company, LLC.

P. O. Box 4096

Big Bear Lake CA, 92315

Project: Routine

Sub Project: Title 21-Annual

Project Manager: Eric Heino

Work Order: 24G2638

Received: 07/24/24 16:00

Reported: 08/21/24

Bottled Water Product

24G2638-01 (Drinking Water)

Sample Date: 07/24/24 8:30

Sampler: Not Listed

Analyte	Method	Result	Units	Rep. Limit	Prepared	Analyzed	Batch	Qualifier
---------	--------	--------	-------	------------	----------	----------	-------	-----------

Synthetic Organic Analyses

Atrazine (AATREX)	EPA 525.2	ND	ug/L	0.50	07/25/24	08/02/24	2430119	
Benzo(a)pyrene	EPA 525.2	ND	ug/L	0.10	07/25/24	08/02/24	2430119	
Diethylhexylphthalate (DEHP)	EPA 525.2	ND	ug/L	3.0	07/25/24	08/02/24	2430119	
Di(2-ethylhexyl) adipate	EPA 525.2	ND	ug/L	5.0	07/25/24	08/02/24	2430119	
Molinate (ORDRAM)	EPA 525.2	ND	ug/L	2.0	07/25/24	08/02/24	2430119	
Simazine (PRINCEP)	EPA 525.2	ND	ug/L	1.0	07/25/24	08/02/24	2430119	
Thiobencarb (BOLERO)	EPA 525.2	ND	ug/L	1.0	07/25/24	08/02/24	2430119	
Surrogate: 1,3-dimethyl-2-nitrobenzene	EPA 525.2	106 %			07/25/24	08/02/24	2430119	
Surrogate: Perylene-d12	EPA 525.2	74 %			07/25/24	08/02/24	2430119	
Surrogate: Triphenylphosphate	EPA 525.2	70 %			07/25/24	08/02/24	2430119	
Oxamyl (VYDATE)	EPA 531.1	ND	ug/L	20	08/18/24	08/20/24	2434001	
Carbofuran (FURADAN)	EPA 531.1	ND	ug/L	5.0	08/18/24	08/20/24	2434001	
Glyphosate	EPA 547	ND	ug/L	25	08/05/24	08/05/24	2432007	
Endothall	EPA 548.1	ND	ug/L	45	07/25/24	07/27/24	2430140	
Diquat	EPA 549.2	ND	ug/L	4.0	07/26/24	07/31/24	2430162	

Haloacetic Acids Analyses

Dibromoacetic Acid	EPA 552.2	ND	ug/L	1.0	07/29/24	07/30/24	2431009	
Dichloroacetic Acid	EPA 552.2	ND	ug/L	1.0	07/29/24	07/30/24	2431009	
Monobromoacetic Acid	EPA 552.2	ND	ug/L	1.0	07/29/24	07/30/24	2431009	
Monochloroacetic Acid	EPA 552.2	ND	ug/L	2.0	07/29/24	07/30/24	2431009	
Trichloroacetic Acid	EPA 552.2	ND	ug/L	1.0	07/29/24	07/30/24	2431009	
Total Haloacetic Acids (HAA5)	EPA 552.2	ND	ug/L	1.0	07/29/24	07/30/24	2431009	
Surrogate: 2,3-Dibromopropionic Acid	EPA 552.2	73 %			07/29/24	07/30/24	2431009	

pH (Lab) was analyzed ASAP but received and analyzed past the 15 minute hold time.

Cl Res Total was analyzed ASAP but received and analyzed past the 15 minute hold time.

ND Analyte NOT DETECTED at or above the reporting limit

Bridget Durand

Project Manager