

Powerful filtering, excellent permeability AirQUEEN Nano Mask

Masks are a must!

Introducing easy-to-breathe mask with exceptional filter





**MADE IN
KOREA**



**Made with our own proud
technology-AirQUEEN Nano Mask**

- Nanofiber Filter Manufactured by: LEMON Co., Ltd.
- Mask Manufacture & Distribution: TOPTEC Co., Ltd.

Breathing New Technology, Nanofiber Nanofiber Filter

Nano fibers are only 1 nanometer, 1 billionth of a meter, in thickness. Nanofiber is a new material made by building these fibers sterically for a fishnet structure that has better permeability, blockability, and durability than any fiber ever invented.



Permeability



Blockability





Durability

Nanofiber Filter has a different fiber thickness / space between fibers / filtering method from the existing electrostatic filter

Melt-Blown Filter
5,000 times
enlarged

The fiber is thick
Average space between the
fibers is over 10 μ m



Nanofiber Filter
15,000 times
enlarged

Very thin, nano thickness
Average space between the fibers
is 1.0 μ m, with self-filtering

The Nanofiber Filter from LEMON Co., Ltd is a cutting-edge technology applied to global brands and products

It is exclusively
distributed to
THE NORTH FACE
as the brand **FUTURELIGHT**.

It is also used for
RENERGIE mask pack
from **LANCÔME** that touches
the skin directly.

This technology is used for
smart devices that sends
sounds but needs protection against
dust and moisture.



We asked professor Kim Iksu, the most renowned professor in the fiber research area.

What kind of material is **Nanofiber**?

The thickness of a nano fiber is only a thousandth of a strand of hair.
Nanofiber is a material made by building these fibers sterically for a fishnet structure,
and nanofiber of 85% processing rate has both permeability and blockability.


The Nanofibers of LEMON have consistent thickness and pore size.
This means the ratio of the pores in the general area becomes very high,
which is relevant to permeability.

Because it touches the **nose** and the **mouth**,
this fiber is **very appropriate for masks**
where permeability is important



We asked professor Kim Iksu, the most renowned professor in the fiber research area.

Is it possible to reuse **Nanofiber masks**?



When MB filter was sprayed with ethanol and reused, the filtration efficiency was reduced by up to 64%. On the other hand, the nanofiber filter retained almost the same efficiency as the first time even after 10 reuses. This difference is due to the decrease in static electricity in the filter after spraying. Unlike MB fibers, nanofibers are filtered according to the morphological characteristics and nano size spiracle of the surface, so efficiency is maintained after cleaning.

As a result of joint research by POSTECH and Shinshu University in Japan, it was found that the efficiency of nanofiber masks remains almost the same even after 10 reuses after spraying ethanol.

[Download PDF of research results : <https://pubs.acs.org/doi/10.1021/acsanm.0c01562>]



**Long-lasting freshness
from morning to evening!**

The effectiveness of electrostatic masks decreases with each breath due to the moisture
Nanofiber Filter has great permeability, so you can feel fresh even after a long term use



Excellent filtering and permeability of Nanofiber

Nanofiber smart filter
blocks pollutants and enables easier breathing

**If you care about the skin,
permeability is even more
important**

Wearing mask all day causes skin trouble (C Newspaper)

Must wear permeable masks (M Newspaper)

Pimple breaks out due to wearing masks
(National Health Insurance Service blog)

There is moisture in our respiratory system
If a mask has bad permeability, the moisture from within
cannot escape, which causes eczema

If you wear a mask for a long time, you really need a permeable mask

Because we need to wear it all day, <http://airqueenmask.com>
we made it lighter.

AirQUEEN Nano mask 4.38g

Lighter than a sheet of paper, so you won't even realize wearing it!



※ a sheet of paper = 4.72g

3D ergonomic design to fit perfectly on all face types.

Nose bridge prevents sliding or fogging up.


Nanofiber-
a breathing smart filter

Wide nose bridge

Comfortable elastic
ear band

High permeability
NO skin trouble

Light weight
for all-day wear





**Individually wrapped
for easy carrying and using**



CE FFP2

AirQUEEN Nano Mask is a FFP2 grade level mask that has officially been certified by European Commission



Certificate No: 2163-PPE-1433

Respiratory protective devices, filtering half masks to protect against particles manufactured by
TOPTEC CO., LTD.
140-22, Cheomdangieop 5-ro, Sandong-myeon, Gumi-si, Gyeongsangbuk-do, Republic of Korea
are tested and evaluated according to

**EN 149:2001 + A1:2009 Respiratory Protective Devices -
Filtering Half Masks to Protect Against Particles -
Requirements, Testing, Marking**

Product Definition

Brand Name: Air Queen **Model:** Breeze Mask
Filtering half mask
Classification: FFP2 NR

- Certificate Holder : TOPTEC CO., LTD.
- Certificate No : 2163-PPE-1433
- Model Name : Air Queen / Breeze Mask FFP2 NR

FDA Establishment Registration & Device Listing

THIS ACKNOWLEDGES THAT

TOPTEC CO., LTD.

WITH OWNER/OPERATOR NUMBER 10072713

HAS BEEN REGISTERED AS OWNER/OPERATOR FOR FOLLOWING ESTABLISHMENTS:

TOPTEC CO., LTD. located at 122 Asanvalley-Ro, Durnpo-Myeon, Asan-Si Chungcheongnamdo, KR 310409
Registration Number: 3016790437

NANO FILTER, INC. located at 3833 McGowan St. Long Beach , CA 90808
Registration Number: 3016766213

LIME CO., LTD. located at 212-27 Hapuri-1-Gil, Ujeong-Eup, Hwasong-Si Gyeonggido, KR 18561
Registration Number: Active and Awaiting Assignment of Registration Number

AND OWNER OF FOLLOWING DEVICE:

Proprietary Name: AIR QUEEN, AIR QUEEN BREEZE, NANO MASK, PURE MASK, TECHNO WEB, TECHNOWEB
Classification Name: MASK, SURGICAL
Product Code: FXX
Device Class: 2
Premarket Submission Number: K172500

THIS REGISTRATION IS VALID UNTIL 12/31/2020.

This document is no representation or warranty, nor does this document make any representation or warranty to any person or entity other than the named document holder. No value shall be placed on it. This document does not denote endorsement or approval of the document holder's device or establishment by the U.S. Food and Drug Administration. This document does not assure or guarantee to any person or entity in connection with the foregoing statements or provisions in paragraph 12 (2) (b) (2) (3). Registration of a device establishment or assignment of a registration number does not in any way denote approval of the establishment or its products. Any representation that creates an impression of official approval because of registration or presence of a registration or presentation of a registration number is misleading and constitutes misbranding. The U.S. Food and Drug Administration does not issue a certificate or registration document, nor does the U.S. Food and Drug Administration recognize a certificate or a registration document.

510(k) Premarket Notification

[FDA Home](#)
[Medical Devices](#)
[Databases](#)

[510\(k\) / Definition / Registration & Listing / Adverse Events / Recalls / PMA / RDE / Classification / Standards](#)
[CFR Title 21 / Regulation-Enabling Products / A-Play-Access / Medical Reports / CLIA / TPLs](#)


[New Search](#)
[Back To Search Results](#)

Device Classification Name	Utah Surgical
510(k) Number	K172050
Device Name	Technovue Surgical Mask
Applicant	YTS GLOBAL, INC 7400 ALBANY STATION CT STE 108 Springfield, VA, 22150
Applicant Contact	Eddie Nguyen
Correspondent	YTS IT GLOBAL, INC 7400 ALBANY STATION COURT SUITE A108 Springfield, VA, 22150
Correspondent Contact	Eddie Nguyen
Regulation Number	87A.650
Classification Product Code	230
Date Received	06/16/2017
Decision Date	03/02/2019
Decision	Substantially Equivalent (SESE)
Regulation Medical Specialty	General & Plastic Surgery
816a Review Panel	General Hospital
Summary	Summary
Type	Traditional
Reviewed By Third Party	No
Combination Product	No

- **Classification : Mask, Surgical**
- **Device Class : Class2**
- **510(K) Number : K172500**

PATENT

The Nanofiber Filter of AirQUEEN Nano Mask is manufactured with patent technology.

등록특허 10-1040055	
 (19) 대한민국특허청(KR) (12) 등록특허공보(B1)	
(45) 공고일자 2011년06월09일 (11) 등록번호 10-1040055 (24) 등록일자 2011년06월02일	
(51) Int. Cl. D01D 5/00 (2006.01) D01D 13/00 (2006.01) B82Y 40/00 (2011.01)	(73) 특허권자 신승 다이가루 일본 나가노켄 마쓰모토시 아사리 2초메 1번 1교 주식회사 토크 경상북도 구미시 산동면 봉산리 366번지
(21) 출원번호 10-2011-0015676 (22) 출원일자 2011년02월24일 실사청구일자 2011년02월24일	(72) 발명자 이재훈 경상북도 구미시 산동면 봉산리 366 (주)토크 김익수 T.38685567나가노현후에다시도키타3-15-1국립대학 법인인규대학실용학부소속
(30) 우선권주장 JP-P-2010-272071 2010년12월06일 일본(JP)	(74) 대리인 오광환
(56) 선행기술조사분원 JP2009052163 A KR1020070026744 A KR1020070047282 A JP2008519175 A	
전체 청구항 수 : 총 14 항	심사관 : 최봉준
(54) 전계방사장치 및 나노섬유 제조장치	
(57) 요약 노즐 끝쪽을 접지 한 상태에서 컬렉터에 고전압을 인가하여 전계 방사를 하는 경우라도, 균일한 품질을 가지는 나노 섬유를 높은 생산성으로 대량생산 하는 것이 가능한 전계방사장치를 제공한다. 컬렉터(110)와, 노즐 끝쪽(110)과, 정전극이 컬렉터(150)에 접속되고, 부전극이 노즐 끝쪽(110)에 접속되는 장치에 해당 부전극의 전위가 접지 전위로 떨어진 전원 장치(160)를 구비하는 전계방사장치(20)로서, 컬렉터(150)를 둘러싸는 위치에 회전 자유롭게 배열된 원연성 또는 다공성의 엔들리스 벨트로 이루어지는 보조 벨트(172) 및 해당 보조 벨트(172)를 장치시트(W)의 이동 속도에 대응하는 회전 속도로 회전시키는 보조 벨트 구동장치(174)를 가지는 보조 벨트 장치(170)를 추가로 구비하는 전계방사장치.	
대표도	
- 1 -	

- Patent Registration no. : 10-1040055
- Name of Invention:
An electrospinning device and an apparatus for manufacturing nano-fiber

Cutting-edge technology that mass-produces consistent nano fiber filters
Selected as one of 100 technologies to lead Korea in 2020
Approximately 50 patents around the world

ISO

The Nanofiber Filter of AirQUEEN Nano Mask is produced in a clean and sanitary facility.

Proven suitable for
ISO 9001
quality management
system



Proven suitable for
ISO 14001
environment
management system

The Nanofiber Filter used for AirQUEEN Masks is produced in a facility approved for environment management system, so it is safe and sanitary.

SGS

The Nanofiber Filter of AirQUEEN Nano Mask is approved by European SGS, REACH, and RoHS, and therefore is safe.

SGS
Test Report No. F690101/LF-CTSAIGA18-00327 Issued Date: 2018. 01. 16. Page 1 of 17

LEMON CO., LTD.
1105-05 Sanho-daein, Sandong-myeon
Gumi-si, Gyeongbuk
Korea

The following sample(s) was/were submitted and identified by/on behalf of the client as:

SGS File No. : AVGA18-00327
Product Name : Nano membrane (sanitary pad)
Item/Part Name : N/A
Received Date : 2018. 01. 09
Test Period : 2018. 01. 09. ~ 2018. 01. 16.
Test Requested : One hundred-Seventy four (174) substances in the Candidate List of Substances of Very High Concern (SVHC) for authorization published by European Chemicals Agency (ECHA) on July 7, 2017 regarding Regulation (EC) No 1907/2006 concerning the REACH.
Eight (8) substances in the Public Consultation List of potential Substances of Very High Concern (SVHC) published by European Chemicals Agency (ECHA) on Sep 5, 2017 regarding Regulation (EC) No 1907/2006 concerning the REACH.
Test Method : Please refer to next page(s).
Test Result(s) : Please refer to next page(s).
Summary : According to the specified scope and evaluation screening, the test results of SVHC are ≤ 0.1% (w/w) in the articles of the submitted sample.

SGS Korea Co., Ltd
Jeff Jang
Jeff Jang / Chemical Lab Mgr

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at www.sgs.com/sgschemical.
This document is the property of the Company and is loaned to the Client for their use only. It is not to be reproduced, stored in a retrieval system, or transmitted in any form or by any means, electronic, mechanical, photocopying, recording, or by any information storage and retrieval system, without the prior written approval of the Company. The Client agrees to indemnify and hold the Company harmless from all claims, damages, costs and expenses, including reasonable attorneys' fees, arising from the use of this document by the Client or any third party. The Client agrees to return this document to the Company upon request. The Client agrees to keep this document confidential and to use it only for the purpose for which it was provided. The Client agrees to notify the Company immediately in writing if it becomes aware of any unauthorized use of this document. The Client agrees to notify the Company immediately in writing if it becomes aware of any unauthorized use of this document. The Client agrees to notify the Company immediately in writing if it becomes aware of any unauthorized use of this document.

File version 3
SGS Korea Co., Ltd
SGS, The U.S. Korea Chemicals Laboratory, Gyeongbuk, Korea 14117
TEL: 053-603-1000 FAX: 053-603-1001 sgs@sgs.com
Branch of the SGS Group (Global Laboratory for Environment)

**RoHS**
Metal
material and
harmful
compound
not detected



Organic
solvent and
compound
not detected

- Cadmium N/D
- Lead N/D
- Mercury N/D
- Hexavalent chromium N/D
- Other 182 categories N/D

KAKEN

Nanofiber filter is approved
its safety by KAKEN which is
Japanese Certification Authority.

試験報告書

試験結果

試験項目	試験結果
圧力損失	0.15
透過率	99.9
静電抵抗	1.0E+10

以上

試験報告書

試験結果

試験項目	試験結果
圧力損失	0.15
透過率	99.9
静電抵抗	1.0E+10

以上

試験報告書

試験結果

試験項目	試験結果
圧力損失	0.15
透過率	99.9
静電抵抗	1.0E+10

以上

試験報告書

試験結果

試験項目	試験結果
圧力損失	0.15
透過率	99.9
静電抵抗	1.0E+10

以上

Approved for safety by KAKEN TEST CENTER which is a Global Certification Authority evaluate a quality and a performance of textile products, household goods, and functional high-tech materials as an analyze center of JSTIIF.

KCL

This is an industrial product that passed the tests of KCL and is [**currently in process**] of being approved by the MFDS.

confidential

Applied for the KF approval at the MFDS early April of 2020 after passing the test done by KCL, a testing institute designated by the MFDS.

It is manufactured with the same technology as Technoweb Disposable dustproof mask (KF94) from FTENE, the subsidiary company of TOPTEC, but if the manufacturer or brand is changed, the approval from the MFDS needs to be renewed, and until then, it cannot be publically distributed.

However, after the KF certification, 80% of the supply will be distributed as public, so there might be a shortage.

Fiti

As a result of testing at Fiti,
a testing agency designated
by MFDS and obtained an appropriate
level of evaluation for KF94 in the test.

fiti (00511) 302-No. 254, Beokkot-ro, Guro-gu, Seoul, Korea
Tel: 02-2113-4138 Fax: 02-2113-8130

TEST REPORT **FOR REFERENCE ONLY**

APPLICANT : SOOMLAB

REPORT NO. : N271-20-00801
SAMPLE RECEIVED DATE : 2020-04-17
TEST STARTED DATE : 2020-04-17
REPORT ISSUED DATE : 2020-05-13
PAGE : 1 OF 5

DESCRIPTION : ONE(1) PIECE OF SUBMITTED CUTTING SAID TO BE MASK.

ITEM : AIRQUEEN NANO MASK, NANOFIBER FILTER

TEST CONDUCTED : AS REQUESTED BY THE APPLICANT, FOR DETAILS PLEASE SEE ATTACHED PAGES.

PREPARED AND CHECKED BY
FOR FITI
Kwang Jae Park
GWANG JAE, PARK
QUALITY MANAGER

AUTHORIZED BY
FOR FITI
Jun Se Jun
JE-GOO JUN
PRESIDENT

■ Report Verification No.: Y4V9-GL2E-LDRU ■
(You can see the authenticity of your test report through the above "Report Verification No." at FITI homepage.)

DOCUMENT SERVICE
The test results contained in this report are limited to results on the sample(s) that is provided by client and are not necessarily evaluative or representative of the quality of the lot from which the sample(s) was taken or of all products. Results contained in this report are not based on the quality verification of sample by the FITI quality verification program unless specifically requested by the client. Further use of the results of this report is prohibited unless allowed under a separate agreement and both is an official document that is established between the client specified on this label and the FITI. This test report is irrelevant to ISO 15189:2013 and ISO 17025 accreditation.

fiti (00511) 302-No. 254, Beokkot-ro, Guro-gu, Seoul, Korea
Tel: 02-2113-4138 Fax: 02-2113-8130

REPORT NO.: N271-20-00801
PAGE: 2 OF 5

01. EFFICIENT DUST COLLECTION
(GUIDELINES FOR REFERENCE STANDARDS FOR HEALTH MASKS,
FOOD AND DRUG SAFETY (2018.12.)) %

ROOM TEMPERATURE	#1
-1	95.7
-2	95.3
-3	95.2
ROOM TEMPERATURE + 4 HOURS AFTER PRE-TREATMENT (20 ± 0.1, 25 ± 0.1, 30 ± 0.1, 35 ± 0.1, 40 ± 0.1, 45 ± 0.1, 50 ± 0.1, 55 ± 0.1, 60 ± 0.1, 65 ± 0.1, 70 ± 0.1, 75 ± 0.1, 80 ± 0.1, 85 ± 0.1, 90 ± 0.1, 95 ± 0.1, 100 ± 0.1)	95.7
-1	95.7
-2	95.3
-3	95.2

NOTE: THIS STANDARD TEST METHOD WAS APPLIED BY THE CLIENT'S REQUEST.

02. MASK INHALATION RESISTANCE TEST
(GUIDELINES FOR REFERENCE STANDARDS FOR HEALTH MASKS,
FOOD AND DRUG SAFETY (2018.12.)) %

ROOM TEMPERATURE	#1
-1	23
-2	24
-3	24
ROOM TEMPERATURE + 4 HOURS AFTER PRE-TREATMENT (20 ± 0.1, 25 ± 0.1, 30 ± 0.1, 35 ± 0.1, 40 ± 0.1, 45 ± 0.1, 50 ± 0.1, 55 ± 0.1, 60 ± 0.1, 65 ± 0.1, 70 ± 0.1, 75 ± 0.1, 80 ± 0.1, 85 ± 0.1, 90 ± 0.1, 95 ± 0.1, 100 ± 0.1)	23
-1	23
-2	24
-3	24

NOTE: THIS STANDARD TEST METHOD WAS APPLIED BY THE CLIENT'S REQUEST.

03. TENSILE STRENGTH
(GUIDELINES FOR REFERENCE STANDARDS FOR HEALTH MASKS,
FOOD AND DRUG SAFETY (2018.12.)) %

	#1
	23

NOTE: THIS STANDARD TEST METHOD WAS APPLIED BY THE CLIENT'S REQUEST.

04. PHASE
(GUIDELINES FOR REFERENCE STANDARDS FOR HEALTH MASKS,
FOOD AND DRUG SAFETY (2018.12.)) %

	#1
	23

NOTE: THIS STANDARD TEST METHOD WAS APPLIED BY THE CLIENT'S REQUEST.

05. ACID OR ALKALINE
(GUIDELINES FOR REFERENCE STANDARDS FOR HEALTH MASKS,
FOOD AND DRUG SAFETY (2018.12.))

METHOD	#1
METHOD: ORANGE FLUID	NO ORANGE COLOR DEVELOPS
METHOD: PHENOL FLUID	NO RED COLOR DEVELOPS

07. DYES
(GUIDELINES FOR REFERENCE STANDARDS FOR HEALTH MASKS,
FOOD AND DRUG SAFETY (2018.12.))

	#1
	NO COLOR DEVELOPS

08. FLUORESCENCE
(GUIDELINES FOR REFERENCE STANDARDS FOR HEALTH MASKS,
FOOD AND DRUG SAFETY (2018.12.))

	#1
	NOT VISIBLE FLUORESCENCE

09. FORMALDEHYDE
(GUIDELINES FOR REFERENCE STANDARDS FOR HEALTH MASKS,
FOOD AND DRUG SAFETY (2018.12.))

	#1
	THE COLOR OF THE SAMPLE SOLUTION IS NOT DARKER THAN THAT OF THE COMPARATIVE SOLUTION

DOCUMENT SERVICE
The test results contained in this report are limited to results on the sample(s) that is provided by client and are not necessarily evaluative or representative of the quality of the lot from which the sample(s) was taken or of all products. Results contained in this report are not based on the quality verification of sample by the FITI quality verification program unless specifically requested by the client. Further use of the results of this report is prohibited unless allowed under a separate agreement and both is an official document that is established between the client specified on this label and the FITI. This test report is irrelevant to ISO 15189:2013 and ISO 17025 accreditation.

Obtained an appropriate level of evaluation for KF94 in test items such as dust
collection efficiency (sodium chloride), dust collection efficiency (paraffin oil),
and mask inhalation resistance test.

NIOSH

AirQUEEN Nano Mask has passed every conformity evaluation test for N95 and CE FFP2 certificates.

Filter efficiency meets all of the N95(NIOSH) and CE FFP2 performance as 96.03%~99.694% and it has been found to have excellent functions in inhalation resistance(under 35mmH2O standard 8.2~8.8mmH2O) and exhalation resistance(under 25mmH2O 7.9~8.3mmH2O standard), making it a differentiated breathing convenience with the filter function of the mask.

 **Nelson Labs**
A Safety Health Company

Sponsor:
Kyunghan Chung
Toptec Co., Ltd
140-22, Cheomdangseop 5-ro, Sandong-myeon
Gumi-si, Gyeongangbuk-do, 39171
KOREA, REPUBLIC OF

Sodium Chloride (NaCl) Aerosol Test Final Report

Test Article: Air Queen Breeze Mask
Study Number: 1295789-001
Study Received Date: 04 May 2020
Testing Facility: Nelson Laboratories, LLC
6280 S. Redwood Rd.
Salt Lake City, UT 84123 U.S.A.
Test Procedure(s): Standard Test Protocol (STP) Number: STP0014 Rev 09
Deviation(s): None

Summary: This procedure was performed to evaluate particulate filter penetration as specified in 42 CFR Part 84 and TEB-APR-STP-0059 for requirements on a N95 respirator. Respirators were conditioned then tested for particle penetration against a polydispersed, sodium chloride (NaCl) particulate aerosol. The challenge aerosol was dried, neutralized, and passed through the test article at a concentration not exceeding 200 mg/m³. The initial airflow resistance and particle penetration for each respirator was determined.

According to 42 CFR Part 84.64, pretesting must be performed by all applicants as part of the application process with NIOSH. Results seen below are part of that pretesting and must be submitted to and accepted by NIOSH for respirator approval.

All test method acceptance criteria were met. Testing was performed in compliance with US FDA good manufacturing practice (GMP) regulations 21 CFR Parts 210, 211 and 820.


 

Robert Decker electronically approved for
Study Director
Curtis Garow
02 Jun 2020 15:37 (+00:00)
Study Completion Date and Time

801-290-7008 | nelsonlabs.com | sales@nelsonlabs.com  NHTSA-002 Rev 0
Page 1 of 3

Please refer to applicable standards referenced within this report for details on test methods and procedures. Results are for informational purposes only and do not constitute a warranty or guarantee of performance.

DMT - Prüf- und Ergebnisprotokoll


DMT

Prüfung Corona SARS-Cov-2 Virus Pandemie Atemschutzmasken

auf Grundlage Prüfgrundsatz Rev. 1 vom 28.02.2020 - erstellt von der DKVRA Testing and Certification GmbH und dem Institut für Arbeitsschutz (IFA) der Deutschen gesetzlichen Unfallversicherung


Allgemeines


Antrag Nr.	Prüfnummer	Datum Prüfungstermin
80020191225-1 Rev. 01	13.05. - 15.05.2020	08.05.2020

Prüfung bewirkt durch: ISO 9001 CERT GmbH Sengenthalstraße 32 41141 Essen	Erzeuger: Ingenieur Deutschland GmbH Friedenstraße 98 51194 Köln	Untersucht Prober: 19.05.2020 <i>g.h.</i> (Gibber)	Untersucht Leder Prüferbeil: 19.05.2020 <i>g.h.</i> (Schönberg)
--	---	---	--

Angaben zum Prüfing

Modell / Produktbezeichnung (*)	MNS
Ausführung	Nasenbügel, Ohrenschlaufen
Hersteller	TOPTEC Co. Ltd.
Modell-/Produktbezeichnung	Air Queen Nanofiber Filter Mask
Ident.-Nr. / Artikelnummer	1. A.
Lieferumfang	
Verpackungseinheit	Kunststoffbeutel
Verpackungseinheit:	1 Stück
Informationsbezeichnung:	auf Verpackung gedruckt





Prüfungen und Test-Ergebnisse

1. Siebprüfung		2. Ansiebprüfung	
Kriterium	Bewertung	Kriterium	Bewertung
Verpackungseinheit	Kunststoffbeutel	CPA nicht anzuwenden	I. O.
Verpackungseinheit:	1 Stück	CPA leicht abzuwischen	I. O.
Verpackungseinheit:	I. O.	Konform der Kopfformel	I. O.
Verpackungseinheit:	I. O.	Gesichtsfeld	I. O.
Zustand CPA:	I. O.	offensichtliche Undichtigkeiten:	nein
Beurteilung:	bestanden	wahrscheinliche Leckagezustand:	bestanden
3. Zustand nach Temperaturkonditionierung (24 h bei 70°C)			
Zustand CPA:	formstabil, Zustand unbeachtet	5. Durchlass-Prüfung (Max.-Wert Prüfgröße 1-3) Parafilm-Konzentration nach Maske + Ranges C ₁ (mg/m³) Parafilm-Konzentration vor Filter - Ranges C ₂ (mg/m³) Durchlasszahl (P ₂) Pross. < 0,1 bestanden	
Beurteilung:	bestanden		
4. Zustand nach Gebrauchssimulation			
Zustand CPA:	formstabil, Zustand unbeachtet	6. Siebprüfung Siebtestergebnis: bestanden Beurteilung: bestanden	
Beurteilung:	bestanden		

Prüfung:		Ergebnis:	Beurteilung:
Eingabe: Zugscherm auf Ventil mit 10 N:		hier nicht anwendbar	-
Funktionsprüfung nach Belastung bei 300 l/min:		hier nicht anwendbar	-

7. Prüfung des Atemschutzstandes CPA			8. Prüfung der Kennzeichnung		
Kriterium	Prüfung	Beurteilung	Kriterium	Prüfung	Beurteilung
Max.-Wert Ausatemstrom: bei 180 l/min (Schleifst 5 min)					
Blattnichtung:	Druckverlust	Beurteilung	Herstellerkennzeichnung:	TOPTEC Co. Ltd.	o. k.
gerade aus	182 Pa	bestanden	Vermerkennzeichnung:	MNS	o. k.
nach links	-	bestanden	Informationsdruck:	auf Verpackung gedruckt	o. k.
nach oben	-	bestanden	Aufdruck:	auf Verpackung gedruckt	o. k.
nach unten	-	bestanden	Hinweise zur Verwendung:	auf Verpackung gedruckt	o. k.
Max.-Wert Einatemstrom: bei 95 l/min (Schleifst 4 min)					
Blattnichtung:	Druckverlust	Beurteilung	Beurteilung:	bestanden	
gerade aus	182 Pa	bestanden			
Gesamtbewertung der geprüften CPA:			bestanden		

DMT GmbH & Co. KG
Am TÜV 1
40372 Essen 1 Germany
TÜV NORD GROUP

5. Durchlass-Prüfung (Max.-Wert Prüflinge 1-3)	
Paraffinöl-Konzentration nach Maske - Reingas C ₂ [mg/m³]:	0,42
Paraffinöl-Konzentration vor Filter - Rohgas C ₁ [mg/m³]:	19,85
Durchlassgrad P [%]:	2,1
Sollwertabgleich:	P _{max} < 6%
Beurteilung:	bestanden
Beurteilung:	bestanden

AirQUEEN Nano Mask
has passed the test by certified
European facility TUV NORD.

TUV NORD Group is a technical service provider with worldwide activities. Founded in 1869 and headquartered in Hanover, Germany.

“bestanden = pass”

HALAL

Korea Halal Authority certifies the AirQUEEN.



Certification No : KHA-20F-00362904

Name of Products : Air Queen / Nano Mask

Certification standard : KHAS 29000-General Standards for HALAL Industrial Products

[illegible]

Quality Indication According to Electric Products and Household Safety Law

- Product: AirQUEEN Mask
- Manufactured in: Korea
- Manufacturer: TOPTEC
- Distributor: TOPTEC
- Weight: 4.38g
- Product Type: Mask
- Expiry Date: 36 months
- Storage Directions: closed container, in room temperature (1~30°C)
- Entire Ingredients:
felt (outer fabric, filter, inner fabric), plastic sheath wire, polypropylene ring, nylon strap











AirQUEEN Mask is an industrial product.

<http://airqueenmask.com>

AirQUEEN Mask uses the Nanofiber Filter technology of LEMON and is manufactured by TOPTEC.
Currently, the MFDS approval is being processed,
and when it is approved as sanitation mask,
80% of the product supply will be distributed as public distribution.



Masks are a must! Prepare in advance with AirQUEEN Nano Mask

SOOMLAB is a partner of LEMON that joint-developed the Hyper Purifying Breathing Mask which is SOOMLAB's unique brand that has exclusive sales right.

SOOMLAB is a primary sales company of AirQUEEN Nano mask and authorized as a distributor by TOPTEC, a parent company of LEMON.

SOOMLAB and BOOSTERZ are in charge of marketing, distribution, and sales of AirQUEEN Nano Mask.

[AirQUEEN Mask mass purchase inquiry]

Email : contact@soomlab.co.kr

Store : <http://airqueenmask.com>

