

Carestream NDT

5 (10 Gallon) Compartment Developing Tank

Manual Processing Cycle (INDUSTREX Film & Chemistry)

	Developer	Stop Bath	Fixer	Wash	Rinse Solution
Time/Temperature	5 Minutes - 68°F 4 Minutes - 72°F 3 Minutes - 75°F 2 Minutes - 79°F 1.5 Minutes - 82°F	30 Seconds	5 - 7 Minutes	10 to 30 Minutes in running water (8 volume changes/hour)	30 Seconds
Notes	Pull and Plunge Film Directly from Developer into Stop Bath. DO NOT ALLOW excess Developer to Drain Back into Developer Tank. Replenishment: Fill the Developer Tank to its original level AS NEEDED (topping off) with Developer Replenisher Solution.	Back into Stop Bath Tank Before Moving Over to Fixer Tank. Replenishment: NOT REQUIRED. Fill the Stop	Pull and Plunge Film Directly from Fixer into Wash. DO NOT ALLOW excess Fixer to Drain Back into Fixer Tank. Replenishment: Fill the Fixer Tank to its original level AS NEEDED (topping off) with Fixer Replenisher Solution.	Replenishment: NOT REQUIRED. Fill the Wash Tank to its original level AS NEEDED (topping off) with water.	Replenishment: NOT REQUIRED. Fill the Final Rinse Tank to its original level as needed (topping off) with water.
N	Additional: Store Developer Replenisher Solution in a closed airtight container. * Stir solutions after each addition	Additional: This odorless stop bath features an indicator that changes color (from orange-yellow to magentared) to alert the user of declining activity, so that the solution can be changed or additional concentrate added.	Additional: Store Fixer Replenisher Solution in a closed airtight container.		

^{*} Discard solutions after adding two tank volumes of replenisher to the tank, or at least once per month, and then refill with fresh solution.