

Certificate of Calibration

Customer: ILRT, INC.

17880 ALTHEA BLUE PLACE
LUTZ FL 33558-6056
(877) 333-4578

P.O. Number: PO220123A R0

ID Number: II-0141



Description: TEMP/HUMIDITY SENSOR
Manufacturer: SENSIRION
Model Number: SHT75
Serial Number: NSN
Technician: TAYLOR FLOYD

Calibration Date: 02/01/2022
Calibration Due: 02/01/2023
Procedure: TMI-HYGROTHERMOGRAPHS
Rev: 8/24/2020
Temperature: 71 °F
Humidity: 35 % RH

On-Site Calibration:
Comments:

As Found Condition: IN TOLERANCE
Calibration Results: IN TOLERANCE

Limiting Attribute:

This instrument has been calibrated using standards traceable to the SI units through the National Institute of Standards and Technology (NIST) or other National Metrological Institute (NMI). The method of calibration is direct comparison to a known standard, derived from natural physical constants, ratio measurements or compared to consensus standards.


Reported uncertainties are expressed as expanded uncertainty values at an approximately 95% confidence level using a coverage factor of k=2. Statements of compliance are based on test results falling within specified limits with no reduction by the uncertainty of the measurement.

TMI's Quality System is accredited to ISO/IEC 17025:2017 and ANSI/NCSL Z540-1-1994. ISO/IEC 17025:2017 is written in a language relevant to laboratory operations, meeting the principles of ISO 9001 and aligned with its pertinent requirements. This calibration is within the current Scope of Accreditation and complies with the requirements of ISO/IEC 17025:2017 and TMI's Quality Manual, QM-1.

Results contained in this document relate only to the item calibrated. Calibration due dates appearing on the certificate or label are determined by the client for administrative purposes and do not imply continued conformance to specifications.

This certificate shall not be reproduced, except in full, without the written permission of Technical Maintenance, Inc. Measurements not currently on TMI's Scope of Accreditation are identified with an asterisk.


WALLY GYNN, BRANCH MANAGER


Scott Chamberlain, QUALITY MANAGER

Calibration Standards

<u>Asset Number</u>	<u>Manufacturer</u>	<u>Model Number</u>	<u>Date Calibrated</u>	<u>Cal Due</u>
0710649	THUNDER SCIENTIFIC	2500ST	8/30/2021	6/10/2022
660TL18010015	ADDITEL CORPORATION	ADT875PC-155	5/20/2021	5/20/2022
71512	FLUKE	5616-12	12/22/2021	12/22/2022
A88072	FLUKE/HART	1502A	9/27/2021	2/7/2022



Technical Maintenance, Inc.

12530 TELECOM DRIVE, TEMPLE TERRACE, FL 33637

Phone: 813-978-3054 Fax 813-978-3758

www.tmicalibration.com



AC-2080

Certificate of Calibration

Data Sheet

<u>Parameter</u>	<u>Nominal</u>	<u>Minimum</u>	<u>Maximum</u>	<u>As Found</u>	<u>As Left</u>	<u>Uncertainty</u>	<u>Unit</u>	<u>ADJ/FAIL</u>
Temperature Accuracy	0.00	-1.00	1.00	0.22	0.22	0.042°F	F	
Temperature Accuracy	30.00	29.00	31.00	30.16	30.16	0.042°F	F	
Temperature Accuracy	60.00	59.00	61.00	60.08	60.08	0.042°F	F	
Temperature Accuracy	90.00	89.00	91.00	89.92	89.92	0.042°F	F	
Temperature Accuracy	120.00	119.00	121.00	119.79	119.79	0.042°F	F	
Humidity Accuracy	10.00	6.50	13.50	12.45	12.45	0.51%RH	%	
Humidity Accuracy	33.00	29.50	36.50	35.22	35.22	0.51%RH	%	
Humidity Accuracy	50.00	46.50	53.50	51.11	51.11	0.51%RH	%	
Humidity Accuracy	75.00	71.50	78.50	72.60	72.60	0.51%RH	%	
Humidity Accuracy	90.00	86.50	93.50	86.98	86.98	0.51%RH	%	



Technical Maintenance, Inc.

12530 TELECOM DRIVE, TEMPLE TERRACE, FL 33637

Phone: 813-978-3054 Fax 813-978-3758

www.tmicalibration.com



AC-2080