



CERTIFICATE OF NIST TRACEABLE CALIBRATION

Calibration Certificate No: 95557

Customer Information

Customer: ILRT Inc.
Address : 237 Phinney Road
Hannibal, NY 13074



Customer PO #: PO220104A R0

Calibration Procedure Information

Procedure ID: GTP Graftel-Temp

Revision #: 11

Revision Date: 4/15/2019

Calibration Standards Information

<u>Graftel ID</u>	<u>Manufacturer</u>	<u>Model #</u>	<u>Description</u>	<u>CAL Due</u>
10017	Hart Scientific/Burns	1502A/12005	PRT, Temperature	10/29/2022
10129	Hart Scientific	1502A/5628	PRT, Temperature	5/8/2022
60030	Paroscientific	760-100A	Pressure, 100 psia	5/20/2022
T1830459	Vaisala	HMW95D	RH/Temp. Logger	6/17/2022
1A01JMGKP36	Graftel	N/A	Digital Barometer	6/16/2022

Sensor Information

Manufacturer: Graftel

Description: Temp. Sensor

Method Used: Temp. Bath/PRT

Model #: 9202

Rated Accuracy: ± 1 Difference

Accuracy Specified By: Graftel

Instrument ID#: II-0106

Range: 50 to 150 °F

Condition: Functional

Serial #: 0392020-2

Comments: Calibration Date: 01/31/2022

Calibration Due: 01/31/2023

This calibration is Nuclear Safety Related. 10CFR50 Appendix B and 10CFR Part 21 apply.

The calibrations within the certificate/report are traceable through NIST or another National Metrology Institute to the International System of Units (SI). The reported calibration uncertainty has a confidence level of 95% ($k=2$). A calibration uncertainty ratio of 4:1 was maintained unless required uncertainty is supported by analysis. Graftel Quality Assurance System complies with applicable requirements of ISO/IEC-17025-2017, ANSI/NCSL Z540-I-1994 and ISO 9001. All results contained within this certificate relate only to item(s) calibrated. This certificate shall not be reproduced except in full and with the written consent of Graftel. Acceptance Criteria per Simple Acceptance Rule: Measurement Uncertainty is not applied to the measured value when in/out of tolerance statement is made.

Performed By:

Philip Davis
Calibration Technician

Date: 1/31/2022

Approved By:

Scott Pickett
Vice President, Lab Services


Date: 1/31/2022

**ATTACHMENT TO CALIBRATION CERTIFICATE 95557
AS FOUND / AS LEFT DATA**

Page 2 of 2

Reading From Standard, °F	Lower Limit of Meter Reading, °F	Measured Reading From Meter, °F	Upper Limit of Meter Reading, °F	Error, °F	Measurement Uncertainty (k=2) °F	CMC (k=2) °F	STATUS
50.054	49.054	50.20	51.054	0.146	0.05	0.05	Pass
74.991	73.991	75.08	75.991	0.089	0.05	0.05	Pass
99.982	98.982	100.12	100.982	0.138	0.05	0.05	Pass
124.989	123.989	125.19	125.989	0.201	0.05	0.05	Pass
149.989	148.989	150.21	150.989	0.221	0.05	0.05	Pass

ERROR CHART



Instrument Specifications

Lower Range:	50	°F
Upper Range:	150	°F
Resolution:	0.01	°F
Rated Accuracy:	1	Difference

Laboratory Ambient Conditions

Pressure:	14.41	psia
Humidity:	15.51	%RH
Temperature:	70.62	°F



FLOW - TEMPERATURE - HUMIDITY - PRESSURE - DESIGN - CONSULTING - ENGINEERING

NIST Traceable Calibration Data Sheet

WWW.GRAFTEL.COM

95 Chancellor Dr., Roselle, IL 60172

Phone: 847-364-2600

Fax: 847-364-3899