


FOLLOW UP CASES

Interesting Data Downloads & Stuff
seen in Follow up.




Compliance Report

Usage		05/09/2017 - 06/07/2017
Usage days		29/30 days (97%)
>= 4 hours		22 days (73%)
< 4 hours		7 days (23%)
Usage hours		178 hours 22 minutes
Average usage (total days)		6 hours 57 minutes
Average usage (days used)		8 hours 9 minutes
Median usage (days used)		6 hours 17 minutes

AirSense 10 AutoSet	
Serial number	
Mode	CPAP
Set pressure	7 cmH ₂ O
EPR	Fulltime
EPR level	2

Therapy						
Leakage - L/min	Median:	0.0	95th percentile:	3.0	Maximum:	17.5
Events per hour	AI:	14.7	HI:	1.8	AHI:	18.5
Apnea Index	Central:	8.2	Obstructive:	5.8	Unknown:	0.6
RERA Index						2.0
Cheyne-Stokes respiration (average duration per night)						20 minutes (6%)

Usage - hours





ResMed

AirView™

PAD1402
3777 JAS1 v8v10 07/01/09/10
PAD1402
Pulmon, 03036

Phone: 0800-362-0182
Fax: 0800-362-1012
Email: sales.view@resmed.co.uk

06/04/2017 - 06/06/2017

DOB: [REDACTED] 1936

Age: 80 years


Compliance Report

Usage		06/04/2017 - 06/06/2017
Usage days		3/3 days (100%)
>= 4 hours		3 days (100%)
< 4 hours		0 days (0%)
Usage hours		23 hours 46 minutes
Average usage (total days)		7 hours 55 minutes
Average usage (days used)		7 hours 55 minutes
Median usage (days used)		7 hours 54 minutes

AirSense 10 AutoSet	
Serial number	[REDACTED]
Mode	CPAP
Set pressure	7 cmH2O
EPR	Fulltime
EPR level	2

Therapy					
Leak - L/min	Median:	0.0	95th percentile:	2.8	Maximum: 10.8
Events per hour	AI:	5.4	HI:	1.1	AHI: 6.5
Apnea Index	Central:	3.0	Obstructive:	2.4	Unknown: 0.0
RERA Index					2.0
Cheyne-Stokes respiration (average duration per night)					25 minutes (5%)

Usage - hours





AirView™

Bernie Miller

Logout

Help

Patients

My profile

Administration

Create report



Card downloaded



Patient ID

00/05/1970
D.O.B.

02/10/2010
Setup date

Compliant

97%
Last 30

AutoSet:
AirSense 10 AutoSet

Back to patients

Charts Patient details Prescription Remote Assist Notes Logs Thresholds

Fixed time period




00

day period ending


05/05/2017

Update

Compliant use Non-compliant use



100%



ResMed AirView™

[Create report](#) [Card downloaded](#)


[Back to patients](#)

Patient ID 10/00/1873 D.O.B. 02/24/2017 Setup date ☒ Compliant 30% Last 30 AutoSet AirSense 10 AutoSet

[Charts](#) [Patient details](#) [Prescription](#) [Remote Assist](#) [Notes](#) [Logs](#) [Thresholds](#)

Fixed time period 90 day period ending 05/24/2017

☒ Compliant use ☐ Non-compliant use



[Back to patients](#)

Patient ID

/1947

D.O.B.

12/15/2014

Setup date



Compliant

100%

Last 30

AutoSet

S9 AutoSet

Charts

Patient details

Prescription

Notes

Logs

Thresholds

Fixed time period



90

day period ending


06/07/2016

Update

Compliant use

Non-compliant use

Usage



[Create report](#)

Card download



--

[Back to patients](#)

Patient ID

12/27/1951
D.O.B.

04/21/2017
Setup date

Compliant

100%
Last 30

AutoSet for Her

AutoSet for Her

AirSense 10 AutoSet For Her

Charts

Patient details

Prescription

Remote Assist

Notes

Logs

Thresholds

Fixed time period



30

day period ending


05/31/2017

Update

- Compliant use

- Non-compliant use

Usage





FEWPE
444 W 2nd St
Suite 103
Ft Worth
TX 76102, 81728
Phone: 480-835-9100
Fax: 480-835-9104
Email: c@200@madonast.com

[Redacted]

05/02/2017 - 05/31/2017

DOB [Redacted]/1951

Age: 65 years


Compliance Report

Usage		05/02/2017 - 05/31/2017
Usage days		30/30 days (100%)
>= 4 hours		30 days (100%)
< 4 hours		0 days (0%)
Usage hours		265 hours 9 minutes
Average usage (total days)		8 hours 50 minutes
Average usage (days used)		8 hours 50 minutes
Median usage (days used)		9 hours 7 minutes

AirSense 10 AutoSet For Her	
Serial number	[Redacted]
Mode	AutoSet for Her
Min Pressure	5 cmH2O
Max Pressure	14 cmH2O
EPR	Fulltime
EPR level	3

Therapy				
Pressure - cmH2O	Median: 10.6	95th percentile: 12.3	Maximum: 13.1	
Leaks - L/min	Median: 0.0	95th percentile: 5.1	Maximum: 82.9	
Events per hour	AI: 10.8	HI: 1.2	AHI: 12.0	
Apnea Index	Central: 8.3	Obstructive: 2.2	Unknown: 0.2	
RERA Index				0.1
Cheyne-Stokes respiration (average duration per night)				0 minutes (0%)

Usage - hours





AirView™

TEWPE
444 W 2nd St
Suite 103
Tempe
Arizona, 85281
Phone: 480-835-9100
Fax: 480-835-9104
Email: cp-ano@resmed.com

05/02/2017 - 05/31/2017


DOB: 7/1951
Age: 65 years

Therapy Report


AirSense 10 AutoSet For Her

St.


Usage (hours)




Leak (L/min)



Pressure (cmH2O)



AHI (events/hour)





ResMed

AirView™

Female
444 W 2nd St
Suite 103
Anaheim, CA 92801
Phone: 480-835-9100
Fax: 480-835-9104
Email: info@resmed.com

05/22/2017 - 05/27/2017

DOB: [REDACTED] /1951

Age: 65 years

Compliance Report

Usage		05/22/2017 - 05/27/2017
Usage days		6/6 days (100%)
>= 4 hours		6 days (100%)
< 4 hours		0 days (0%)
Usage hours		48 hours 37 minutes
Average usage (total days)		8 hours 6 minutes
Average usage (days used)		8 hours 6 minutes
Median usage (days used)		7 hours 43 minutes


AirSense 10 AutoSet For Her

Serial number	[REDACTED]
Mode	AutoSet for Her
Min Pressure	5 cmH2O
Max Pressure	14 cmH2O
EPR	Fulltime
EPR level	3

Therapy

Pressure - cmH2O	Median: 11.3	95th percentile: 12.7	Maximum: 13.3
Leaks - L/min	Median: 0.0	95th percentile: 1.4	Maximum: 74.6
Events per hour	AI: 33.5	HI: 1.6	AHI: 35.1
Apnea Index	Central: 30.4	Obstructive: 2.6	Unknown: 0.4
RERA Index			0.1
Cheyne-Stokes respiration (average duration per night)			0 minutes (0%)

Usage - hours





AirView™

FEU PE
444 W 2nd St
Suite 103
Fremont
Alameda, 94529
Phone: 480-835-9100
Fax: 480-835-9104
Email: cp-acc@resmed.com

05/22/2017 - 05/27/2017


D 1951
Age: 65 years

Therapy Report


AirSense 10 AutoSet For Her

SN: [REDACTED]


Usage (hours)




Leak (L/min)



Pressure (cmH2O)



AHI (events/hour)





FEW PE
 444 W 2nd St
 Suite 103
 Tempe
 Arizona, 85281
 Phone: 480-835-9100
 Fax: 480-835-9104
 Email: cp-2006@resmedusa.com

04/22/2017 - 05/21/2017

DOB: 12/27/1951

Age

Compliance Report

Usage	04/22/2017 - 05/21/2017
Usage days	30/30 days (100%)
>= 4 hours	30 days (100%)
< 4 hours	0 days (0%)
Usage hours	260 hours 2 minutes
Average usage (total days)	8 hours 40 minutes
Average usage (days used)	8 hours 40 minutes
Median usage (days used)	8 hours 32 minutes


AirSense 10 AutoSet For Her

Serial number	8
Mode	AutoSet for Her
Min Pressure	5 cmH2O
Max Pressure	14 cmH2O
EPR	Fulltime
EPR level	3

Therapy

Pressure - cmH2O	Median: 10.6	95th percentile: 12.3	Maximum: 13.1
Leaks - L/min	Median: 0.0	95th percentile: 3.5	Maximum: 83.1
Events per hour	AI: 5.0	HI: 1.2	AHI: 6.2
Apnea Index	Central: 2.7	Obstructive: 2.1	Unknown: 0.1
RERA Index	0.1		
Cheyne-Stokes respiration (average duration per night)	0 minutes (0%)		

Usage - hours





AirView™

TEW PE
date: 04/22/2017
Site: 103
Name:
Address: 8528
Phone: 480-835-9100
Fax: 480-835-9104
Email: cpg@resmed.com

04/22/2017 - 05/21/2017


DOB: 951
Age: 65 years

Therapy Report


AirSense 10 AutoSet For Her

SN: 2


Usage (hours)




Leak (L/min)



Pressure (cmH2O)



AHI (events/hour)





ResMed
 440 W 2nd St
 Suite 100
 Fremont, CA 94536
 Phone: 480-835-9100
 Fax: 480-835-9104
 Email: cpap@resmed.com

04/22/2017 - 05/21/2017
 ID: [REDACTED] 951
 Age: 65 years

Compliance Report

Usage	04/22/2017 - 05/21/2017
Usage days	30/30 days (100%)
>= 4 hours	30 days (100%)
< 4 hours	0 days (0%)
Usage hours	260 hours 2 minutes
Average usage (total days)	8 hours 40 minutes
Average usage (days used)	8 hours 40 minutes
Median usage (days used)	8 hours 32 minutes


AirSense 10 AutoSet For Her

Serial number	[REDACTED]
Mode	AutoSet for Her
Min Pressure	5 cmH2O
Max Pressure	14 cmH2O
EPR	Fulltime
EPR level	3

Therapy

Pressure - cmH2O	Median: 10.6	95th percentile: 12.3	Maximum: 13.1
Leaks - L/min	Median: 0.0	95th percentile: 3.5	Maximum: 83.1
Events per hour	AI: 5.0	HI: 1.2	AHI: 6.2
Apnea Index	Central: 2.7	Obstructive: 2.1	Unknown: 0.1
RERA Index	0.1		
Cheyne-Stokes respiration (average duration per night)	0 minutes (0%)		

Usage - hours





AirView™

TEMP
444 W 2nd St
Suite 103
Ft Worth
Arlington, 76010
Phone: 800-835-5100
Fax: 800-835-5104
Email: cg-200@medcoast.com

04/22/2017 - 05/21/2017

DOE
Age: 65 years

Therapy Report


AirSense 10 AutoSet For Her

SN




Usage (hours)

Usage days 30/30 (100%)
≥ 4 hours days 30 (100%)
≤ 4 hours days 0 (0%)
Days not used 0 (0%)
Days no data 0 (0%)
Used/day (avg) 8.7 hrs




Leak (L/min)

Set threshold 24.0 L/min
Maximum (avg) 82.1
95th % (avg) 3.6
Median (avg) 0.0




Pressure (cmH2O)

Mode AutoSet for Her
Set EPR Fulltime
Set Max Pressure 14.0
Set Min Pressure 5.0
Maximum (avg) 13.1
95th % (avg) 12.3
Median (avg) 10.8



AHI (events/hour)

AHI 6.2
HI 1.2
AI 5.0
CAI 2.7
OAI 2.1
UAI 0.1
RERA 0.1
CSRR% (avg) 0.0





ResMed
 444 W. 2nd St
 Suite 103
 Tempe, AZ 85281
 Phone: 480-835-9100
 Fax: 480-835-9104
 Email: cg-2006@resmed.com

05/28/2017 - 05/28/2017

DOB: [REDACTED] 951

Age: 65 years

Compliance Report

Usage		05/28/2017 - 05/28/2017
Usage days		1/1 days (100%)
>= 4 hours		1 days (100%)
< 4 hours		0 days (0%)
Usage hours		9 hours 40 minutes
Average usage (total days)		9 hours 40 minutes
Average usage (days used)		9 hours 40 minutes
Median usage (days used)		9 hours 40 minutes


AirSense 10 AutoSet For Her

Serial number	[REDACTED]
Mode	AutoSet for Her
Min Pressure	5 cmH2O
Max Pressure	14 cmH2O
EPR	Fulltime
EPR level	3

Therapy

Pressure - cmH2O	Median: 10.0	95th percentile: 11.9	Maximum: 12.8
Leaks - L/min	Median: 0.0	95th percentile: 18.0	Maximum: 114.0
Events per hour	AI: 4.0	HI: 1.6	AHI: 5.6
Apnea Index	Central: 2.0	Obstructive: 1.7	Unknown: 0.2
RERA Index			0.1
Cheyne-Stokes respiration (average duration per night)			0 minutes (0%)

Usage - hours





ResMed

AirView™

TYPE
d44 W 2 airBr
Size 103
Type
Airbox, 8528
Phone: 480-835-9100
Fax: 480-835-9104
Email: cgaroto@resmed.com

05/28/2017 - 05/28/2017

951
Age: 65 years


Therapy Report

AirSense 10 AutoSet For Her

SN. [REDACTED]


Usage (hours)

Usage days 1/1 (100%)
≥ 4 hours days 1 (100%)
≤ 4 hours days 0 (0%)
Days not used 0 (0%)
Days no data 0 (0%)
Used/day (avg) 9.7 hrs




Leak (L/min)

Set threshold 24.0 L/min
Maximum (avg) 114.0
95th % (avg) 18.0
Median (avg) 0.0




Pressure (cmH2O)

Mode AutoSet for Her
Set EPR Fulltime, 3.0
Set Max Pressure 14.0
Set Min Pressure 5.0
Maximum (avg) 12.8
95th % (avg) 11.9
Median (avg) 10.0



AHI (events/hour)

AHI 5.6
HI 1.6
AI 4.0
CAI 2.0
OAI 1.7
UAI 0.2
RERA 0.1
CSRR% (avg) 0.0



09/02/2013

Setup date



Compliant

100%

Last 30

CPAP

S9 Elite

[Back to patients](#)

- Charts
- Patient details
- Prescription
- Notes
- Logs
- Thresholds

Fixed time period




90

day period ending

07/31/2017

Update

■ Compliant use ■ Non-compliant use



[Back to patients](#)

Charts

Patient details

Prescription

Notes

Logs


Thresholds


Fixed time period


90 day period ending 09/06/2017


Update


■ Compliant use ■ Non-compliant use














Compliance Report

Compliance

Payor

Medicare (Part A & B)

Usage

01/23/2018 - 02/21/2018

Usage days	29/30 days (97%)
>= 4 hours	18 days (60%)
< 4 hours	11 days (37%)
Usage hours	115 hours 10 minutes
Average usage (total days)	3 hours 50 minutes
Average usage (days used)	3 hours 58 minutes
Median usage (days used)	4 hours 17 minutes
Total used hours (value since last reset)	276 hours


AirSense 10 AutoSet For Her


Serial number	2
Mode	AutoSet for Her
Min Pressure	6 cmH2O
Max Pressure	14 cmH2O
EPR	Fulltime
EPR level	3

Therapy

Pressure - cmH2O	Median: 9.1	95th percentile: 10.0	Maximum: 10.2
Leak - L/min	Median: 13.2	95th percentile: 40.6	Maximum: 54.4
Events per hour	AI: 27.9	HI: 0.2	AHI: 28.1
Apnea Index	Central: 0.1	Obstructive: 23.1	Unknown: 4.6
RERA Index	0.5		
Cheyne-Stokes respiration (average duration per night)	0 minutes (0%)		

Usage - hours





ResMed

AirView™

Valley Respiratory Services-Mesa
2350 W Broadway Road
Suite 112
Mesa
Arizona, 85202
Phone: 480-630-7700
Email: revans@valleyo2.com

01/23/2018 - 02/21/2018

Therapy Report


AirSense 10 AutoSet For Her

SN: [REDACTED]

Usage (hours)


Usage days 29/30 (97%)
>= 4 hour days 18 (60%)
< 4 hour days 11 (37%)
Days not used 1 (3%)
Days no data 0 (0%)

Used/day (avg.) 4.0 hrs.




Leak (L/min)

Set threshold 24.0 L/min
Maximum (avg) 64.4
95th % (avg) 40.6
Median (avg) 13.2




Pressure (cmH2O)

Mode AutoSet for Her
Set EPR Fulltime, 3.0
Set Max Pressure 14.0
Set Min Pressure 8.0
Maximum (avg) 10.2
95th % (avg) 10.0
Median (avg) 9.1



AHI (events/hour)

AHI 28.1
HI 0.2
AI 27.9
CAI 0.1
OAI 23.1
UAI 4.6
RERA 0.6
CSR% (avg) 0.0



patients

Patient details

Prescription

Remote Assist

Notes

Logs

Preferences

Fixed time period



30

day period ending


06/04/2018

Update

Compliant use

Non-compliant use

Usage




Fixed time period ☐

90 day period ending 11/21/2017

Update

Compliant use

Non-compliant use





[Back to patients](#)

Patient ID

06/05/1948
D.O.B.

07/22/2013
Setup date

✓
Compliant

97%
Last 30

Spont
S9 VPAP Auto

Charts

Patient details

Prescription

Notes

Logs

Preferences

Fixed time period



90


day period ending

03/14/2018

Update

■ Compliant use ■ Non-compliant use

Usage



100%

THE END SO FAR...