



813-A Woodcrest Drive · Houston, Texas · 77018-2127
 (713)695-1939 · Fax: (713)695-3001
 http://www.thermotemp.com

Certificate Of Calibration

Calibrated for :P & B Testing
 6645 West Tidwell
 Houston Texas,77092

Report No:	62865	Calibration Interval:	6 Months
PO. No.:	85532	Calibration Date:	03/10/2021
Company No:	103008	Calibration Due Date:	09/10/2021
Description:	Gould-Bass 3509B	Serial No:	103008
Model No:	3509B	Manufacturer:	Gould-Bass
Temperature:	64°F	Condition As Found :	In Tolerance
Humidity :	35%RH	Condition As Left :	In Tolerance-No Adjustment
Cal. Procedure:	FCP-9		

The standards used in this certification have measurement traceability to the International System of Units (SI). The Management System of Thermo-Temp, Inc. has been assessed and accredited to ISO 17025, ANSI/NCSL Z540-1-1994 and ANSI/NCSL Z540.3-2006. The procedures used for this calibration is as listed above. The results stated in this report relate only to the items tested or calibrated. This calibration was performed to customer requirements. The company's responsibility shall in no event, and for no cause, exceed the cost of calibration of the above documented equipment. Where uncertainties are provided, the uncertainty stated is the expanded uncertainty of the measurement, where k=2. This report shall not be reproduced, except in full, without the written permission of Thermo-Temp, Inc.

Asset Requirements: 6K AMPS DC Shot Time = .75 second Quick Break - OK

Standards Utilized

I.D. No.	Mfg	Model No.	Description	Rpt. #	Cal. Date	Due Date
MM144	Fluke	87 V	Fluke 87 V True RMS Voltmeter	94685	10/23/2020	04/23/2021
WD153	Empro Manufacturing Co.	E2000-100	Shunt, 200 Amp	50795	05/04/2017	05/04/2021

Test Point Data

S/N	Function Tested	Nominal	As Found Condition	As Left Condition	Error as Left	Test Limits Low	Test Limits High	Test Result
103008 DC Ammeter	DC Amps	5000	5026	5026	26	4400	5600	Pass
	DC Amps	500	510	510	10	-100	1100	Pass
	DC Amps	1000	1016	1016	16	400	1600	Pass
	DC Amps	2000	2022	2022	22	1400	2600	Pass
	DC Amps	3000	3016	3016	16	2400	3600	Pass
	DC Amps	4000	4028	4028	28	3400	4600	Pass

Calibration Performed By:			Certificate Authorized By:		
Jeromy Holman	Technician		Kevin Green	3/12/2021	
Name	Title	Signature	Field Services Manager	Date	Signature