



813-A Woodcrest Drive · Houston, Texas · 77018-2127  
(713)695-1939 · Fax: (713)695-3001  
<http://www.thermotemp.com>

## Certificate Of Calibration

Calibrated for :P & B Testing  
6645 West Tidwell  
Houston Texas,77092

Report No:	125483	Calibration Interval:	1 Year
PO. No.:	86725	Calibration Date:	08/08/2024
Company No:	11039	Calibration Due Date:	08/08/2025
Description:	Rockwell Hardness Tester	Serial No:	11039
Model No:	CPTXL	Manufacturer:	Clark
Temperature:	84°F	Humidity :	35.00 %RH
Cal. Procedure:	CP-02		



THIS DOCUMENT IS TO CERTIFY THAT THE AFOREMENTIONED HARDNESS TESTER PASSED VERIFICATION BY INDIRECT METHOD IN THE RANGES LISTED BELOW USING A TEST BLOCK CALIBRATED IN THE STANDARDIZATION LABORATORY OF THE BLOCK MANUFACTURER (SUNTEC). ALL TESTS WERE PERFORMED IN ACCORDANCE WITH OUR QUALITY MANUAL AND THE LATEST REVISION OF ASTM E18 AND / OR E103. THERMO-TEMP INC. QUALITY ASSURANCE PROGRAM FOLLOWS THE REQUIREMENTS STATED IN ANSI / NCSL Z540-1.

This report shall not be reproduced, except in full, without the written approval of Thermo-Temp Incorporated

### Asset Requirements:

### VERIFICATION TEST DATA

TEST BLOCK		SERIAL NO	HARDNESS TESTER READINGS					(AS FOUND/AS LEFT)		
			RD1	RD2	RD3	RD4	RD5	Average	Error	Repeatability
HR C	24.80+/-1.00	22250101	24.80	24.80	25.00	25.00	24.70	24.86	0.06	0.30
HR C	45.10+/-1.80	23450257	45.00	45.00	44.80	44.90	45.00	44.94	-0.16	0.20
HR C	60.40+/-0.50	22610189	60.50	60.50	60.60	60.60	60.40	60.52	0.12	0.20
HR Bw	44.40+/-2.50	22400088	44.00	43.80	43.90	43.60	44.20	43.90	-0.50	0.60
HR Bw	72.90+/-2.50	22700266	73.40	73.30	73.50	73.30	73.10	73.32	0.42	0.40
HR Bw	96.20+/-1.00	22950123	96.50	96.30	96.80	96.60	96.50	96.54	0.34	0.50

Calibration Performed By:			Certificate Authorized By:		
Jeromy Holman	Technician		Ashley Holman	8/15/2024	
Name	Title	Signature	Quality Control	Date	Signature

*Buck Snider*