



813-A Woodcrest Drive · Houston, Texas · 77018-2127
 (713)695-1939 · Fax: (713)695-3001
<http://www.thermotemp.com>

Certificate Of Calibration

Calibrated for :P & B Testing
 6645 West Tidwell
 Houston Texas,77092

Report No:	52087	Calibration Interval:	1 Year
PO. No.:	85512	Calibration Date:	02/10/2021
Company No:	7088	Calibration Due Date:	02/10/2022
Description:	ROCKWELL	Serial No:	7088
Model No:	AR-12	Manufacturer:	Clark
Temperature:	64°F	Humidity :	60.00 %RH
Cal. Procedure:	FCP-6, Rev 0		

THIS DOCUMENT IS TO CERTIFY THAT THE AFOREMENTIONED HARDNESS TESTER PASSED VERIFICATION BY INDIRECT METHOD IN THE RANGES LISTED BELOW USING A TEST BLOCK CALIBRATED IN THE STANDARDIZATION LABORATORY OF THE BLOCK MANUFACTURER (SUNTEC). ALL TESTS WERE PERFORMED IN ACCORDANCE WITH OUR QUALITY MANUAL AND THE LATEST REVISION OF ASTM E18 AND / OR E103. THERMO-TEMP INC. QUALITY ASSURANCE PROGRAM FOLLOWS THE REQUIREMENTS STATED IN ANSI / NCSL Z540-1.

This report shall not be reproduced, except in full, without the written approval of Thermo-Temp Incorporated

Asset Requirements: 0

VERIFICATION TEST DATA

TEST BLOCK	SERIAL NO	HARDNESS TESTER READINGS					(AS FOUND/AS LEFT)		
		RD1	RD2	RD3	RD4	RD5	Average	Error	Repeatability
HR C 29.40+/-1.00	182500 22	29.60	29.30	29.50	29.40	29.30	29.42	0.02	0.30
HR C 48.40+/-1.00	164779	48.80	48.20	48.60	48.50	48.50	48.52	0.12	0.60
HR C 62.00+/-0.50	166059	61.70	61.70	62.00	62.00	61.90	61.86	-0.14	0.30
HR Bw 42.00+/-2.50	194700 43	42.30	42.40	42.80	42.90	42.60	42.60	0.60	0.60
HR Bw 73.10+/-2.50	196401 56	73.00	73.00	72.60	73.30	73.20	73.02	-0.08	0.70
HR Bw 92.60+/-1.00	199003 10	92.60	93.20	92.20	93.00	92.80	92.76	0.16	1.00

Calibration Performed By:			Certificate Authorized By:		
Jeromy Holman	Technician		Kevin Green	2/12/2021	
Name	Title	Signature	Field Services Manager	Date	Signature