



813-A Woodcrest Drive · Houston, Texas · 77018-2127  
(713)695-1939 · Fax: (713)695-3001  
<http://www.thermotemp.com>

## Certificate Of Calibration

Calibrated for :P & B Testing  
6645 West Tidwell  
Houston Texas,77092

Report No:	130391	Calibration Interval:	1 Year
PO. No.:	86899	Calibration Date:	02/06/2025
Company No:	7088	Calibration Due Date:	02/06/2026
Description:	ROCKWELL	Serial No:	7088
Model No:	AR-12	Manufacturer:	Clark
Temperature:	74°F	Humidity :	50.00 %RH
Cal. Procedure:	CP-02		



THIS DOCUMENT IS TO CERTIFY THAT THE AFOREMENTIONED HARDNESS TESTER PASSED VERIFICATION BY INDIRECT METHOD IN THE RANGES LISTED BELOW USING A TEST BLOCK CALIBRATED IN THE STANDARDIZATION LABORATORY OF THE BLOCK MANUFACTURER (SUNTEC). ALL TESTS WERE PERFORMED IN ACCORDANCE WITH OUR QUALITY MANUAL AND THE LATEST REVISION OF ASTM E18 AND / OR E103. THERMO-TEMP INC. QUALITY ASSURANCE PROGRAM FOLLOWS THE REQUIREMENTS STATED IN ANSI / NCSL Z540-1.

This report shall not be reproduced, except in full, without the written approval of Thermo-Temp Incorporated

Asset Requirements: 0

### VERIFICATION TEST DATA

TEST BLOCK		SERIAL NO	HARDNESS TESTER READINGS					(AS FOUND/AS LEFT)		
			RD1	RD2	RD3	RD4	RD5	Average	Error	Repeatability
HR C	28.90+/-1.00	24270214	28.40	28.50	28.50	28.30	28.40	28.42	-0.48	0.20
HR C	46.40+/-1.00	24440262	46.20	46.10	46.20	46.30	46.00	46.16	-0.24	0.30
HR C	62.30+/-0.50	24610256	62.00	62.00	62.10	62.00	62.10	62.04	-0.26	0.10
HR Bw	46.30+/-2.50	24440170	46.50	46.40	46.50	46.60	46.50	46.50	0.20	0.20
HR Bw	81.30+/-2.50	24720105	81.90	81.90	82.00	82.10	81.90	81.96	0.66	0.20
HR Bw	100.30+/-1.00	23970156	100.50	100.50	100.40	100.50	100.50	100.48	0.18	0.10

Calibration Performed By:			Certificate Authorized By:		
Jeromy Holman	Technician		Ashley Holman	2/7/2025	
Name	Title	Signature	Quality Control	Date	Signature

*Buck Snider*