



813-A Woodcrest Drive · Houston, Texas · 77018-2127
(713)695-1939 · Fax: (713)695-3001
<http://www.thermotemp.com>

Certificate Of Calibration

Calibrated for :P & B Testing
6645 West Tidwell
Houston Texas,77092

Report No:	137545	Calibration Interval:	1 Year
PO. No.:	87071	Calibration Date:	10/28/2025
Company No:	CPT18979	Calibration Due Date:	10/28/2026
Description:	Rockwell Hardness Tester	Serial No:	CPT18979
Model No:	CPT	Manufacturer:	Clark
Temperature:	73°F	Humidity :	29.00 %RH
Cal. Procedure:	CP-02		

THIS DOCUMENT IS TO CERTIFY THAT THE AFOREMENTIONED HARDNESS TESTER PASSED VERIFICATION BY INDIRECT METHOD IN THE RANGES LISTED BELOW USING A TEST BLOCK CALIBRATED IN THE STANDARDIZATION LABORATORY OF THE BLOCK MANUFACTURER (SUNTEC). ALL TESTS WERE PERFORMED IN ACCORDANCE WITH OUR QUALITY MANUAL AND THE LATEST REVISION OF ASTM E18 AND / OR E103. THERMO-TEMP INC. QUALITY ASSURANCE PROGRAM FOLLOWS THE REQUIREMENTS STATED IN ANSI / NCSL Z540-1.


This report shall not be reproduced, except in full, without the written approval of Thermo-Temp Incorporated

Asset Requirements:

VERIFICATION TEST DATA

TEST BLOCK		SERIAL NO	HARDNESS TESTER READINGS					(AS FOUND/AS LEFT)		
			RD1	RD2	RD3	RD4	RD5	Average	Error	Repeatability
HR C	24.70+/-1.00	22250114	24.30	24.20	24.10	24.20	24.30	24.22	-0.48	0.20
HR C	45.10+/-1.00	25440111	44.30	44.20	44.40	44.20	44.10	44.24	-0.86	0.30
HR C	63.00+/-0.50	24630178	62.80	63.00	62.70	62.80	62.90	62.84	-0.16	0.30
HR Bw	44.20+/-2.50	24440154	45.00	45.10	44.90	45.30	44.90	45.04	0.84	0.40
HR Bw	75.70+/-2.50	25740149	76.00	75.90	76.20	76.00	76.30	76.08	0.38	0.40
HR Bw	96.40+/-1.00	25240140	96.00	96.20	96.00	96.10	96.20	96.10	-0.30	0.20

Buck Snider

Calibration Performed By:			Certificate Authorized By:		
Jeromy Holman	Technician		Ashley Holman	10/30/2025	
Name	Title	Signature	Quality Control	Date	Signature