



813-A Woodcrest Drive · Houston, Texas · 77018-2127
(713)695-1939 · Fax: (713)695-3001
<http://www.thermotemp.com>

Certificate Of Calibration

Calibrated for :P & B Testing
6645 West Tidwell
Houston Texas,77092

Report No:	135315	Calibration Interval:	1 Year
PO. No.:	87012	Calibration Date:	08/06/2025
Company No:	11039	Calibration Due Date:	08/06/2026
Description:	Rockwell Hardness Tester	Serial No:	11039
Model No:	CPTXL	Manufacturer:	Clark
Temperature:	90°F	Humidity :	46.00 %RH
Cal. Procedure:	CP-02		



THIS DOCUMENT IS TO CERTIFY THAT THE AFOREMENTIONED HARDNESS TESTER PASSED VERIFICATION BY INDIRECT METHOD IN THE RANGES LISTED BELOW USING A TEST BLOCK CALIBRATED IN THE STANDARDIZATION LABORATORY OF THE BLOCK MANUFACTURER (SUNTEC). ALL TESTS WERE PERFORMED IN ACCORDANCE WITH OUR QUALITY MANUAL AND THE LATEST REVISION OF ASTM E18 AND / OR E103. THERMO-TEMP INC. QUALITY ASSURANCE PROGRAM FOLLOWS THE REQUIREMENTS STATED IN ANSI / NCSL Z540-1.

This report shall not be reproduced, except in full, without the written approval of Thermo-Temp Incorporated

Asset Requirements:

VERIFICATION TEST DATA

TEST BLOCK		SERIAL NO	HARDNESS TESTER READINGS					(AS FOUND/AS LEFT)		
			RD1	RD2	RD3	RD4	RD5	Average	Error	Repeatability
HR C	28.90+/-1.00	24270235	28.30	28.20	28.20	28.10	28.30	28.22	-0.68	0.20
HR C	44.90+/-1.00	23450269	44.40	44.50	44.30	44.50	44.20	44.38	-0.52	0.30
HR C	62.30+/-0.50	24610318	62.00	62.00	62.00	61.90	61.80	61.94	-0.36	0.20
HR Bw	44.00+/-2.50	24440215	44.00	44.30	44.50	44.00	44.50	44.26	0.26	0.50
HR Bw	68.00+/-2.50	24650181	67.50	67.50	68.00	67.70	67.50	67.64	-0.36	0.50
HR Bw	92.30+/-1.00	24890201	91.70	91.80	91.70	91.60	91.80	91.72	-0.58	0.20

Calibration Performed By:			Certificate Authorized By:		
Jeromy Holman	Technician		Ashley Holman	8/7/2025	
Name	Title	Signature	Quality Control	Date	Signature

Buck Snider