



813-A Woodcrest Drive · Houston, Texas · 77018-2127
(713)695-1939 · Fax: (713)695-3001
<http://www.thermotemp.com>

Certificate Of Calibration

Calibrated for :P & B Testing
6645 West Tidwell
Houston Texas,77092

Report No:	145364	Calibration Interval:	1 Year
PO. No.:	87215	Calibration Date:	08/06/2026
Company No:	11039	Calibration Due Date:	08/06/2027
Description:	Rockwell Hardness Tester	Serial No:	11039
Model No:	CPTXL	Manufacturer:	Clark
Temperature:	87°F	Humidity :	64.00 %RH
Cal. Procedure:	CP-02		

THIS DOCUMENT IS TO CERTIFY THAT THE AFOREMENTIONED HARDNESS TESTER PASSED VERIFICATION BY INDIRECT METHOD IN THE RANGES LISTED BELOW USING A TEST BLOCK CALIBRATED IN THE STANDARDIZATION LABORATORY OF THE BLOCK MANUFACTURER (SUNTEC). ALL TESTS WERE PERFORMED IN ACCORDANCE WITH OUR QUALITY MANUAL AND THE LATEST REVISION OF ASTM E18 AND / OR E103. THERMO-TEMP INC. QUALITY ASSURANCE PROGRAM FOLLOWS THE REQUIREMENTS STATED IN ANSI / NCSL Z540-1.


This report shall not be reproduced, except in full, without the written approval of Thermo-Temp Incorporated

Asset Requirements:

VERIFICATION TEST DATA

TEST BLOCK		SERIAL NO	HARDNESS TESTER READINGS					(AS FOUND/AS LEFT)		
			RD1	RD2	RD3	RD4	RD5	Average	Error	Repeatability
HR C	24.30+/-1.00	25240959	23.90	24.00	23.60	23.50	23.80	23.76	-0.54	0.50
HR C	46.40+/-1.00	25440546	45.90	46.00	46.00	45.80	45.90	45.92	-0.48	0.20
HR C	63.20+/-0.50	25620273	63.00	62.90	63.00	63.00	62.70	62.92	-0.28	0.30
HR Bw	45.70+/-2.50	24450020	45.40	45.00	45.00	45.00	45.20	45.12	-0.58	0.40
HR Bw	70.30+/-2.50	25620083	69.20	69.00	69.10	68.80	68.90	69.00	-1.30	0.40
HR Bw	92.60+/-1.00	25920198	92.00	92.00	91.80	91.90	92.10	91.96	-0.64	0.30

Buck Snider

Calibration Performed By:			Certificate Authorized By:		
Jeromy Holman	Technician		Ashley Holman	8/6/2026	
Name	Title	Signature	Quality Control	Date	Signature