



Hocker Incorporated

13402 Weiman Road Houston, TX 77041

713-464-5829 Fax 713-464-3192

www.hockerinc.com

Performance Evaluation

UV & White Light Meter

Customer PO #:	86904
Certification #:	25-0056
Evaluation Date:	Due Date:
1/17/2025	7/17/2025

F-LTMT-0		Operational When Received?		No	DLM-1000		Equipment Condition When Received?	
Company: P & B Testing, Inc.		In Tolerance when Received?		Yes		<div></div> <div>New</div> <div>Good</div> <div><div>X</div>Fair</div> <div>Poor</div> <div>Failed</div>		
Address: 6645 W. Tidwell								
City: Houston		Equipment Description:		Dual Light Meter Model;DLM-1000 S/N: 204182 A,B,C				
State: TX								
Zip: 77092								
Country: USA		Phone:	(713)290-8490					
Contact: Buck Snider		Email:	qa@pbtesting.com					
		File Loc.	P & B Testing, Inc. 45674_25-0056					

Equipment Used For This Eval.:	Serial #:	Cal. Date:	Due Date:	NIST#:
Gould Bass DLM-1000 A Readout	142154 A	11/22/2024	5/22/2025	A=116415/1533001/Q1734
Gould Bass DLM-1000 B UV Sensor	142154 B	11/22/2024	5/22/2025	B=116415/00249,10540/Q1734
Gould Bass DLM-1000 C WL Sensor	142154 C	11/22/2024	5/22/2025	C=116415/811,7481/Y25189/W10962 Q1734

Tolerance %:	5.0%	(+/-)	Meter Type:	Digital	3 PCS	Equipment Mfg.:
						Gould Bass

Equipment Owner:	Address:	City:	State:	Zip:
P & B Testing, Inc.	6645 W. Tidwell	Houston	TX	77092

DLM-1000 "B" Sensor	Temperature:	70°F	Humid.:	51%
---------------------	--------------	------	---------	-----

UV As Found	UV Meter As Received				UV As Left	UV Meter After Adjustment			
uW/sq.cm	uW/sq.cm	uW/sq.cm	% Diff.	Tolerance	uW/sq.cm	uW/sq.cm	uW/sq.cm	% Diff.	Tolerance
Ideal	Actual	Difference			Ideal	Actual	Difference		
0	0	0		In	0	0	0		In
720	710	-10	-1.4%	In	720	730	10	1.4%	In
2200	2170	-30	-1.4%	In	2200	2230	30	1.4%	In
4670	4530	-140	-3.0%	In	4670	4600	-70	-1.5%	In
Maximum Deviation		-140	-3.0%		Maximum Deviation		-70	-1.5%	

DLM-1000 "C" Sensor									
WL As Found	White Light Meter As Received				WL As Left	White Light Meter....After Adjustment			
Foot Candles	Foot Candles	Foot Candles	% Diff.	Tolerance	Foot Candles	Foot Candles	Foot Candles	% Diff.	Tolerance
Ideal	Actual	Difference			Ideal	Actual	Difference		
0.0	0.0	0.0		In	0.0	0.0	0.0		In
3.0	3.0	0.0	0.0%	In	3.0	3.0	0.0	0.0%	In
100.0	99.0	-1.0	-1.0%	In	100.0	100.0	0.0	0.0%	In
400.0	395.0	-5.0	-1.3%	In	400.0	401.0	1.0	0.3%	In
Maximum Deviation		-5	-1.3%		Maximum Deviation		1	0.3%	


Notes: AS received condition- Unit has white powder covering sensor lenses : 9v Battery was dead and broke off the terminal when removed exposing corrosion on the terminal inside meter, Tech cleaned battery connection terminal & removed white powder from the sensor(s)/ Meter , Lab Battery was used for evaluation.
 AS FOUND /AS LEFT (A,B,& C) <5% Adjustment
 not required for calibration , but made to lower %

P & B Testing, Inc.


Certification #
Evaluation Date

25-0056
1/17/2025


UV Sensor AS Found




UV Sensor AS Left



White Light Sensor AS Found



White Light Sensor AS Left



Hocker Incorporated calibration procedure CP-LTMT- 1 Was utilized for this performance evaluation /report of findings. CP-LTMT-1 is in accordance with and complies to, but not limited to: ANSI/ NCSL-Z540 ; ASTM-E1444, ASTM E-709, ASTM E-1417. Master Equipment or transfer standards & meterology equipment used to perform this evaluation are traceable to the National Institute of Standards and Technology (NIST) through SI. Referenced traceability numbers listed in this document are on file with supporting documentation. This certificate or report shall not be reproduced except in full without the written approval of HOCKER INCORPORATED

In Conformance with 10CFR21, ANSI/NCSL 2540.3-2006 & ISO10012.

Technician Signature:

[Signature]

Technician Performing Evaluation:

Derrick Schumann

Approval Signature:

[Signature]

Approved By:

Lewis Brittain

Date

1/17/2025

F-LTMT-0 05/01/2018
CP-LTMT-1 05/24/2022

An ISO 9001:2015 Registered Company

Buck Snider