



813-A Woodcrest Drive · Houston, Texas · 77018-2127
(713)695-1939 · Fax: (713)695-3001
<http://www.thermotemp.com>

Certificate Of Calibration

Calibrated for :P & B Testing
6645 West Tidwell
Houston Texas,77092

Report No:	135314	Calibration Interval:	6 Months
PO. No.:	87012	Calibration Date:	08/06/2025
Company No:	103008	Calibration Due Date:	02/06/2026
Description:	Gould-Bass 3509B	Serial No:	103008
Model No:	3509B	Manufacturer:	Gould-Bass
Temperature:	94°F	Condition As Found :	In Tolerance
Humidity :	46%RH	Condition As Left :	In Tolerance-No Adjustment
Cal. Procedure:	FCP-9		

The standards used in this certification have measurement traceability to the International System of Units (SI), through National Metrology Institutes (NIST, PTB, NIM). The Management System of Thermo-Temp, Inc. has been assessed and accredited to ISO 17025, ANSI/NCSL Z540-1-1994 and ANSI/NCSL Z540.3-2006. The procedures used for this calibration is as listed above. The results stated in this report relate only to the items tested or calibrated. This calibration was performed to customer requirements. The company's responsibility shall in no event, and for no cause, exceed the cost of calibration of the above documented equipment. Where uncertainties are provided, the uncertainty stated is the expanded uncertainty of the measurement, where k=2. This report shall not be reproduced, except in full, without the written permission of Thermo-Temp, Inc.



Asset Requirements: 6K AMPS DC Shot Time = .75 second Quick Break - OK

Standards Utilized

I.D. No.	Mfg	Model No.	Description	Rpt. #	Cal. Date	Due Date
CL114	Fluke	787	Multimeter, Process	134172	10/22/2024	10/22/2025
WD164	Empro Manufacturing Co.	E2000-100x	2000 amp Shunt	124997	10/02/2023	10/02/2025
TL290	Extech	44550	Temp/Humidity Pen	137439	05/02/2025	05/02/2026

Test Point Data

S/N	Function Tested	Nominal	As Found Condition	As Left Condition	Error as Left	Test Limits Low	Test Limits High	Test Result
103008 DC Ammeter	DC Amps	5000	5042	5042	42	4400	5600	Pass
	DC Amps	500	516	516	16	-100	1100	Pass
	DC Amps	1000	1028	1028	28	400	1600	Pass
	DC Amps	2000	2030	2030	30	1400	2600	Pass
	DC Amps	3000	3048	3048	48	2400	3600	Pass
	DC Amps	4000	4042	4042	42	3400	4600	Pass

Calibration Performed By:			Certificate Authorized By:		
Jeromy Holman	Technician		Ashley Holman	8/7/2025	
Name	Title	Signature	Quality Control	Date	Signature

Buck Snider