Thermo-Temp INCORPORATED
813-A Woodcrest Drive · Houston, Texas · 77018-2127
(713)695-1939 · Fax: (713)695-3001
http://www.thermotemp.com

Certificate Of Calibration

Calibrated for :P & B Testing
6645 West Tidwell
Houston Texas,77092

Report No:	147228	Calibration Interval:	1 Year
PO. No.:	87283	Calibration Date:	10/12/2026
Company No:	CPT18979	Calibration Due Date:	10/12/2027
Description:	Rockwell Hardness Tester	Serial No:	CPT18979
Model No:	CPT	Manufacturer:	Clark
Temperature:	87°F	Humidity :	64.00 %RH
Cal. Procedure:	CP-02		

THIS DOCUMENT IS TO CERTIFY THAT THE AFOREMENTIONED HARDNESS TESTER PASSED VERIFICATION BY INDIRECT METHOD IN THE RANGES LISTED BELOW USING A TEST BLOCK CALIBRATED IN THE STANDARDIZATION LABORATORY OF THE BLOCK MANUFACTURER (SUNTEC). ALL TESTS WERE PERFORMED IN ACCORDANCE WITH OUR QUALITY MANUAL AND THE LATEST REVISION OF ASTM E18 AND / OR E103. THERMO-TEMP INC. QUALITY ASSURANCE PROGRAM FOLLOWS THE REQUIREMENTS STATED IN ANSI / NCSL Z540-1.

This report shall not be reproduced, except in full, without the written approval of Thermo-Temp Incorporated

---

Asset Requirements:

VERIFICATION TEST DATA

TEST BLOCK		SERIAL NO	HARDNESS TESTER READINGS					(AS FOUND/AS LEFT)		
			RD1	RD2	RD3	RD4	RD5	Average	Error	Repeatability
HR C	29.20+/-1.00	25270326	29.00	28.50	28.70	28.50	28.60	28.66	-0.54	0.50
HR C	46.60+/-1.00	25440475	45.80	45.90	45.80	45.70	45.80	45.80	-0.80	0.20
HR C	64.30+/-0.50	23600833	64.00	64.00	64.00	63.90	64.00	63.98	-0.32	0.10
HR Bw	46.10+/-2.50	25430099	46.00	45.70	45.70	45.90	45.50	45.76	-0.34	0.50
HR Bw	74.10+/-2.50	26700071	74.00	74.50	74.70	74.60	74.50	74.46	0.36	0.70
HR Bw	92.90+/-1.00	26900140	92.50	92.20	92.30	92.20	92.30	92.30	-0.60	0.30

Buck Snider

Calibration Performed By:			Certificate Authorized By:		
Jeromy Holman	Technician		Ashley Holman	10/15/2026	
Name	Title	Signature	Quality Control	Date	Signature