



813-A Woodcrest Drive · Houston, Texas · 77018-2127
(713)695-1939 · Fax: (713)695-3001
<http://www.thermotemp.com>

Certificate Of Calibration

Calibrated for :P & B Testing
6645 West Tidwell
Houston Texas,77092

Report No:	145363	Calibration Interval:	6 Months
PO. No.:	87215	Calibration Date:	08/06/2026
Company No:	103008	Calibration Due Date:	02/06/2027
Description:	Gould-Bass 3509B	Serial No:	103008
Model No:	3509B	Manufacturer:	Gould-Bass
Temperature:	87°F	Condition As Found :	In Tolerance
Humidity :	64%RH	Condition As Left :	In Tolerance-No Adjustment
Cal. Procedure:	FCP-9		

The standards used in this certification have measurement traceability to the International System of Units (SI), through National Metrology Institutes (NIST, PTB, NIM). The Management System of Thermo-Temp, Inc. has been assessed and accredited to ISO 17025, ANSI/NCSL Z540-1-1994 and ANSI/NCSL Z540.3-2006. The procedures used for this calibration is as listed above. The results stated in this report relate only to the items tested or calibrated. This calibration was performed to customer requirements. The company's responsibility shall in no event, and for no cause, exceed the cost of calibration of the above documented equipment. Where uncertainties are provided, the uncertainty stated is the expanded uncertainty of the measurement, where k=2. This report shall not be reproduced, except in full, without the written permission of Thermo-Temp, Inc.

Asset Requirements: 6K AMPS DC Shot Time = .80 seconds Quick Break - OK

Standards Utilized

I.D. No.	Mfg	Model No.	Description	Rpt. #	Cal. Date	Due Date
WD163	Empro Manufacturing Co.	E2000-100 X	Shunt, 2000 Amp	125962	06/01/2024	08/12/2026
MM145	Fluke	87 Series V	Handheld Digital True RMS Multimeter	149139	07/02/2026	07/02/2027
TL308	Extech	44550	Temp/Humidity Pen	147927	05/12/2026	05/12/2027

Test Point Data

S/N	Function Tested	Nominal	As Found Condition	As Left Condition	Error as Left	Test Limits Low	Test Limits High	Test Result
103008 DC Ammeter	DC Amps	5000	5072	5072	72	4400	5600	Pass
	DC Amps	500	520	520	20	-100	1100	Pass
	DC Amps	1000	1016	1016	16	400	1600	Pass
	DC Amps	2000	2022	2022	22	1400	2600	Pass
	DC Amps	3000	3028	3028	28	2400	3600	Pass
	DC Amps	4000	4040	4040	40	3400	4600	Pass

Calibration Performed By:			Certificate Authorized By:		
Jeromy Holman	Technician		Ashley Holman	8/6/2026	
Name	Title	Signature	Quality Control	Date	Signature

Buck Snider