



813-A Woodcrest Drive · Houston, Texas · 77018-2127
(713)695-1939 · Fax: (713)695-3001
<http://www.thermotemp.com>

Certificate Of Calibration

Calibrated for :P & B Testing

6645 West Tidwell

Houston Texas,77092

Report No:	140223	Calibration Interval:	1 Year
PO. No.:	87125	Calibration Date:	02/05/2026
Company No:	7088	Calibration Due Date:	02/05/2027
Description:	ROCKWELL	Serial No:	7088
Model No:	AR-12	Manufacturer:	Clark
Temperature:	59°F	Humidity :	37.00 %RH
Cal. Procedure:	CP-02		

THIS DOCUMENT IS TO CERTIFY THAT THE AFOREMENTIONED HARDNESS TESTER PASSED VERIFICATION BY INDIRECT METHOD IN THE RANGES LISTED BELOW USING A TEST BLOCK CALIBRATED IN THE STANDARDIZATION LABORATORY OF THE BLOCK MANUFACTURER (SUNTEC). ALL TESTS WERE PERFORMED IN ACCORDANCE WITH OUR QUALITY MANUAL AND THE LATEST REVISION OF ASTM E18 AND / OR E103. THERMO-TEMP INC. QUALITY ASSURANCE PROGRAM FOLLOWS THE REQUIREMENTS STATED IN ANSI / NCSL Z540-1.

This report shall not be reproduced, except in full, without the written approval of Thermo-Temp Incorporated

Asset Requirements: 0

VERIFICATION TEST DATA

TEST BLOCK	SERIAL NO	HARDNESS TESTER READINGS					(AS FOUND/AS LEFT)			
		RD1	RD2	RD3	RD4	RD5	Average	Error	Repeatability	
HR C	24.70+/-1.00	222501 14	24.30	24.20	24.30	24.40	24.00	24.24	-0.46	0.40
HR C	45.10+/-1.00	254401 11	45.00	45.00	45.10	44.90	45.00	45.00	-0.10	0.20
HR C	63.00+/-0.50	246301 78	62.70	62.80	62.60	62.80	63.10	62.80	-0.20	0.50
HR Bw	44.20+/-2.50	244401 54	43.40	43.70	43.80	44.20	43.40	43.70	-0.50	0.80
HR Bw	75.70+/-2.50	257401 49	76.20	75.70	76.30	76.50	76.40	76.22	0.52	0.80
HR Bw	96.40+/-1.00	252401 40	95.80	96.30	96.10	96.00	95.80	96.00	-0.40	0.50

Calibration Performed By:			Certificate Authorized By:		
Jeromy Holman	Technician		Ashley Holman	2/6/2026	
Name	Title	Signature	Quality Control	Date	Signature

Buck Snider