

DIGITAL READOUT



CERTIFICATION



We certify that the instrument described has had its performance evaluated according to the procedures set forth by the manufacturer or procedures judged to be a suitable substitute. Where possible, data was obtained relating to the "as found" condition of the instrument. "Calibrated" readings were obtained after a recommended alignment procedure was performed. The equipment used in this certification have been calibrated with traceability to NIST. These services have been provided with reference to the requirements of ASTM E 1444 & ANSI/NCSL Z540 (Latest Editions). The customer should consider the uncertainty of the measurement in determining compliance. Any reported uncertainty is based on the standard uncertainty multiplied by a cover factor of K=2, providing a level of confidence of approximately ninety five percent.

Certification Date: September 3, 2026

Instrument: GOULD BASS DLM-1000

Serial Number: 204182A

Report Number: R69108

Prev. Report Number: N/A

Received: IN TOLERANCE

Returned: IN TOLERANCE

Order Number: *007187

Expires: March 3, 2027

Certification Officer: Rodolfo Garza
(Rodolfo Garza)

CERTIFICATION RESULTS

MAX. ACCEPTABLE DEVIATION:	<u>5.0000%</u>		
	ZERO	LOW	HIGH
Standard Readings:	<u>0</u>	<u>500</u>	<u>1500</u>
Unit Readings:	<u>0</u>	<u>520</u>	<u>1560</u>
% DEVIATION	<u>0.00%</u>	<u>4.00%</u>	<u>4.00%</u>
AVERAGE DEVIATION BEFORE ADJUSTMENTS:	<u>4.00%</u>		
	ZERO	LOW	HIGH
Standard Readings:	<u>0</u>	<u>500</u>	<u>1500</u>
Unit Readings:	<u>0</u>	<u>500</u>	<u>1500</u>
% DEVIATION	<u>0.00%</u>	<u>0.00%</u>	<u>0.00%</u>
AVERAGE DEVIATION AFTER ADJUSTMENTS:	<u>0.00%</u>		

Temperature: 72
Humidity: 40%

All measurements made will be within +/- 5% of tolerance required by manufacturer specifications and/or applicable standard(s).

Please Contact Rodolfo Garza at (713) 451-6506 with your questions.
Thank you for your business!

Buck Snider

AIP
7100 N. LOOP EAST, SUITE A-6
HOUSTON, TX 77028

CUSTOMER NUMBER: 1782

P & B TESTING INC.
P.O. BOX 924469
HOUSTON, TX 77292

NIST traceable certifications are maintained
on site in our files.

AIP	NIST #	NIST #	NIST #
	QA107		

ISO 10012:2003E