


# Infrared Thermometer

Art.No : AET-R1D1

HOT  
SALE

- ✓ Non-touch measurement;
- ✓ High precision infrared sensor;
- ✓ Practical patent design
- ✓ High definition display




Forehead mode



Measure in 1s




Fever warning



memory recall



Replaceable battery



auto power-off

- Product Size : 37mm × 37mm × 138mm (L x W x H)
- Product Weight : 47g
- Memory : 32 groups
- Operating temp : 5°C~40°C
- Measuring mode : Forehead mode
- Measurement range : 32.0°C~42.2°C
- Accuracy :  $\pm 0.2^{\circ}\text{C}$  ( 35.0°C~42.0°C ) ;  $\pm 0.3^{\circ}\text{C}$  ( 32.0°C~34.9°C / 42.1°C~42.2°C )
- Voltage source : 2 x 1.5 V Batteries: size AAA
- Display : LCD/LED digital display
- Automatic power off : 60s




➤ LCD


➤ LED



# Product mix



## Display information





Benannt durch/Designated by  
Zentralstelle der Länder  
für Gesundheitsschutz  
bei Arzneimitteln und  
Medizinprodukten  
www.zlg.de  
ZLG-BS-244.10.08



Product Service

# EC Certificate

Production Quality Assurance System  
Directive 93/42/EEC on Medical Devices (MDD), Annex V  
(Devices in Class IIa, IIb or III)

**No. G2 078453 0014 Rev. 01**

## Manufacturer:

**Alicn Medical Shenzhen, Inc**

4/F, B Building  
Shenfubao Modern Optical Factory, Kengzi Street  
Pingshan District  
518122 Shenzhen City  
PEOPLE'S REPUBLIC OF CHINA

## Facility(ies):

Alicn Medical Shenzhen, Inc  
4/F, B Building, Shenfubao Modern Optical Factory, Kengzi Street,  
Pingshan District, 518122 Shenzhen City, PEOPLE'S REPUBLIC  
OF CHINA

## Product

## Category(ies):

**Clinical Infrared Thermometer,  
Electronic Thermometer and Blood  
Pressure Monitor**

The Certification Body of TÜV SÜD Product Service GmbH declares that the aforementioned manufacturer has implemented a quality assurance system for manufacture and final inspection of the respective devices / device categories in accordance with MDD Annex V. This quality assurance system conforms to the requirements of this Directive and is subject to periodical surveillance. For marketing of class IIb and III devices an additional Annex III certificate is mandatory. See also notes overleaf.

## Report No.:

GZ1805201

## Valid from:

2019-07-15

## Valid until:

2022-05-07

## Date,

2019-07-15

Stefan Preiß  
Head of Certification/Notified Body