

TEST REPORT

(Electronic version)



No:200056318

VERIFICATION WEBSITE: www.gttc.net.cn

VERIFICATION CODE: OHHF-3043-34

ISSUE DATE:2020-04-09

扫码下载报告



APPLICANT: KANG DE SHI MEDICAL EQUIPMENT (ZHAOQING) CO.,LTD
ADDRESS: NO.11, JINFENG ROAD, JINLI TOWN, GAOYAO DISTRICT, ZHAOQING CITY, GUANGDONG PROVINCE

SAMPLE INFORMATION:

THIRTY DISPOSABLE MASKS

BRAND: Kang De Shi

STYLE NO. : FLAT RESPIRATOR

DATE RECEIVED/DATE TEST STARTED: 2020-04-02

CONCLUSION:

pH VALUE	M
FORMALDEHYDE CONTENT	M
APPEARANCE REQUIREMENT[10 PIECES]	M
BANNED AZO COLOURANTS	M

NOTE: "M" -MEET THE STANDARD'S REQUIREMENT "F" -FAIL TO MEET THE STANDARD'S REQUIREMENT
"---" -NO COMMENT

REMARK:

THIS REPORT IS THE ENGLISH TRANSLATION VERSION OF THE REPORT 200056317.
ALL THE TESTED ITEMS ARE TESTED UNDER THE STANDARD CONDITION (EXCEPT FOR INDICATION).
COPIES OF THE REPORT ARE VALID ONLY RE-STAMPED.
THE EXPERIMENT WAS CARRIED OUT AT No.1, ZHUJIANG ROAD, PANYU DISTRICT, GUANGZHOU, GUANGDONG, P. R. CHINA.

APPROVED BY:

Nan Ma ENGINEER

马楠



PAGE 1 OF 4



GTT




中国认可
国际互认
检测
TESTING
CNAS L10314

TEST REPORT

(Electronic version)

No:200056318



TEST REPORT

(Electronic version)

No:200056318

pH VALUE

(GB/T 7573-2009, 0.1mol/L KCl EXTRACTING SOLUTION)
(GB/T 7573-2009 MOD ISO 3071:2005)

6.4

REQUIREMENT
4.0~8.5
(GB/T 32610-2016)

FORMALDEHYDE CONTENT

(GB/T 2912.1-2009, WATER EXTRACTION METHOD)
(GB/T 2912.1-2009 MOD ISO 14184-1:1998)

NOT DETECTED

REQUIREMENT
≤20 mg/kg
(GB/T 32610-2016)

REMARK: DETECTION LIMIT 20mg/kg

APPEARANCE REQUIREMENT[10 PIECES]

(GB/T 32610-2016 6.1)

PASS

REQUIREMENT
THE SURFACE OF THE MASK IS FREE FROM
BREAKAGE, OIL STAINS, DEFORMATION AND
OTHER OBVIOUS DEFECTS.
(GB/T 32610-2016)



PAGE 3 OF 4

TEST REPORT

(Electronic version)

No: 200056318

BANNED AZO COLOURANTS

TEST METHOD: GB/T 17592-2011, GENERAL PRETREATMENT

<u>FORBIDDEN AMINE</u>	<u>CAS NO.</u>	<u>REQUIREMENT</u> (mg/kg)	<u>RESULT</u> (mg/kg)
4-aminobiphenyl	[92-67-1]	≤20	N
benzidine	[92-87-5]	≤20	N
4-chloro-o-toluidine	[95-69-2]	≤20	N
2-naphthylamine	[91-59-8]	≤20	N
o-aminoazotoluene	[97-56-3]	≤20	N
5-nitro-o-toluidine	[99-55-8]	≤20	N
p-chloroaniline	[106-47-8]	≤20	N
2,4-diaminoanisole	[615-05-4]	≤20	N
4,4'-diaminobiphenylmethane	[101-77-9]	≤20	N
3,3'-dichlorobenzidine	[91-94-1]	≤20	N
3,3'-dimethoxybenzidine	[119-90-4]	≤20	N
3,3'-dimethylbenzidine	[119-93-7]	≤20	N
3,3'-dimethyl-4,4'-diaminobiphenylmethane	[838-88-0]	≤20	N
p-cresidine	[120-71-8]	≤20	N
4,4'-methylene-bis-(2-chloroaniline)	[101-14-4]	≤20	N
4,4'-oxydianiline	[101-80-4]	≤20	N
4,4'-thiodianiline	[139-65-1]	≤20	N
o-toluidine	[95-53-4]	≤20	N
2,4-toluyldiamine	[95-80-7]	≤20	N
2,4,5-trimethylaniline	[137-17-7]	≤20	N
o-anisidine	[90-04-0]	≤20	N
2,4-xylidine	[95-68-1]	≤20	N
2,6-xylidine	[87-62-7]	≤20	N
4-aminoazobenzene	[60-09-3]	≤20	N

"N" MEANS LESS THAN THE DETECTION LIMIT OF 5mg/kg
RATING ACCORDING TO GB/T 32610-2016.

OVERALL RATING PASS



—End of Report—