

ACCREDITED REFERENCE MATERIAL (GA ARM) : CATALOGUE

ACCREDITED TO ISO 17034 : General requirements for the competence of reference material producers.

(Available in 250g units, except GA ARM048, GA ARM 062, GA ARM 066 which is in 50g units.) (GA ARM045 onwards available in 200g units)

LIST OF SPECIFICATIONS FOR GENERAL ANALYSIS SAMPLES-**Version 31/05/2026**

PULVERISED COAL LESS THAN 212 μ m, < 60 MESH (DRY BASE RESULTS)

PRODUCT	ISO ASH%	ISO VOLATILES (%)	CALORIFIC VALUE (MJ/Kg)	TOTAL SULPHUR (%)	PHOSPHORUS IN COAL (%)	TOTAL CARBON (%)	HYDROGEN (%)	NITROGEN (%)	AFT (REDUCING) DEFORMATION (°C)	AFT (REDUCING) SOFTENING (°C)	AFT (REDUCING) HEMISPHERE (°C)	AFT (REDUCING) FLOW (°C)
GA -ARM025	-	-	-	0.59	0.036	-	-	-	-	-	-	-
GA -ARM026	-	-	-	2.10	-	-	-	-	-	-	-	-
GA- ARM 027	-	-	-	0.82	0.015	-	-	-	1350	1375	1405	1430
GA -ARM 028	-	-	-	0.94	-	-	-	-	1228	1253	1277	1325
GA ARM 029	-	-	-	2.24	-	-	-	-	-	-	-	-
GA -ARM 034	-	-	-	0.74	-	-	-	-	-	-	-	-
GA- ARM 036	13.18	10.03	29.70	2.23	0.091	-	-	-	-	-	-	-
GA- ARM 039	14.47	19.13	29.27	1.65	0.036	-	-	-	-	-	-	-
GA ARM 045	10.44	27.44	28.79	0.60	0.034	73.35	4.02	1.77	-	-	-	-
GA- ARM 046	16.57	20.08	28.65	1.47	0.037	71.40	3.77	1.86	1330	1359	1390	1430

PRODUCT	ISO ASH%	ISO VOLATILES (%)	CALORIFIC VALUE (MJ/Kg)	TOTAL SULPHUR (%)	PHOSPHORUS IN COAL (%)	TOTAL CARBON (%)	HYDROGEN (%)	NITROGEN (%)	AFT (REDUCING) DEFORMATION (°C)	AFT (REDUCING) SOFTENING (°C)	AFT (REDUCING) HEMISPHERE (°C)	AFT (REDUCING) FLOW (°C)
GA- ARM 048	4.91	3.33	31.98	0.25	0.018	91.25	1.34	0.95	-	-	-	-
GA- ARM 049	25.02	17.98	23.86	0.96	0.053	63.55	2.99	1.39	-	-	-	-
GA-ARM 051	30.70	22.06	21.10	1.14	0.075	54.85	3.09	1.32	1370	1400	1420	1460
GA- ARM 052	23.76	23.19	24.01	0.90	0.030	61.88	3.38	1.49	-	-	-	-
GA- ARM 054	22.67	25.27	24.56	0.93	0.036	62.75	3.53	1.51	1420	1445	1465	1490
GA- ARM 055	18.79	24.07	25.23	0.55	0.093	65.96	3.49	1.58	1420	1440	1460	1490
GA- ARM 056	33.47	19.64	20.13	1.00	0.044	53.38	2.98	1.20	-	-	-	-
GA- ARM 058	15.62	28.37	27.40	0.54	0.089	69.27	3.95	1.68	1405	1430	1450	1490
GA- ARM 060	18.88	21.05	27.12	1.59	0.035	68.54	3.62	1.78	1370	1400	1440	1470
GA- ARM 061	12.26	35.60	29.44	1.07	0.024	71.34	4.73	1.57	1405	1430	1460	1490
GA- ARM 062	-	-	-	6.00								
GA-ARM 063	19.49	33.04	27.24	1.21	0.014	66.59	4.41	1.45	1230	1300	1340	1380
GA -ARM 064	35.55	24.92	19.93	1.56	0.057	50.38	3.25	1.13	1340	1370	1400	1450
GA-ARM 065	23.21	24.29	24.10	0.62	0.047	62.65	3.44	1.52	1370	1400	1430	1460
GA -ARM 066	-	-	-	4.39	-	-	-	-	-	-	-	-
GA-ARM 067	31.78	22.43	20.62	1.14	0.075	-	-	-	1390	1420	1445	1480
GA-ARM 068	42.79	21.35	15.82	0.88	-	-	-	-	1439	1465	1485	1500

PRODUCT	ASTM ASH%	ASTM VOLATILES (%)	Chlorine	FLUORINE
GA- ARM039	14.65	19.78	-	-
GA- ARM045	-	-	-	163
GA- ARM046	16.65	20.44	-	142
GA- ARM048	4.99	3.62	-	-
GA- ARM 049	25.24	18.79	-	-
GA- ARM 054	22.71	25.98	-	-
GA- ARM 055	18.90	24.74	-	158
GA- ARM 056	33.55	20.28	-	-
GA- ARM 058	15.70	29.03	-	-
GA- ARM 060	19.35	21.19	-	-
GA- ARM 064	35.49	25.46	-	-
GA- ARM 067	31.86	23.17	-	-

**ASTM Ash, Volatile matter and Quick ash not accredited

**Chlorine /Florine not accredited, and values are indicative

Accredited Benzoic acid pellets, traceable to NIST

**Available in 1g and 0.5g weights
100 pills per bottle.**

CONTACT DETAILS

Mail: ravi@coalconcepts.co.za

Phone: +27 35 751 2446 Mobile: +27 83 6500 151

Unit 10, 81 Ceramic Curve , Richards Bay, KZN , South Africa