



Division of Criminal Justice Services

CERTIFICATION PURSUANT TO CPLR 4518 OF RECORDS MAINTAINED IN THE REGULAR COURSE OF BUSINESS

I, Steven J. Carluccio, Highway Safety Equipment Technical Supervisor, Office of Public Safety, New York State Division of Criminal Justice Services, 80 South Swan Street, Albany, New York 12210, having been duly designated and authorized by the Commissioner of the Division of Criminal Justice Services, do hereby certify and authenticate, as provided by subdivision c of Rule 4518 of the Civil Practice Law and Rules, that the document annexed hereto is an exact copy of an electronic record of the Office of Public Safety which is in my possession, custody and control. This record, once created, is stored in an electronic format that cannot thereafter be altered or modified.

I further certify that the original electronic record of inspection/maintenance/calibration of INTOXIMETERS INC., ALCO-SENSOR FST SCREENER, serial number 052816, performed on July 20, 2022 by Office of Public Safety employee Jonathan Mascolo, of which the annexed is an exact copy, was made in the regular course of business of the New York State Division of Criminal Justice Services, Office of Public Safety, that such record was made at the time such inspection/ maintenance/calibration was performed or within a reasonable time thereafter, and further that it is the regular course of the Office of Public Safety's business to make such records at the time such inspection/maintenance /calibration is performed, or within a reasonable time thereafter, and to provide such instrument's records to the agency that requested them.

A handwritten signature in black ink that reads "Steven J. Carluccio".

Digitally signed under ESRA
by Steven J. Carluccio
on 07/28/2022 02:51 PM

Steven J. Carluccio
Highway Safety Equipment Technical Supervisor
Office of Public Safety

**BREATH SCREENING DEVICE
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: North Syracuse Village Police Department

Manufacturer: INTOXIMETERS INC.

Model: ALCO-SENSOR FST

Serial Number: 052816

Date: July 20, 2022

Reference Solution Lot Number : 22100

Reference Solution Standard : 0.10% at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.100 %	34.00 degrees centigrade
2	0.102 %	34.02 degrees centigrade
3	0.103 %	34.02 degrees centigrade
4	0.102 %	34.01 degrees centigrade

[Faint, illegible text or stamp]

I hereby certify that I have performed any necessary maintenance procedures and verified the calibration of INTOXIMETERS INC. screener model ALCO-SENSOR FST, serial number 052816, and have determined that it is accurate.

I further certify that I made the entries in these records of inspection/maintenance/calibration at the time that such inspection/maintenance/calibration of the above identified screener was performed, or within a reasonable time thereafter.

Jonathan Mascolo

Digitally signed under ESRA
by Jonathan Mascolo
on 07/20/2022 01:41 PM

**Jonathan Mascolo
Highway Safety Equipment Technician
Office of Public Safety**

**BREATH SCREENING DEVICE
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: North Syracuse Village Police Department

Manufacturer: INTOXIMETERS INC.

Model: ALCO-SENSOR FST

Serial Number: 052816

Date: July 20, 2022

1. General appearance and condition of this device -----: OK

2. Check battery indicators -----: OK

3. Check zero flashing -----: OK

**4. Record any repairs made to this instrument and list services provided : ADJUSTED,
VERIFIED CALIBRATION**

[Faint, illegible handwritten text]

Jonathan Mascolo

Digitally signed under ESRA
by Jonathan Mascolo
on 07/20/2022 01:41 PM

**Jonathan Mascolo
Highway Safety Equipment Technician
Office of Public Safety**


SERVICE AUTHORIZATION FORM
 Highway Safety Technology Unit
 80 South Swan Street
 Albany, NY 12210
 Ph: 518-485-7636 / 518-402-0689
 Fax: 518-457-6869

THIS FORM IS USED TO AUTHORIZE THE DIVISION OF CRIMINAL JUSTICE SERVICES TO PERFORM SERVICE ON THE INSTRUMENT IDENTIFIED. THIS AUTHORIZATION MUST ACCOMPANY EACH PIECE OF EQUIPMENT WHEN IT IS PRESENTED TO THE EQUIPMENT REPAIR CENTER FOR SERVICE. **PLEASE SIGN AND DATE FORM.**

SECTION I - AGENCY INFORMATION

Name of Agency North Syracuse Police Department		Today's Date 07/20/2022
Address 600 South Bay Road	City, State, ZIP North Syracuse, NY 13212	
Contact Person Sergeant Jeffrey D. Tripp	Contact Telephone 315-418-7938/ 315-458-5670	
Contact Email jtripp@northsyracuse.ny.org	Best Way to Reach You During Appointment Period 315-418-7938 (work cell)	

SECTION II - INSTRUMENT INFORMATION

Instrument Manufacturer (Please Circle) NPAS DMT GUTH Applied Concepts(Stalker) Decatur(Genesis) Kustom(Eagle/Falcon) MPH Alco-Sensor Other:		
Model Number Alco Sensor FST	Instrument Serial Number 052816	
Antenna Serial Number (if applicable) Antenna 1 - / Antenna 2 -	Type of maintenance (check appropriate) <input checked="" type="checkbox"/> Calibration Check <input type="checkbox"/> Repair	
Brief Description of Malfunction Calibration Check		
List all parts, cables and/or accessories being submitted. Tuning forks must accompany all radar units (if applicable). (Please Circle)		
DMT/SIM: DMT SIM Tubes Keyboard Other: <u>CASE/TUBES</u>	Radar: Control Unit Antenna(s) Antenna Cable(s) Power Cord Forks Other:	
I acknowledge I am signing this Service Authorization Form as an act and deed of said agency, organization, or municipality, and that I am duly authorized to sign same for the uses and purposes mentioned herein. PLEASE SIGN AND DATE.		
Signature 	Date 07/20/2022	

DCJS USE ONLY

Initial Test: <u>0.109</u> % <u>34.01</u> °C	Test 1: <u>0.100</u> % <u>34.00</u> °C
Ref Std Lot # <u>22100</u> <u>0.100</u> %	Test 2: <u>0.102</u> % <u>34.02</u> °C
Re-adjusted	Test 3: <u>0.103</u> % <u>34.02</u> °C
	Test 4: <u>0.102</u> % <u>34.01</u> °C



Division of Criminal Justice Services

CERTIFICATION PURSUANT TO CPLR 4518 OF RECORDS MAINTAINED IN THE REGULAR COURSE OF BUSINESS

I, Steven J. Carluccio, Highway Safety Equipment Technical Supervisor, Office of Public Safety, New York State Division of Criminal Justice Services, 80 South Swan Street, Albany, New York 12210, having been duly designated and authorized by the Commissioner of the Division of Criminal Justice Services, do hereby certify and authenticate, as provided by subdivision c of Rule 4518 of the Civil Practice Law and Rules, that the document annexed hereto is an exact copy of an electronic record of the Office of Public Safety which is in my possession, custody and control. This record, once created, is stored in an electronic format that cannot thereafter be altered or modified.

I further certify that the original electronic record of inspection/maintenance/calibration of INTOXIMETERS INC., ALCO-SENSOR FST SCREENER, serial number 052827, performed on July 20, 2022 by Office of Public Safety employee George Coffin, of which the annexed is an exact copy, was made in the regular course of business of the New York State Division of Criminal Justice Services, Office of Public Safety, that such record was made at the time such inspection/ maintenance/calibration was performed or within a reasonable time thereafter, and further that it is the regular course of the Office of Public Safety's business to make such records at the time such inspection/maintenance /calibration is performed, or within a reasonable time thereafter, and to provide such instrument's records to the agency that requested them.

A handwritten signature in black ink that reads "Steven J. Carluccio".

Digitally signed under ESRA
by Steven J. Carluccio
on 07/28/2022 02:52 PM

Steven J. Carluccio
Highway Safety Equipment Technical Supervisor
Office of Public Safety

**BREATH SCREENING DEVICE
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: North Syracuse Village Police Department

Manufacturer: INTOXIMETERS INC.

Model: ALCO-SENSOR FST

Serial Number: 052827

Date: July 20, 2022

Reference Solution Lot Number : 22100

Reference Solution Standard : 0.10% at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.097 %	34.00 degrees centigrade
2	0.099 %	34.00 degrees centigrade
3	0.099 %	34.01 degrees centigrade
4	0.099 %	34.01 degrees centigrade

[Faint, illegible signature or stamp]

I hereby certify that I have performed any necessary maintenance procedures and verified the calibration of INTOXIMETERS INC. screener model ALCO-SENSOR FST, serial number 052827, and have determined that it is accurate.

I further certify that I made the entries in these records of inspection/maintenance/calibration at the time that such inspection/maintenance/calibration of the above identified screener was performed, or within a reasonable time thereafter.

George Coffin

Digitally signed under ESRA
by George Coffin
on 07/20/2022 02:51 PM

**George Coffin
Highway Safety Equipment Technician
Office of Public Safety**

**BREATH SCREENING DEVICE
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: North Syracuse Village Police Department

Manufacturer: INTOXIMETERS INC.

Model: ALCO-SENSOR FST

Serial Number: 052827

Date: July 20, 2022

1. General appearance and condition of this device -----: OK

2. Check battery indicators -----: OK

3. Check zero flashing -----: OK

**4. Record any repairs made to this instrument and list services provided : VERIFIED
CALIBRATION**

[Faint, illegible text, possibly a stamp or watermark]



Digitally signed under ESRA
by George Coffin
on 07/20/2022 02:51 PM

**George Coffin
Highway Safety Equipment Technician
Office of Public Safety**


SERVICE AUTHORIZATION FORM
Highway Safety Technology Unit
80 South Swan Street
Albany, NY 12210
Ph: 518-485-7636 / 518-402-0689
Fax: 518-457-6869

THIS FORM IS USED TO AUTHORIZE THE DIVISION OF CRIMINAL JUSTICE SERVICES TO PERFORM SERVICE ON THE INSTRUMENT IDENTIFIED. THIS AUTHORIZATION MUST ACCOMPANY EACH PIECE OF EQUIPMENT WHEN IT IS PRESENTED TO THE EQUIPMENT REPAIR CENTER FOR SERVICE. **PLEASE SIGN AND DATE FORM.**

SECTION I - AGENCY INFORMATION

Name of Agency North Syracuse Police Department		Today's Date 07/20/2022
Address 600 South Bay Road		City, State, ZIP North Syracuse, NY 13212
Contact Person Sergeant Jeffrey D. Tripp		Contact Telephone 315-418-7938/ 315-458-5670
Contact Email jtripp@northsyracuse.ny.gov		Best Way to Reach You During Appointment Period 315-418-7938 (work cell)

SECTION II - INSTRUMENT INFORMATION

Instrument Manufacturer (Please Circle) NPAS DMT GUTH Applied Concepts(Stalker) Decatur(Genesis) Kustom(Eagle/Falcon) MPH Alco-Sensor Other:		
Model Number Alco Sensor FST		Instrument Serial Number 052827 ✓
Antenna Serial Number (if applicable) Antenna 1 - / Antenna 2 -		Type of maintenance (check appropriate) <input checked="" type="checkbox"/> Calibration Check <input type="checkbox"/> Repair
Brief Description of Malfunction Calibration Check		
List all parts, cables and/or accessories being submitted. Tuning forks must accompany all radar units (if applicable). (Please Circle) DMT/SIM: Radar: DMT SIM Tubes Control Unit Antenna(s) Antenna Cable(s) Keyboard Other: <u>CASE / 2 FUNNELS / GUIDE / TUBES</u> Power Cord Forks Other:		
<i>I acknowledge I am signing this Service Authorization Form as an act and deed of said agency, organization, or municipality, and that I am duly authorized to sign same for the uses and purposes mentioned herein. PLEASE SIGN AND DATE.</i>		
Signature 		Date 07/20/2022

DCJS USE ONLY

Initial Test: <u>0.097</u> % <u>34.00</u> °C	Test 1: <u>0.097</u> % <u>34.00</u> °C
Ref Std Lot # <u>22100</u> <u>0.10</u> %	Test 2: <u>0.099</u> % <u>34.00</u> °C
	Test 3: <u>0.099</u> % <u>34.01</u> °C
	Test 4: <u>0.099</u> % <u>34.01</u> °C



Division of Criminal Justice Services

CERTIFICATION PURSUANT TO CPLR 4518 OF RECORDS MAINTAINED IN THE REGULAR COURSE OF BUSINESS

I, Steven J. Carluccio, Highway Safety Equipment Technical Supervisor, Office of Public Safety, New York State Division of Criminal Justice Services, 80 South Swan Street, Albany, New York 12210, having been duly designated and authorized by the Commissioner of the Division of Criminal Justice Services, do hereby certify and authenticate, as provided by subdivision c of Rule 4518 of the Civil Practice Law and Rules, that the document annexed hereto is an exact copy of an electronic record of the Office of Public Safety which is in my possession, custody and control. This record, once created, is stored in an electronic format that cannot thereafter be altered or modified.

I further certify that the original electronic record of inspection/maintenance/calibration of INTOXIMETERS INC., ALCO-SENSOR FST SCREENER, serial number 116206, performed on July 20, 2022 by Office of Public Safety employee Jonathan Mascolo, of which the annexed is an exact copy, was made in the regular course of business of the New York State Division of Criminal Justice Services, Office of Public Safety, that such record was made at the time such inspection/ maintenance/calibration was performed or within a reasonable time thereafter, and further that it is the regular course of the Office of Public Safety's business to make such records at the time such inspection/maintenance /calibration is performed, or within a reasonable time thereafter, and to provide such instrument's records to the agency that requested them.

A handwritten signature in black ink that reads "Steven J. Carluccio".

Digitally signed under ESRA
by Steven J. Carluccio
on 07/28/2022 02:52 PM

Steven J. Carluccio
Highway Safety Equipment Technical Supervisor
Office of Public Safety

**BREATH SCREENING DEVICE
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: North Syracuse Village Police Department

Manufacturer: INTOXIMETERS INC.

Model: ALCO-SENSOR FST

Serial Number: 116206

Date: July 20, 2022

Reference Solution Lot Number : 22100

Reference Solution Standard : 0.10% at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.102 %	34.01 degrees centigrade
2	0.100 %	34.00 degrees centigrade
3	0.101 %	34.00 degrees centigrade
4	0.101 %	34.01 degrees centigrade

[Faint, illegible text, possibly a signature or stamp]

I hereby certify that I have performed any necessary maintenance procedures and verified the calibration of INTOXIMETERS INC. screener model ALCO-SENSOR FST, serial number 116206, and have determined that it is accurate.

I further certify that I made the entries in these records of inspection/maintenance/calibration at the time that such inspection/maintenance/calibration of the above identified screener was performed, or within a reasonable time thereafter.

Jonathan Mascolo

Digitally signed under ESRA
by Jonathan Mascolo
on 07/20/2022 01:44 PM

**Jonathan Mascolo
Highway Safety Equipment Technician
Office of Public Safety**

**BREATH SCREENING DEVICE
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: North Syracuse Village Police Department

Manufacturer: INTOXIMETERS INC.

Model: ALCO-SENSOR FST

Serial Number: 116206

Date: July 20, 2022

1. General appearance and condition of this device -----: OK

2. Check battery indicators -----: OK

3. Check zero flashing -----: OK

**4. Record any repairs made to this instrument and list services provided : VERIFIED
CALIBRATION**

[Faint, illegible text, possibly a stamp or watermark]

Jonathan Mascolo

Digitally signed under ESRA
by Jonathan Mascolo
on 07/20/2022 01:44 PM

**Jonathan Mascolo
Highway Safety Equipment Technician
Office of Public Safety**


SERVICE AUTHORIZATION FORM
Highway Safety Technology Unit
80 South Swan Street
Albany, NY 12210
Ph: 518-485-7636 / 518-402-0689
Fax: 518-457-6869

THIS FORM IS USED TO AUTHORIZE THE DIVISION OF CRIMINAL JUSTICE SERVICES TO PERFORM SERVICE ON THE INSTRUMENT IDENTIFIED. THIS AUTHORIZATION MUST ACCOMPANY EACH PIECE OF EQUIPMENT WHEN IT IS PRESENTED TO THE EQUIPMENT REPAIR CENTER FOR SERVICE. PLEASE SIGN AND DATE FORM.

SECTION I - AGENCY INFORMATION

Name of Agency North Syracuse Police Department	Today's Date 07/20/2022
Address 600 South Bay Road	City, State, ZIP North Syracuse, NY 13212
Contact Person Sergeant Jeffrey D. Tripp	Contact Telephone 315-418-7938/ 315-458-5670
Contact Email jtripp@northsyracuse.ny.gov	Best Way to Reach You During Appointment Period 315-418-7938 (work cell)

SECTION II - INSTRUMENT INFORMATION

Instrument Manufacturer (Please Circle) NPAS DMT GUTH Applied Concepts(Stalker) Decatur(Genesis) Kustom(Eagle/Falcon) MPH Alco-Sensor Other:		
Model Number Alco Sensor FST	Instrument Serial Number 116206	
Antenna Serial Number (if applicable) Antenna 1 -- / Antenna 2 --	Type of maintenance (check appropriate) <input checked="checked" type="checkbox"/> Calibration Check <input type="checkbox"/> Repair	
Brief Description of Malfunction Calibration Check		
List all parts, cables and/or accessories being submitted. Tuning forks must accompany all radar units (if applicable). (Please Circle) DMT/SIM: DMT SIM Tubes Keyboard Other: CASE/TUBES/2 FUNNELS/GUIDE/YLW ALL	Radar: Control Unit Antenna(s) Antenna Cable(s) Power Cord Forks Other:	
<i>I acknowledge I am signing this Service Authorization Form as an act and deed of said agency, organization, or municipality, and that I am duly authorized to sign same for the uses and purposes mentioned herein. PLEASE SIGN AND DATE.</i>		
Signature 	Date 07/20/2022	

DCJS USE ONLY

Initial Test: <u>0.101</u> % <u>34.01</u> °C	Test 1: <u>0.102</u> % <u>34.01</u> °C
Ref Std Lot # <u>22100</u> <u>0.100</u> %	Test 2: <u>0.100</u> % <u>34.00</u> °C
	Test 3: <u>0.101</u> % <u>34.00</u> °C
	Test 4: <u>0.101</u> % <u>34.01</u> °C



Division of Criminal Justice Services

CERTIFICATION PURSUANT TO CPLR 4518 OF RECORDS MAINTAINED IN THE REGULAR COURSE OF BUSINESS

I, Steven J. Carluccio, Highway Safety Equipment Technical Supervisor, Office of Public Safety, New York State Division of Criminal Justice Services, 80 South Swan Street, Albany, New York 12210, having been duly designated and authorized by the Commissioner of the Division of Criminal Justice Services, do hereby certify and authenticate, as provided by subdivision c of Rule 4518 of the Civil Practice Law and Rules, that the document annexed hereto is an exact copy of an electronic record of the Office of Public Safety which is in my possession, custody and control. This record, once created, is stored in an electronic format that cannot thereafter be altered or modified.

I further certify that the original electronic record of inspection/maintenance/calibration of INTOXIMETERS INC., ALCO-SENSOR FST SCREENER, serial number 246861, performed on July 20, 2022 by Office of Public Safety employee Jonathan Mascolo, of which the annexed is an exact copy, was made in the regular course of business of the New York State Division of Criminal Justice Services, Office of Public Safety, that such record was made at the time such inspection/ maintenance/calibration was performed or within a reasonable time thereafter, and further that it is the regular course of the Office of Public Safety's business to make such records at the time such inspection/maintenance /calibration is performed, or within a reasonable time thereafter, and to provide such instrument's records to the agency that requested them.

A handwritten signature in black ink that reads "Steven J. Carluccio".

Digitally signed under ESRA
by Steven J. Carluccio
on 07/28/2022 02:53 PM

Steven J. Carluccio
Highway Safety Equipment Technical Supervisor
Office of Public Safety

**BREATH SCREENING DEVICE
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: North Syracuse Village Police Department

Manufacturer: INTOXIMETERS INC.

Model: ALCO-SENSOR FST

Serial Number: 246861

Date: July 20, 2022

Reference Solution Lot Number : 22100

Reference Solution Standard : 0.10% at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.100 %	34.01 degrees centigrade
2	0.100 %	34.00 degrees centigrade
3	0.101 %	34.00 degrees centigrade
4	0.100 %	34.01 degrees centigrade

I hereby certify that I have performed any necessary maintenance procedures and verified the calibration of INTOXIMETERS INC. screener model ALCO-SENSOR FST, serial number 246861, and have determined that it is accurate.

I further certify that I made the entries in these records of inspection/maintenance/calibration at the time that such inspection/maintenance/calibration of the above identified screener was performed, or within a reasonable time thereafter.



Digitally signed under ESRA
by Jonathan Mascolo
on 07/20/2022 01:43 PM

**Jonathan Mascolo
Highway Safety Equipment Technician
Office of Public Safety**

**BREATH SCREENING DEVICE
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: North Syracuse Village Police Department

Manufacturer: INTOXIMETERS INC.

Model: ALCO-SENSOR FST

Serial Number: 246861

Date: July 20, 2022

1. General appearance and condition of this device -----: OK

2. Check battery indicators -----: OK

3. Check zero flashing -----: OK

**4. Record any repairs made to this instrument and list services provided : VERIFIED
CALIBRATION**

[Faint, illegible text, possibly a stamp or watermark]

Jonathan Mascolo

Digitally signed under ESRA
by Jonathan Mascolo
on 07/20/2022 01:43 PM

**Jonathan Mascolo
Highway Safety Equipment Technician
Office of Public Safety**


SERVICE AUTHORIZATION FORM
Highway Safety Technology Unit
80 South Swan Street
Albany, NY 12210
Ph: 518-485-7636 / 518-402-0689
Fax: 518-457-6869

THIS FORM IS USED TO AUTHORIZE THE DIVISION OF CRIMINAL JUSTICE SERVICES TO PERFORM SERVICE ON THE INSTRUMENT IDENTIFIED. THIS AUTHORIZATION MUST ACCOMPANY EACH PIECE OF EQUIPMENT WHEN IT IS PRESENTED TO THE EQUIPMENT REPAIR CENTER FOR SERVICE. **PLEASE SIGN AND DATE FORM.**

SECTION I - AGENCY INFORMATION

Name of Agency North Syracuse Police Department		Today's Date 07/20/2022
Address 600 South Bay Road	City, State, ZIP North Syracuse, NY 13212	
Contact Person Sergeant Jeffrey D. Tripp	Contact Telephone 315-418-7938/ 315-458-5670	
Contact Email jtripp@northsyracuse.ny.org	Best Way to Reach You During Appointment Period 315-418-7938 (work cell)	

SECTION II - INSTRUMENT INFORMATION

Instrument Manufacturer (Please Circle) NPAS DMT GUTH Applied Concepts(Stalker) Decatur(Genesis) Kustom(Eagle/Falcon) MPH Alco-Sensor Other:		
Model Number Alco Sensor FST	Instrument Serial Number 246861	
Antenna Serial Number (if applicable) Antenna 1 - / Antenna 2 -	Type of maintenance (check appropriate) <input checked="" type="checkbox"/> Calibration Check <input type="checkbox"/> Repair	
Brief Description of Malfunction Calibration Check		
List all parts, cables and/or accessories being submitted. Tuning forks must accompany all radar units (if applicable). (Please Circle)		
DMT/SIM: DMT SIM Tubes Keyboard Other: <u>CASE/TUBES/GUIDE/2 FUNDERS/YLW ACC.</u>	Radar: Control Unit Antenna(s) Antenna Cable(s) Power Cord Forks Other: _____	
I acknowledge I am signing this Service Authorization Form as an act and deed of said agency, organization, or municipality, and that I am duly authorized to sign same for the uses and purposes mentioned herein. PLEASE SIGN AND DATE.		
Signature 	Date 07/20/2022	

DCJS USE ONLY

Initial Test: <u>0.101</u> % <u>34.02</u> °C Ref Std Lot # <u>22100</u> <u>0.100</u> %	Test 1: <u>0.100</u> % <u>34.01</u> °C Test 2: <u>0.100</u> % <u>34.00</u> °C Test 3: <u>0.101</u> % <u>34.00</u> °C Test 4: <u>0.100</u> % <u>34.01</u> °C
---	--



Division of Criminal Justice Services

CERTIFICATION PURSUANT TO CPLR 4518 OF RECORDS MAINTAINED IN THE REGULAR COURSE OF BUSINESS

I, Steven J. Carluccio, Highway Safety Equipment Technical Supervisor, Office of Public Safety, New York State Division of Criminal Justice Services, 80 South Swan Street, Albany, New York 12210, having been duly designated and authorized by the Commissioner of the Division of Criminal Justice Services, do hereby certify and authenticate, as provided by subdivision c of Rule 4518 of the Civil Practice Law and Rules, that the document annexed hereto is an exact copy of an electronic record of the Office of Public Safety which is in my possession, custody and control. This record, once created, is stored in an electronic format that cannot thereafter be altered or modified.

I further certify that the original electronic record of inspection/maintenance/calibration of INTOXIMETERS INC., ALCO-SENSOR FST SCREENER, serial number 247349, performed on July 20, 2022 by Office of Public Safety employee George Coffin, of which the annexed is an exact copy, was made in the regular course of business of the New York State Division of Criminal Justice Services, Office of Public Safety, that such record was made at the time such inspection/ maintenance/calibration was performed or within a reasonable time thereafter, and further that it is the regular course of the Office of Public Safety's business to make such records at the time such inspection/maintenance /calibration is performed, or within a reasonable time thereafter, and to provide such instrument's records to the agency that requested them.

A handwritten signature in black ink that reads "Steven J. Carluccio".

Digitally signed under ESRA
by Steven J. Carluccio
on 07/28/2022 02:53 PM

Steven J. Carluccio
Highway Safety Equipment Technical Supervisor
Office of Public Safety

**BREATH SCREENING DEVICE
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: North Syracuse Village Police Department

Manufacturer: INTOXIMETERS INC.

Model: ALCO-SENSOR FST

Serial Number: 247349

Date: July 20, 2022

Reference Solution Lot Number : 22100

Reference Solution Standard : 0.10% at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.099 %	34.00 degrees centigrade
2	0.096 %	34.01 degrees centigrade
3	0.100 %	34.01 degrees centigrade
4	0.099 %	34.00 degrees centigrade

[Faint, illegible signature or stamp]

I hereby certify that I have performed any necessary maintenance procedures and verified the calibration of INTOXIMETERS INC. screener model ALCO-SENSOR FST, serial number 247349, and have determined that it is accurate.

I further certify that I made the entries in these records of inspection/maintenance/calibration at the time that such inspection/maintenance/calibration of the above identified screener was performed, or within a reasonable time thereafter.

George Coffin

Digitally signed under ESRA
by George Coffin
on 07/20/2022 02:47 PM

**George Coffin
Highway Safety Equipment Technician
Office of Public Safety**

**BREATH SCREENING DEVICE
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: North Syracuse Village Police Department

Manufacturer: INTOXIMETERS INC.

Model: ALCO-SENSOR FST

Serial Number: 247349

Date: July 20, 2022

1. General appearance and condition of this device -----: OK

2. Check battery indicators -----: OK

3. Check zero flashing -----: OK

**4. Record any repairs made to this instrument and list services provided : VERIFIED
CALIBRATION**

[Faint, illegible handwritten notes or signature]



Digitally signed under ESRA
by George Coffin
on 07/20/2022 02:47 PM

**George Coffin
Highway Safety Equipment Technician
Office of Public Safety**


SERVICE AUTHORIZATION FORM
Highway Safety Technology Unit
80 South Swan Street
Albany, NY 12210
Ph: 518-485-7636 / 518-402-0689
Fax: 518-457-6869

THIS FORM IS USED TO AUTHORIZE THE DIVISION OF CRIMINAL JUSTICE SERVICES TO PERFORM SERVICE ON THE INSTRUMENT IDENTIFIED. THIS AUTHORIZATION MUST ACCOMPANY EACH PIECE OF EQUIPMENT WHEN IT IS PRESENTED TO THE EQUIPMENT REPAIR CENTER FOR SERVICE. **PLEASE SIGN AND DATE FORM.**

SECTION I - AGENCY INFORMATION

Name of Agency North Syracuse Police Department		Today's Date 07/20/2022
Address 600 South Bay Road		City, State, ZIP North Syracuse, NY 13212
Contact Person Sergeant Jeffrey D. Tripp	Contact Telephone 315-418-7938 / 315-458-5670	
Contact Email jtripp@northsyracuse.ny.org	Best Way to Reach You During Appointment Period 315-418-7938 (work cell)	

SECTION II - INSTRUMENT INFORMATION

Instrument Manufacturer (Please Circle) NPAS DMT GUTH Applied Concepts(Stalker) Decatur(Genesis) Kustom(Eagle/Falcon) MPH Alco-Sensor Other: _____			
Model Number Alco Sensor FST		Instrument Serial Number 247349 ✓	
Antenna Serial Number (if applicable) Antenna 1 - _____ / Antenna 2 - _____		Type of maintenance (check appropriate) <input checked="" type="checkbox"/> Calibration Check <input type="checkbox"/> Repair	
Brief Description of Malfunction Calibration Check			
List all parts, cables and/or accessories being submitted. Tuning forks must accompany all radar units (if applicable). (Please Circle) DMT/SIM: _____ Radar: _____ DMT SIM Tubes Control Unit Antenna(s) Antenna Cable(s) Keyboard Other: CASE/TUBES/2 FUNNELS/GUIDE/YELLOW ACC. Power Cord Forks Other: _____			
<i>I acknowledge I am signing this Service Authorization Form as an act and deed of said agency, organization, or municipality, and that I am duly authorized to sign same for the uses and purposes mentioned herein. PLEASE SIGN AND DATE.</i>			
Signature 		Date 07/20/2022	

DCJS USE ONLY

Initial Test: <u>0.044</u> % <u>34.00</u> °C	Test 1: <u>0.099</u> % <u>34.00</u> °C
Ref Std Lot # <u>22100</u> <u>0.10</u> %	Test 2: <u>0.096</u> % <u>34.01</u> °C
	Test 3: <u>0.100</u> % <u>34.01</u> °C
	Test 4: <u>0.099</u> % <u>34.00</u> °C



Division of Criminal Justice Services

CERTIFICATION PURSUANT TO CPLR 4518 OF RECORDS MAINTAINED IN THE REGULAR COURSE OF BUSINESS

I, Steven J. Carluccio, Highway Safety Equipment Technical Supervisor, Office of Public Safety, New York State Division of Criminal Justice Services, 80 South Swan Street, Albany, New York 12210, having been duly designated and authorized by the Commissioner of the Division of Criminal Justice Services, do hereby certify and authenticate, as provided by subdivision c of Rule 4518 of the Civil Practice Law and Rules, that the document annexed hereto is an exact copy of an electronic record of the Office of Public Safety which is in my possession, custody and control. This record, once created, is stored in an electronic format that cannot thereafter be altered or modified.

I further certify that the original electronic record of inspection/maintenance/calibration of INTOXIMETERS INC., ALCO-SENSOR SCREENER, serial number 1038116, performed on July 20, 2022 by Office of Public Safety employee Jonathan Mascolo, of which the annexed is an exact copy, was made in the regular course of business of the New York State Division of Criminal Justice Services, Office of Public Safety, that such record was made at the time such inspection/ maintenance/calibration was performed or within a reasonable time thereafter, and further that it is the regular course of the Office of Public Safety's business to make such records at the time such inspection/maintenance /calibration is performed, or within a reasonable time thereafter, and to provide such instrument's records to the agency that requested them.

A handwritten signature in black ink that reads "Steven J. Carluccio".

Digitally signed under ESRA
by Steven J. Carluccio
on 07/28/2022 02:56 PM

Steven J. Carluccio
Highway Safety Equipment Technical Supervisor
Office of Public Safety

**BREATH SCREENING DEVICE
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: North Syracuse Village Police Department

Manufacturer: INTOXIMETERS INC.

Model: ALCO-SENSOR

Serial Number: 1038116

Date: July 20, 2022

Reference Solution Lot Number : 22100

Reference Solution Standard : 0.10% at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.10 %	33.99 degrees centigrade
2	0.10 %	34.00 degrees centigrade
3	0.10 %	33.99 degrees centigrade
4	0.10 %	33.99 degrees centigrade

I hereby certify that I have performed any necessary maintenance procedures and verified the calibration of INTOXIMETERS INC. screener model ALCO-SENSOR, serial number 1038116, and have determined that it is accurate.

I further certify that I made the entries in these records of inspection/maintenance/calibration at the time that such inspection/maintenance/calibration of the above identified screener was performed, or within a reasonable time thereafter.



Digitally signed under ESRA
by Jonathan Mascolo
on 07/20/2022 02:28 PM

**Jonathan Mascolo
Highway Safety Equipment Technician
Office of Public Safety**

**BREATH SCREENING DEVICE
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: North Syracuse Village Police Department

Manufacturer: INTOXIMETERS INC.

Model: ALCO-SENSOR

Serial Number: 1038116

Date: July 20, 2022

1. General appearance and condition of this device -----: OK

2. Check battery indicators -----: OK

3. Check zero flashing -----: OK

**4. Record any repairs made to this instrument and list services provided : ADJUSTED,
VERIFIED CALIBRATION**

[Faint, illegible handwritten notes or signatures]

Jonathan Mascolo

Digitally signed under ESRA
by Jonathan Mascolo
on 07/20/2022 02:28 PM

**Jonathan Mascolo
Highway Safety Equipment Technician
Office of Public Safety**

SERVICE AUTHORIZATION FORM
Highway Safety Technology Unit
80 South Swan Street
Albany, NY 12210
Ph: 518-485-7636 / 518-402-0689
Fax: 518-457-6869

THIS FORM IS USED TO AUTHORIZE THE DIVISION OF CRIMINAL JUSTICE SERVICES TO PERFORM SERVICE ON THE INSTRUMENT IDENTIFIED. THIS AUTHORIZATION MUST ACCOMPANY EACH PIECE OF EQUIPMENT WHEN IT IS PRESENTED TO THE EQUIPMENT REPAIR CENTER FOR SERVICE. **PLEASE SIGN AND DATE FORM.**

SECTION I – AGENCY INFORMATION

Name of Agency North Syracuse Police Department		Today's Date 07/20/2022
Address 600 South Bay Road	City, State, ZIP North Syracuse, NY 13212	
Contact Person Sergeant Jeffrey D. Tripp	Contact Telephone 315-418-7938/ 315-458-5670	
Contact Email jtripp@northsyracuse.ny.gov	Best Way to Reach You During Appointment Period 315-418-7938 (work cell)	

SECTION II – INSTRUMENT INFORMATION

Instrument Manufacturer (Please Circle) NPAS DMT GUTH Applied Concepts(Stalker) Decatur(Genesis) Kustom(Eagle/Falcon) MPH Alco-Sensor Other:		
Model Number Alco Sensor 2 Digit	Instrument Serial Number 1038116	
Antenna Serial Number (if applicable) Antenna 1 – / Antenna 2 –	Type of maintenance (check appropriate) <input checked="" type="checkbox"/> Calibration Check <input type="checkbox"/> Repair	
Brief Description of Malfunction Calibration Check		
List all parts, cables and/or accessories being submitted. Tuning forks must accompany all radar units (if applicable). (Please Circle)		
DMT/SIM: DMT SIM Tubes Keyboard Other: <i>CASE / TUBES</i>	Radar: Control Unit Antenna(s) Antenna Cable(s) Power Cord Forks Other: _____	
I acknowledge I am signing this Service Authorization Form as an act and deed of said agency, organization, or municipality, and that I am duly authorized to sign same for the uses and purposes mentioned herein. PLEASE SIGN AND DATE.		
Signature <i>Sgt Jeffrey D. Tripp</i>		Date 07/20/2022

DCJS USE ONLY

Initial Test: <u>0.09</u> % <u>34.02</u> °C	Test 1: <u>0.10</u> % <u>33.99</u> °C
Ref Std Lot # <u>22100</u> <u>0.100</u> %	Test 2: <u>0.10</u> % <u>34.00</u> °C
Re-adjusted	Test 3: <u>0.10</u> % <u>33.99</u> °C
	Test 4: <u>0.10</u> % <u>33.99</u> °C



Division of Criminal Justice Services

CERTIFICATION PURSUANT TO CPLR 4518 OF RECORDS MAINTAINED IN THE REGULAR COURSE OF BUSINESS

I, Steven J. Carluccio, Highway Safety Equipment Technical Supervisor, Office of Public Safety, New York State Division of Criminal Justice Services, 80 South Swan Street, Albany, New York 12210, having been duly designated and authorized by the Commissioner of the Division of Criminal Justice Services, do hereby certify and authenticate, as provided by subdivision c of Rule 4518 of the Civil Practice Law and Rules, that the document annexed hereto is an exact copy of an electronic record of the Office of Public Safety which is in my possession, custody and control. This record, once created, is stored in an electronic format that cannot thereafter be altered or modified.

I further certify that the original electronic record of inspection/maintenance/calibration of INTOXIMETERS INC., ALCO-SENSOR III SCREENER, serial number 1062753, performed on July 20, 2022 by Office of Public Safety employee Jonathan Mascolo, of which the annexed is an exact copy, was made in the regular course of business of the New York State Division of Criminal Justice Services, Office of Public Safety, that such record was made at the time such inspection/ maintenance/calibration was performed or within a reasonable time thereafter, and further that it is the regular course of the Office of Public Safety's business to make such records at the time such inspection/maintenance /calibration is performed, or within a reasonable time thereafter, and to provide such instrument's records to the agency that requested them.

A handwritten signature in black ink that reads "Steven J. Carluccio".

Digitally signed under ESRA
by Steven J. Carluccio
on 07/29/2022 07:43 AM

Steven J. Carluccio
Highway Safety Equipment Technical Supervisor
Office of Public Safety

**BREATH SCREENING DEVICE
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: North Syracuse Village Police Department

Manufacturer: INTOXIMETERS INC.

Model: ALCO-SENSOR III

Serial Number: 1062753

Date: July 20, 2022

Reference Solution Lot Number : 22100

Reference Solution Standard : 0.10% at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.099 %	34.00 degrees centigrade
2	0.100 %	34.01 degrees centigrade
3	0.100 %	34.02 degrees centigrade
4	0.102 %	34.01 degrees centigrade

[Faint, illegible text, possibly a signature or stamp]

I hereby certify that I have performed any necessary maintenance procedures and verified the calibration of INTOXIMETERS INC. screener model ALCO-SENSOR III, serial number 1062753, and have determined that it is accurate.

I further certify that I made the entries in these records of inspection/maintenance/calibration at the time that such inspection/maintenance/calibration of the above identified screener was performed, or within a reasonable time thereafter.

Jonathan Mascolo

Digitally signed under ESRA
by Jonathan Mascolo
on 07/28/2022 03:40 PM

**Jonathan Mascolo
Highway Safety Equipment Technician
Office of Public Safety**

**BREATH SCREENING DEVICE
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: North Syracuse Village Police Department

Manufacturer: INTOXIMETERS INC.

Model: ALCO-SENSOR III

Serial Number: 1062753

Date: July 20, 2022

1. General appearance and condition of this device -----: OK

2. Check battery indicators -----: OK

3. Check zero flashing -----: OK

**4. Record any repairs made to this instrument and list services provided : ADJUSTED,
VERIFIED CALIBRATION**

[Faint, illegible text, possibly a stamp or watermark]

Jonathan Mascolo

Digitally signed under ESRA
by Jonathan Mascolo
on 07/28/2022 03:40 PM

**Jonathan Mascolo
Highway Safety Equipment Technician
Office of Public Safety**


SERVICE AUTHORIZATION FORM
Highway Safety Technology Unit
80 South Swan Street
Albany, NY 12210
Ph: 518-485-7636 / 518-402-0689
Fax: 518-457-6869

THIS FORM IS USED TO AUTHORIZE THE DIVISION OF CRIMINAL JUSTICE SERVICES TO PERFORM SERVICE ON THE INSTRUMENT IDENTIFIED. THIS AUTHORIZATION MUST ACCOMPANY EACH PIECE OF EQUIPMENT WHEN IT IS PRESENTED TO THE EQUIPMENT REPAIR CENTER FOR SERVICE. **PLEASE SIGN AND DATE FORM.**

SECTION I - AGENCY INFORMATION

Name of Agency North Syracuse Police Department		Today's Date 07/20/2022
Address 600 South Bay Road	City, State, ZIP North Syracuse, NY 13212	
Contact Person Sergeant Jeffrey D. Tripp	Contact Telephone 315-418-7938 / 315-458-5670	
Contact Email jtripp@northsyracuse.ny.gov	Best Way to Reach You During Appointment Period 315-418-7938 (work cell)	

SECTION II - INSTRUMENT INFORMATION

Instrument Manufacturer (Please Circle)			
NPAS DMT	GUTH	Applied Concepts(Stalker)	Decatur(Genesis) Kustom(Eagle/Falcon) MPH Alco-Sensor Other:
Model Number Alco Sensor III		Instrument Serial Number 1062753	
Antenna Serial Number (if applicable) Antenna 1 - / Antenna 2 -		Type of maintenance (check appropriate) <input checked="" type="checkbox"/> Calibration Check <input type="checkbox"/> Repair	
Brief Description of Malfunction Calibration Check			
List all parts, cables and/or accessories being submitted. Tuning forks must accompany all radar units (if applicable). (Please Circle)			
DMT/SIM: DMT SIM Tubes Keyboard Other: <u>CASE / R10E3</u>		Radar: Control Unit Antenna(s) Antenna Cable(s) Power Cord Forks Other:	
<i>I acknowledge I am signing this Service Authorization Form as an act and deed of said agency, organization, or municipality, and that I am duly authorized to sign same for the uses and purposes mentioned herein. PLEASE SIGN AND DATE.</i>			
Signature 		Date 07/20/2022	

DCJS USE ONLY

Initial Test: <u>0.095</u> % <u>34.00</u> °C	Test 1: <u>0.099</u> % <u>34.00</u> °C
Ref Std Lot # <u>22100</u> <u>0.100</u> %	Test 2: <u>0.100</u> % <u>34.01</u> °C
Re-adjusted	Test 3: <u>0.100</u> % <u>34.02</u> °C
	Test 4: <u>0.102</u> % <u>34.01</u> °C



Division of Criminal Justice Services

CERTIFICATION PURSUANT TO CPLR 4518 OF RECORDS MAINTAINED IN THE REGULAR COURSE OF BUSINESS

I, Steven J. Carluccio, Highway Safety Equipment Technical Supervisor, Office of Public Safety, New York State Division of Criminal Justice Services, 80 South Swan Street, Albany, New York 12210, having been duly designated and authorized by the Commissioner of the Division of Criminal Justice Services, do hereby certify and authenticate, as provided by subdivision c of Rule 4518 of the Civil Practice Law and Rules, that the document annexed hereto is an exact copy of an electronic record of the Office of Public Safety which is in my possession, custody and control. This record, once created, is stored in an electronic format that cannot thereafter be altered or modified.

I further certify that the original electronic record of inspection/maintenance/calibration of INTOXIMETERS INC., ALCO-SENSOR SCREENER, serial number 1221225, performed on July 20, 2022 by Office of Public Safety employee Jonathan Mascolo, of which the annexed is an exact copy, was made in the regular course of business of the New York State Division of Criminal Justice Services, Office of Public Safety, that such record was made at the time such inspection/ maintenance/calibration was performed or within a reasonable time thereafter, and further that it is the regular course of the Office of Public Safety's business to make such records at the time such inspection/maintenance /calibration is performed, or within a reasonable time thereafter, and to provide such instrument's records to the agency that requested them.

A handwritten signature in black ink that reads "Steven J. Carluccio".

Digitally signed under ESRA
by Steven J. Carluccio
on 07/28/2022 02:57 PM

Steven J. Carluccio
Highway Safety Equipment Technical Supervisor
Office of Public Safety

**BREATH SCREENING DEVICE
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: North Syracuse Village Police Department

Manufacturer: INTOXIMETERS INC.

Model: ALCO-SENSOR

Serial Number: 1221225

Date: July 20, 2022


Reference Solution Lot Number : 22100

Reference Solution Standard : 0.10% at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.10 %	34.00 degrees centigrade
2	0.10 %	34.00 degrees centigrade
3	0.10 %	34.02 degrees centigrade
4	0.10 %	34.02 degrees centigrade

I hereby certify that I have performed any necessary maintenance procedures and verified the calibration of INTOXIMETERS INC. screener model ALCO-SENSOR, serial number 1221225, and have determined that it is accurate.

I further certify that I made the entries in these records of inspection/maintenance/calibration at the time that such inspection/maintenance/calibration of the above identified screener was performed, or within a reasonable time thereafter.



Digitally signed under ESRA
by Jonathan Mascolo
on 07/20/2022 01:45 PM

**Jonathan Mascolo
Highway Safety Equipment Technician
Office of Public Safety**

**BREATH SCREENING DEVICE
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: North Syracuse Village Police Department

Manufacturer: INTOXIMETERS INC.

Model: ALCO-SENSOR

Serial Number: 1221225

Date: July 20, 2022

1. General appearance and condition of this device -----: OK

2. Check battery indicators -----: OK

3. Check zero flashing -----: OK

**4. Record any repairs made to this instrument and list services provided : VERIFIED
CALIBRATION**



Jonathan Mascolo

Digitally signed under ESRA
by Jonathan Mascolo
on 07/20/2022 01:45 PM

**Jonathan Mascolo
Highway Safety Equipment Technician
Office of Public Safety**


SERVICE AUTHORIZATION FORM
 Highway Safety Technology Unit
 80 South Swan Street
 Albany, NY 12210
 Ph: 518-485-7636 / 518-402-0689
 Fax: 518-457-6869

THIS FORM IS USED TO AUTHORIZE THE DIVISION OF CRIMINAL JUSTICE SERVICES TO PERFORM SERVICE ON THE INSTRUMENT IDENTIFIED. THIS AUTHORIZATION MUST ACCOMPANY EACH PIECE OF EQUIPMENT WHEN IT IS PRESENTED TO THE EQUIPMENT REPAIR CENTER FOR SERVICE. **PLEASE SIGN AND DATE FORM.**

SECTION I - AGENCY INFORMATION

Name of Agency North Syracuse Police Department		Today's Date 07/20/2022	
Address 600 South Bay Road		City, State, ZIP North Syracuse, NY 13212	
Contact Person Sergeant Jeffrey D. Tripp		Contact Telephone 315-418-7938 / 315-458-5670	
Contact Email jtripp@northsyracuse.ny.gov		Best Way to Reach You During Appointment Period 315-418-7938 (work cell)	

SECTION II - INSTRUMENT INFORMATION

Instrument Manufacturer (Please Circle) NPAS DMT GUTH Applied Concepts(Stalker) Decatur(Genesis) Kustom(Eagle/Falcon) MPH Alco-Sensor Other:			
Model Number Alco Sensor 2 Digit		Instrument Serial Number 1221225	
Antenna Serial Number (if applicable) Antenna 1 - / Antenna 2 -		Type of maintenance (check appropriate) <input checked="checked" type="checkbox"/> Calibration Check <input type="checkbox"/> Repair	
Brief Description of Malfunction Calibration Check			
List all parts, cables and/or accessories being submitted. Tuning forks must accompany all radar units (if applicable). (Please Circle)			
DMT/SIM: DMT SIM Tubes Keyboard Other: CASE/TUBES		Radar: Control Unit Antenna(s) Antenna Cable(s) Power Cord Forks Other:	
I acknowledge I am signing this Service Authorization Form as an act and deed of said agency, organization, or municipality, and that I am duly authorized to sign same for the uses and purposes mentioned herein. PLEASE SIGN AND DATE.			
Signature 		Date 07/20/2022	

DCJS USE ONLY

Initial Test: <u>0.10</u> % <u>34.02</u> °C	Test 1: <u>0.10</u> % <u>34.00</u> °C
Ref Std Lot # <u>22100</u> <u>0.100</u> %	Test 2: <u>0.10</u> % <u>34.00</u> °C
	Test 3: <u>0.10</u> % <u>34.02</u> °C
	Test 4: <u>0.10</u> % <u>34.02</u> °C



Division of Criminal Justice Services

CERTIFICATION PURSUANT TO CPLR 4518 OF RECORDS MAINTAINED IN THE REGULAR COURSE OF BUSINESS

I, Steven J. Carluccio, Highway Safety Equipment Technical Supervisor, Office of Public Safety, New York State Division of Criminal Justice Services, 80 South Swan Street, Albany, New York 12210, having been duly designated and authorized by the Commissioner of the Division of Criminal Justice Services, do hereby certify and authenticate, as provided by subdivision c of Rule 4518 of the Civil Practice Law and Rules, that the document annexed hereto is an exact copy of an electronic record of the Office of Public Safety which is in my possession, custody and control. This record, once created, is stored in an electronic format that cannot thereafter be altered or modified.

I further certify that the original electronic record of inspection/maintenance/calibration of INTOXIMETERS INC., ALCO-SENSOR SCREENER, serial number 1221226, performed on July 20, 2022 by Office of Public Safety employee George Coffin, of which the annexed is an exact copy, was made in the regular course of business of the New York State Division of Criminal Justice Services, Office of Public Safety, that such record was made at the time such inspection/ maintenance/calibration was performed or within a reasonable time thereafter, and further that it is the regular course of the Office of Public Safety's business to make such records at the time such inspection/maintenance /calibration is performed, or within a reasonable time thereafter, and to provide such instrument's records to the agency that requested them.

A handwritten signature in black ink that reads "Steven J. Carluccio".

Digitally signed under ESRA
by Steven J. Carluccio
on 07/28/2022 02:57 PM

Steven J. Carluccio
Highway Safety Equipment Technical Supervisor
Office of Public Safety

**BREATH SCREENING DEVICE
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: North Syracuse Village Police Department

Manufacturer: INTOXIMETERS INC.

Model: ALCO-SENSOR

Serial Number: 1221226

Date: July 20, 2022


Reference Solution Lot Number : 22100

Reference Solution Standard : 0.10% at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.10 %	34.01 degrees centigrade
2	0.09 %	34.00 degrees centigrade
3	0.10 %	34.01 degrees centigrade
4	0.09 %	34.01 degrees centigrade

I hereby certify that I have performed any necessary maintenance procedures and verified the calibration of INTOXIMETERS INC. screener model ALCO-SENSOR, serial number 1221226, and have determined that it is accurate.

I further certify that I made the entries in these records of inspection/maintenance/calibration at the time that such inspection/maintenance/calibration of the above identified screener was performed, or within a reasonable time thereafter.



Digitally signed under ESRA
by George Coffin
on 07/20/2022 02:44 PM

**George Coffin
Highway Safety Equipment Technician
Office of Public Safety**

**BREATH SCREENING DEVICE
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: North Syracuse Village Police Department

Manufacturer: INTOXIMETERS INC.

Model: ALCO-SENSOR

Serial Number: 1221226

Date: July 20, 2022


1. General appearance and condition of this device -----: OK

2. Check battery indicators -----: OK

3. Check zero flashing -----: OK

**4. Record any repairs made to this instrument and list services provided : VERIFIED
CALIBRATION**

[Faint, illegible text]



Digitally signed under ESRA
by George Coffin
on 07/20/2022 02:44 PM

**George Coffin
Highway Safety Equipment Technician
Office of Public Safety**


SERVICE AUTHORIZATION FORM
Highway Safety Technology Unit
80 South Swan Street
Albany, NY 12210
Ph: 518-485-7636 / 518-402-0689
Fax: 518-457-6869

THIS FORM IS USED TO AUTHORIZE THE DIVISION OF CRIMINAL JUSTICE SERVICES TO PERFORM SERVICE ON THE INSTRUMENT IDENTIFIED. THIS AUTHORIZATION MUST ACCOMPANY EACH PIECE OF EQUIPMENT WHEN IT IS PRESENTED TO THE EQUIPMENT REPAIR CENTER FOR SERVICE. **PLEASE SIGN AND DATE FORM.**

SECTION I - AGENCY INFORMATION

Name of Agency North Syracuse Police Department		Today's Date 07/20/2022
Address 600 South Bay Road		City, State, ZIP North Syracuse, NY 13212
Contact Person Sergeant Jeffrey D. Tripp		Contact Telephone 315-418-7938/ 315-458-5670
Contact Email jtripp@northsyracuse.ny.org		Best Way to Reach You During Appointment Period 315-418-7938 (work cell)

SECTION II - INSTRUMENT INFORMATION

Instrument Manufacturer (Please Circle) NPAS DMT GUTH Applied Concepts(Stalker) Decatur(Genesis) Kustom(Eagle/Falcon) MPH Alco-Sensor Other:		
Model Number Alco Sensor 2 Digit		Instrument Serial Number 1221226
Antenna Serial Number (if applicable) Antenna 1 - / Antenna 2 -		Type of maintenance (check appropriate) <input checked="" type="checkbox"/> Calibration Check <input type="checkbox"/> Repair
Brief Description of Malfunction Calibration Check		
List all parts, cables and/or accessories being submitted. Tuning forks must accompany all radar units (if applicable). (Please Circle)		
DMT/SIM: <u>NO BATTERY COVER</u>		Radar:
DMT SIM Tubes	Control Unit Antenna(s) Antenna Cable(s)	
Keyboard Other: <u>CASE/TUBES/TOOL</u>	Power Cord Forks Other:	
I acknowledge I am signing this Service Authorization Form as an act and deed of said agency, organization, or municipality, and that I am duly authorized to sign same for the uses and purposes mentioned herein. PLEASE SIGN AND DATE.		
Signature 		Date 07/20/2022

DCJS USE ONLY

Initial Test: <u>0.10</u> % <u>34.01</u> °C	Test 1: <u>0.10</u> % <u>34.01</u> °C
Ref Std Lot # <u>22100</u> <u>0.10</u> %	Test 2: <u>0.09</u> % <u>34.00</u> °C
	Test 3: <u>0.10</u> % <u>34.01</u> °C
	Test 4: <u>0.09</u> % <u>34.01</u> °C



Division of Criminal Justice Services

CERTIFICATION PURSUANT TO CPLR 4518 OF RECORDS MAINTAINED IN THE REGULAR COURSE OF BUSINESS

I, Steven J. Carluccio, Highway Safety Equipment Technical Supervisor, Office of Public Safety, New York State Division of Criminal Justice Services, 80 South Swan Street, Albany, New York 12210, having been duly designated and authorized by the Commissioner of the Division of Criminal Justice Services, do hereby certify and authenticate, as provided by subdivision c of Rule 4518 of the Civil Practice Law and Rules, that the document annexed hereto is an exact copy of an electronic record of the Office of Public Safety which is in my possession, custody and control. This record, once created, is stored in an electronic format that cannot thereafter be altered or modified.

I further certify that the original electronic record of inspection/maintenance/calibration of INTOXIMETERS INC., ALCO-SENSOR SCREENER, serial number 1221227, performed on July 20, 2022 by Office of Public Safety employee Jonathan Mascolo, of which the annexed is an exact copy, was made in the regular course of business of the New York State Division of Criminal Justice Services, Office of Public Safety, that such record was made at the time such inspection/ maintenance/calibration was performed or within a reasonable time thereafter, and further that it is the regular course of the Office of Public Safety's business to make such records at the time such inspection/maintenance /calibration is performed, or within a reasonable time thereafter, and to provide such instrument's records to the agency that requested them.

A handwritten signature in cursive script that reads "Steven J. Carluccio".

Digitally signed under ESRA
by Steven J. Carluccio
on 07/28/2022 02:58 PM

Steven J. Carluccio
Highway Safety Equipment Technical Supervisor
Office of Public Safety

**BREATH SCREENING DEVICE
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: North Syracuse Village Police Department

Manufacturer: INTOXIMETERS INC.

Model: ALCO-SENSOR

Serial Number: 1221227

Date: July 20, 2022

Reference Solution Lot Number : 22100

Reference Solution Standard : 0.10% at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.10 %	34.00 degrees centigrade
2	0.10 %	34.00 degrees centigrade
3	0.10 %	34.00 degrees centigrade
4	0.10 %	34.01 degrees centigrade

I hereby certify that I have performed any necessary maintenance procedures and verified the calibration of INTOXIMETERS INC. screener model ALCO-SENSOR, serial number 1221227, and have determined that it is accurate.

I further certify that I made the entries in these records of inspection/maintenance/calibration at the time that such inspection/maintenance/calibration of the above identified screener was performed, or within a reasonable time thereafter.



Digitally signed under ESRA
by Jonathan Mascolo
on 07/20/2022 01:42 PM

**Jonathan Mascolo
Highway Safety Equipment Technician
Office of Public Safety**

**BREATH SCREENING DEVICE
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: North Syracuse Village Police Department

Manufacturer: INTOXIMETERS INC.

Model: ALCO-SENSOR

Serial Number: 1221227

Date: July 20, 2022

1. General appearance and condition of this device -----: OK

2. Check battery indicators -----: OK

3. Check zero flashing -----: OK

**4. Record any repairs made to this instrument and list services provided : ADJUSTED,
VERIFIED CALIBRATION**



Jonathan Mascolo

Digitally signed under ESRA
by Jonathan Mascolo
on 07/20/2022 01:42 PM

**Jonathan Mascolo
Highway Safety Equipment Technician
Office of Public Safety**


SERVICE AUTHORIZATION FORM
Highway Safety Technology Unit
80 South Swan Street
Albany, NY 12210
Ph: 518-485-7636 / 518-402-0689
Fax: 518-457-6869

THIS FORM IS USED TO AUTHORIZE THE DIVISION OF CRIMINAL JUSTICE SERVICES TO PERFORM SERVICE ON THE INSTRUMENT IDENTIFIED. THIS AUTHORIZATION MUST ACCOMPANY EACH PIECE OF EQUIPMENT WHEN IT IS PRESENTED TO THE EQUIPMENT REPAIR CENTER FOR SERVICE. PLEASE SIGN AND DATE FORM.

SECTION I - AGENCY INFORMATION

Name of Agency North Syracuse Police Department	Today's Date 07/20/2022
Address 600 South Bay Road	City, State, ZIP North Syracuse, NY 13212
Contact Person Sergeant Jeffrey D. Tripp	Contact Telephone 315-418-7938/ 315-458-5670
Contact Email jttripp@northsyracuseny.org	Best Way to Reach You During Appointment Period 315-418-7938 (work cell)

SECTION II - INSTRUMENT INFORMATION

Instrument Manufacturer (Please Circle) NPAS DMT GUTH Applied Concepts(Stalker) Decatur(Genesis) Kustom(Eagle/Falcon) MPH Alco-Sensor Other:		
Model Number Alco Sensor 2 Digit	Instrument Serial Number 1221227	
Antenna Serial Number (if applicable) Antenna 1 - / Antenna 2 -	Type of maintenance (check appropriate) <input checked="" type="checkbox"/> Calibration Check <input type="checkbox"/> Repair	
Brief Description of Malfunction Calibration Check		
List all parts, cables and/or accessories being submitted. Tuning forks must accompany all radar units (if applicable). (Please Circle)		
DMT/SIM: DMT SIM Tubes Keyboard Other: CASE/GUIDE/TOOL/TUBES		Radar: Control Unit Antenna(s) Antenna Cable(s) Power Cord Forks Other:
<small>I acknowledge I am signing this Service Authorization Form as an act and deed of said agency, organization, or municipality, and that I am duly authorized to sign same for the uses and purposes mentioned herein. PLEASE SIGN AND DATE.</small>		
Signature 	Date 07/20/2022	

DCJS USE ONLY

Initial Test: <u>0.09</u> % <u>34.01</u> °C	Test 1: <u>0.10</u> % <u>34.00</u> °C
Ref Std Lot # <u>22100</u> <u>0.100</u> %	Test 2: <u>0.10</u> % <u>34.00</u> °C
<u>Re-adjusted</u>	Test 3: <u>0.10</u> % <u>34.00</u> °C
	Test 4: <u>0.10</u> % <u>34.01</u> °C