



Certificate of Compliance

Certificate: 70045910

Master Contract: 162029

Project: 70045910

Date Issued: September 18, 2015

Issued to: Fluke Corp.
6920 Seaway Blvd.
Everett, WA 98203
USA

Attention: Mr. Joe Swanzy

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only



Issued by:

A handwritten signature in blue ink that appears to read 'Alvin Kuang'.
Alvin Kuang, Certifier

PRODUCTS

CLASS 3631 06 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT

CLASS 3631 86 - ELECTRICAL EQUIPMENT FOR MEASUREMENT USE (Certified to U.S. Standards)

True RMS Clamp Meters, battery operated 3Vdc,
Fluke Models 374 FC, 375 FC and 376 FC, rated CAT III 1000 V, CAT IV 600 V, 600 A; and
Fluke Model 376, rated CAT III 1000 V, CAT IV 600 V, 1000 A.

Note:

1. The above model is Equipment Class II, Pollution Degree 2, Measurement Category: See ratings above.
2. Mode of operation: Continuous
3. Environmental Conditions: Extended as specified by manufacturer

For details related to rating, size, configuration, etc. reference should be made to the CSA Certification Record or the descriptive report.



Certificate: 70045910

Master Contract: 162029

Project: 70045910

Date Issued: September 18, 2015

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 61010-1-12	-	Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements
CAN/CSA-C22.2 No. 61010-2-032-12	-	Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 2-032: Particular requirements for hand-held and hand-manipulated current sensors for electrical test and measurement
CAN/CSA-C22.2 No. 61010-2-033-12	-	Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 2-033: Particular requirements for hand-held multimeters and other meters, for domestic and professional use, capable of measuring mains voltage
CSA C22.2 No. 115	-	Meter-Mounting Devices
UL Std. No. 61010-1 (3 rd ed.)	-	Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements
UL Std. No. 61010-2-032 (3rd Edition)	-	Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 2-032: Particular requirements for hand-held and hand-manipulated current sensors for electrical test and measurement
UL Std. No. 61010-2-033	-	Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 2-033: Particular requirements for hand-held multimeters and other meters, for domestic and professional use, capable of measuring mains voltage
CSA C17:M84 (R2020)	-	Alternating-Current Electricity-Metering Devices
CSA C17	-	Electricity Meters
CAN/CSA-C22.2 No. 61010-1	-	Test & Measurement Equipment

Reference Standards

IEC 61010-1:2010	-	Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements
IEC 61010-2-032:2012	-	Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 2-032: Particular requirements for hand-held and hand-manipulated current sensors for electrical test and measurement
IEC 61010-2-033:2012	-	Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 2-033: Safety requirements for hand-held probe assemblies for electrical measurement and test



IEC 62052-11: 2020

- Electricity metering equipment General requirements, tests and test conditions - Part 11: Metering equipment"

IEC 62053-41: 2021

- Electricity metering equipment Particular requirements - Part 41: Static meters for DC energy (classes 0,5 and 1)"

IEC 62052-31: 2015

- Electricity metering equipment -General requirements, tests and test conditions- Part 31: Product safety requirements and tests"



Supplement to Certificate of Compliance

Certificate: 70045910

Master Contract: 162029

*The products listed, including the latest revision described below,
are eligible to be marked in accordance with the referenced Certificate.*

Product Certification History

Project	Date	Description
70045910	September 18, 2015	Original certification