

## Limited Warranty

Nanometer Technologies products shall be free of defects in material and workmanship for a period of 1 year from the date of purchase.

In the event of a defect in materials or workmanship, we will either replace or repair without charge (not including shipping costs) at our option any part which in our judgement shows evidence of such defect within 1 year from the date of purchase.


***This warranty does not apply to misuse, abuse, tampered, or altered items.*** At the end of the warranty period Nanometer Technologies shall be under no further obligation expressed or implied. This warranty is in lieu of any other warranty, under no circumstances will Nanometer Technologies be liable for any loss, damage, expense or consequential damages of any kind arising in connection with the use or inability to use Nanometer Technologies products.

Nanometer Technologies Inc  
2501 Golden Hill Rd.  
Paso Robles, CA 93446

Tel: (805) 226-7332

Fax: (805) 226-8753

[www.nanometer.com](http://www.nanometer.com)



Instruction Manual for the


# NTI-32



NANOMETER TECHNOLOGIES

*Fiber Optic Polishers and Components*

### Step 3 Inserting Connectors




Place the connectors into the holes of the **Heating Brick**. Make sure the connector is resting in the bottom of the hole.

**CAUTION: The Heating Brick Will Be Hot!**

# NTI-32

Diagram Layout



Power Indicator

Temperature Dial

## Specifications

Temperature Levels.....	0-200 Celsius 0 - 392 Fahrenheit
Accuracy.....	+/- 3c
Number of Ports.....	32

### NTI-32 INSTRUCTION MANUAL

Document Number RE:00A1-NTI-32

Nanometer Technologies (NMT) has prepared this manual for use by NMT personnel, licensees, and customers. The information contained herein is the property of NMT and shall not be reproduced in whole or in part without the prior written approval of NMT.

NMT reserves the right to make changes without notice to the specification(s) and materials contained herein and shall not be responsible for any damages (including consequential) caused by reliance on the materials presented, including but not limited to typographical or listing errors.

Please address any questions, comments and suggestions to:

### NANOMETER TECHNOLOGIES


2501 Golden Hill Rd.  
Paso Robles, CA 93446  
Tel: (805) 226-7332  
Fax: (805) 266-8753

[www.nanometer.com](http://www.nanometer.com)

December 7th, 2006

## Step 4

### Securing Connector In Foam Handle




Place the **Fiber Cable** into the **Foam Slot** above the connector. Again, make sure the connector is resting at the bottom of the hole in the **Heating Brick**.

Leave the connectors in the oven according to the amount of time suggested to the epoxy manufacturing instructions.

When finished, turn off oven.

## Step 1

### Attaching Foam To NTI-32



Check the top of the handle for dust, particles, etc. Clean if necessary.

Remove the protective layer from the **Adhesive Tape** on the back of the **Foam Slots**.

Place the **Foam Slots** along the top part of the handle.  
(Refer to cover image for help on placement)

## Step 2

### Setting The Temperature



Set the **Oven** temperature to the epoxy manufacturer's temperature recommendation. Turn the dial to the desired temperature level.

The **NTI-32** takes about 15 minutes to reach temperature level.

The Red Light (to the left side of **Temperature Dial**) will show that the **NTI-32** is on.